

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP012422.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 25 00:52:27 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP043208.D 1000 =PP043206.D 750 =PP043207.D
 250 =PP043209.D 50 =PP043210.D

Compound			500	1000	750	250	50	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	2.385	2.373	2.397	2.462	1.876	2.299	E4	10.40
2)	SA	Decachlorobiphenyl	1.411	1.402	1.433	1.490	1.346	1.416	E4	3.68
3)	L1	AR-1016-1	6.787	6.468	6.724	7.004	5.896	6.576	E2	6.47
4)	L1	AR-1016-2	1.059	1.026	1.040	1.097	0.886	1.021	E3	7.87
5)	L1	AR-1016-3	6.495	6.217	6.437	6.431	6.245	6.365	E2	1.97
6)	L1	AR-1016-4	5.197	5.063	5.161	5.338	3.818	4.916	E2	12.65
7)	L1	AR-1016-5	5.137	4.966	5.067	5.181	4.755	5.021	E2	3.37
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1	7.982	7.600	7.809	8.608	6.596	7.719	E2	9.48
32)	L7	AR-1260-2	8.607	8.208	8.390	9.041	8.322	8.514	E2	3.86
33)	L7	AR-1260-3	6.773	6.566	6.639	6.904	6.094	6.595	E2	4.68
34)	L7	AR-1260-4	7.748	7.522	7.674	8.207	7.079	7.646	E2	5.32
35)	L7	AR-1260-5	1.354	1.352	1.377	1.415	1.104	1.320	E3	9.38
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP043208.D 1000 =PP043206.D 750 =PP043207.D
 250 =PP043209.D 50 =PP043210.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.284	2.256	2.283	2.374	1.944	2.228	E4	7.40
2) SA	Decachlorobiphenyl	1.858	1.752	1.823	2.020	1.868	1.864	E4	5.27
3) L1	AR-1016-1	9.243	8.739	9.060	9.936	8.363	9.068	E2	6.50
4) L1	AR-1016-2	1.316	1.267	1.288	1.404	1.174	1.290	E3	6.44
5) L1	AR-1016-3	7.167	6.861	7.092	7.559	6.976	7.131	E2	3.73

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP012422.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 25 00:52:27 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP043208.D 1000 =PP043206.D 750 =PP043207.D
 250 =PP043209.D 50 =PP043210.D

Compound			500	1000	750	250	50	Avg	%RSD	
6)	L1	AR-1016-4	5.588	5.225	5.460	5.981	5.338	5.519	E2	5.29
7)	L1	AR-1016-5	7.028	6.675	6.897	7.343	5.524	6.693	E2	10.41
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1	1.167	1.093	1.134	1.270	1.185	1.170	E3	5.63
32)	L7	AR-1260-2	1.343	1.263	1.309	1.438	1.223	1.315	E3	6.26
33)	L7	AR-1260-3	1.247	1.191	1.228	1.354	1.204	1.245	E3	5.20
34)	L7	AR-1260-4	1.015	0.970	1.001	1.095	0.824	0.981	E3	10.11
35)	L7	AR-1260-5	2.239	2.182	2.236	2.384	2.025	2.213	E3	5.84
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

(#) = Out of Range