

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP012422.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 25 01:19:57 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP043215.D 1000 =PP043213.D 750 =PP043214.D
 250 =PP043216.D 50 =PP043217.D

Compound			500	1000	750	250	50	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	2.478	2.387	2.439	2.486	2.356	2.429	E4	2.35
2)	SA	Decachlorobiphenyl	1.445	1.381	1.434	1.466	1.237	1.393	E4	6.64
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1	5.336	4.965	5.187	5.309	4.739	5.107	E2	4.94
17)	L4	AR-1242-2	8.170	7.744	8.006	8.365	7.563	7.969	E2	4.04
18)	L4	AR-1242-3	4.874	4.786	4.907	4.968	4.751	4.857	E2	1.82
19)	L4	AR-1242-4	4.030	3.869	3.965	3.888	3.856	3.922	E2	1.89
20)	L4	AR-1242-5	3.683	3.627	3.683	3.773	3.481	3.649	E2	2.95
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP043215.D 1000 =PP043213.D 750 =PP043214.D
 250 =PP043216.D 50 =PP043217.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.404	2.309	2.366	2.474	2.234	2.357	E4	3.89
2) SA	Decachlorobiphenyl	1.983	1.818	1.893	2.128	2.009	1.966	E4	6.00
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP012422.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 25 01:19:57 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP043215.D 1000 =PP043213.D 750 =PP043214.D
 250 =PP043216.D 50 =PP043217.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1	7.477	6.864	7.156	7.895	7.574	7.393 E2	5.36
17) L4	AR-1242-2	1.068	0.978	1.018	1.111	1.015	1.038 E3	4.99
18) L4	AR-1242-3	5.836	5.415	5.621	5.931	5.984	5.757 E2	4.11
19) L4	AR-1242-4	5.320	4.916	5.117	5.427	4.214	4.999 E2	9.61
20) L4	AR-1242-5	6.743	6.126	6.464	6.543	6.093	6.394 E2	4.36
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range