

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP012422.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 25 01:42:19 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP043220.D 1000 =PP043218.D 750 =PP043219.D
 250 =PP043221.D 50 =PP043222.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.470	2.387	2.409	2.447	2.373	2.417	E4	1.68
2) SA	Decachlorobiphenyl	1.445	1.373	1.410	1.491	1.326	1.409	E4	4.51
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	3.947	3.785	3.870	3.950	4.051	3.921	E2	2.53
22) L5	AR-1248-2	5.671	5.276	5.415	5.610	5.235	5.442	E2	3.58
23) L5	AR-1248-3	6.698	6.290	6.442	6.694	6.363	6.497	E2	2.91
24) L5	AR-1248-4	6.586	6.145	6.322	6.495	5.551	6.220	E2	6.60
25) L5	AR-1248-5	6.374	5.954	6.174	6.220	5.674	6.079	E2	4.47
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP043220.D 1000 =PP043218.D 750 =PP043219.D
 250 =PP043221.D 50 =PP043222.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.509	2.403	2.434	2.501	2.294	2.428	E4	3.59
2) SA	Decachlorobiphenyl	1.979	1.832	1.899	2.134	1.962	1.961	E4	5.75
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP012422.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 25 01:42:19 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP043220.D 1000 =PP043218.D 750 =PP043219.D
 250 =PP043221.D 50 =PP043222.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	5.463	5.121	5.272	5.634	5.879	5.474 E2	5.44
22) L5	AR-1248-2	7.683	6.999	7.323	8.109	7.938	7.610 E2	5.94
23) L5	AR-1248-3	7.885	7.207	7.546	8.150	7.389	7.635 E2	4.98
24) L5	AR-1248-4	9.406	8.690	8.965	9.752	9.018	9.166 E2	4.53
25) L5	AR-1248-5	9.282	8.665	8.893	9.611	7.719	8.834 E2	8.16
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range