

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012422\
 Data File : PP043216.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Jan 2022 22:17
 Operator : AJ\MA
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 01:15:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 25 01:04:38 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.870	4.035	621560	618598	25.396	25.901
2)	SA Decachlor...	10.797	9.366	366605	531968	25.609	27.208
Target Compounds							
16)	L4 AR-1242-1	6.174	5.300	132718	197380	255.272	268.614
17)	L4 AR-1242-2	6.197	5.320	209115	277653	259.088	266.072
18)	L4 AR-1242-3	6.262	5.514	124192	148285	254.303	260.111
19)	L4 AR-1242-4	6.367	5.609	97201	135679	246.830	261.166
20)	L4 AR-1242-5	7.147	6.177	94325	163579	255.519	252.874



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012422\
Data File : PP043216.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Jan 2022 22:17
Operator : AJ\MA
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 01:15:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012422.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 25 01:04:38 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.870 min
Delta R.T.: 0.000 min
Response: 621560
Conc: 25.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 4.035 min
Delta R.T.: 0.000 min
Response: 618598
Conc: 25.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.797 min
Delta R.T.: 0.000 min
Response: 366605
Conc: 25.61 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.366 min
Delta R.T.: 0.000 min
Response: 531968
Conc: 27.21 ng/ml



#16 AR-1242-1


R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 132718
Conc: 255.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 5.300 min
Delta R.T.: 0.000 min
Response: 197380
Conc: 268.61 ng/ml




#17 AR-1242-2

R.T.: 6.197 min
Delta R.T.: 0.000 min
Response: 209115
Conc: 259.09 ng/ml



#17 AR-1242-2


R.T.: 5.320 min
Delta R.T.: 0.000 min
Response: 277653
Conc: 266.07 ng/ml



#18 AR-1242-3


R.T.: 6.262 min
Delta R.T.: 0.000 min
Response: 124192
Conc: 254.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 5.514 min
Delta R.T.: 0.000 min
Response: 148285
Conc: 260.11 ng/ml




#19 AR-1242-4

R.T.: 6.367 min
Delta R.T.: 0.000 min
Response: 97201
Conc: 246.83 ng/ml



#19 AR-1242-4


R.T.: 5.609 min
Delta R.T.: 0.000 min
Response: 135679
Conc: 261.17 ng/ml



#20 AR-1242-5

R.T.: 7.147 min
Delta R.T.: 0.000 min
Response: 94325
Conc: 255.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 6.177 min
Delta R.T.: 0.000 min
Response: 163579
Conc: 252.87 ng/ml