

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012422\
 Data File : PP043221.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Jan 2022 23:40
 Operator : AJ\MA
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 01:37:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 25 01:29:48 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.871	4.036	611815	625332	25.194	25.399
2)	SA Decachlor...	10.798	9.368	372763	533519	26.073	27.208
Target Compounds							
21)	L5 AR-1248-1	6.175	5.301	98760	140852	254.008	262.170
22)	L5 AR-1248-2	6.470	5.566	140261	202725	255.334	269.274
23)	L5 AR-1248-3	6.685	5.610	167339	203759	256.223	264.723
24)	L5 AR-1248-4	7.109	5.797	162370	243789	254.223	264.900
25)	L5 AR-1248-5	7.149	6.222	155495	240280	251.590	263.673



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012422\
Data File : PP043221.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Jan 2022 23:40
Operator : AJ\MA
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 01:37:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012422.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 25 01:29:48 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.871 min
Delta R.T.: 0.000 min
Response: 611815
Conc: 25.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#1 Tetrachloro-m-xylene

R.T.: 4.036 min
Delta R.T.: 0.000 min
Response: 625332
Conc: 25.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.798 min
Delta R.T.: 0.000 min
Response: 372763
Conc: 26.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.368 min
Delta R.T.: 0.000 min
Response: 533519
Conc: 27.21 ng/ml



#21 AR-1248-1


R.T.: 6.175 min
Delta R.T.: 0.000 min
Response: 98760
Conc: 254.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#21 AR-1248-1

R.T.: 5.301 min
Delta R.T.: 0.000 min
Response: 140852
Conc: 262.17 ng/ml




#22 AR-1248-2

R.T.: 6.470 min
Delta R.T.: 0.000 min
Response: 140261
Conc: 255.33 ng/ml



#22 AR-1248-2


R.T.: 5.566 min
Delta R.T.: 0.000 min
Response: 202725
Conc: 269.27 ng/ml



#23 AR-1248-3


R.T.: 6.685 min
Delta R.T.: 0.000 min
Response: 167339
Conc: 256.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#23 AR-1248-3

R.T.: 5.610 min
Delta R.T.: 0.000 min
Response: 203759
Conc: 264.72 ng/ml




#24 AR-1248-4

R.T.: 7.109 min
Delta R.T.: 0.000 min
Response: 162370
Conc: 254.22 ng/ml



#24 AR-1248-4


R.T.: 5.797 min
Delta R.T.: 0.000 min
Response: 243789
Conc: 264.90 ng/ml



#25 AR-1248-5

R.T.: 7.149 min
Delta R.T.: 0.000 min
Response: 155495
Conc: 251.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250



#25 AR-1248-5

R.T.: 6.222 min
Delta R.T.: 0.000 min
Response: 240280
Conc: 263.67 ng/ml