

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012423\
 Data File : PP054846.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2023 02:55
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP012423AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 03:27:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:26:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.331	3.578	90063782	76725835	52.292	52.216
2)	SA Decachlor...	10.088	8.595	91713495	73009633	51.950	51.418
Target Compounds							
26)	L6 AR-1254-1	6.361	5.467	39889730	41012524	505.622	507.844
27)	L6 AR-1254-2	6.581	5.615	60594714	35995442	500.688	502.992
28)	L6 AR-1254-3	6.948	6.020	61892729	57615150	505.288	517.157
29)	L6 AR-1254-4	7.236	6.249	44688837	31569761	507.182	513.268
30)	L6 AR-1254-5	7.656	6.669	50483309	48830610	499.040	515.069



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012423\
Data File : PP054846.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2023 02:55
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP012423AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 24 03:27:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:26:58 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.331 min
Delta R.T.: 0.000 min
Response: 90063782
Conc: 52.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP012423AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.578 min
Delta R.T.: 0.000 min
Response: 76725835
Conc: 52.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.088 min
Delta R.T.: 0.000 min
Response: 91713495
Conc: 51.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.595 min
Delta R.T.: 0.000 min
Response: 73009633
Conc: 51.42 ng/ml



#26 AR-1254-1


R.T.: 6.361 min
Delta R.T.: 0.000 min
Response: 39889730
Conc: 505.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP012423AR1254




#26 AR-1254-1

R.T.: 5.467 min
Delta R.T.: 0.000 min
Response: 41012524
Conc: 507.84 ng/ml




#27 AR-1254-2

R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 60594714
Conc: 500.69 ng/ml



#27 AR-1254-2


R.T.: 5.615 min
Delta R.T.: 0.000 min
Response: 35995442
Conc: 502.99 ng/ml



#28 AR-1254-3


R.T.: 6.948 min
Delta R.T.: 0.000 min
Response: 61892729
Conc: 505.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP012423AR1254




#28 AR-1254-3

R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 57615150
Conc: 517.16 ng/ml




#29 AR-1254-4

R.T.: 7.236 min
Delta R.T.: 0.000 min
Response: 44688837
Conc: 507.18 ng/ml



#29 AR-1254-4


R.T.: 6.249 min
Delta R.T.: 0.000 min
Response: 31569761
Conc: 513.27 ng/ml



#30 AR-1254-5

R.T.: 7.656 min
Delta R.T.: 0.000 min
Response: 50483309
Conc: 499.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP012423AR1254



#30 AR-1254-5

R.T.: 6.669 min
Delta R.T.: 0.000 min
Response: 48830610
Conc: 515.07 ng/ml