

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012424\
 Data File : PP063094.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jan 2024 01:30
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 02:56:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 25 02:56:18 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.409	3.674	118.6E6	104.4E6	73.357	73.083
2)	SA Decachlor...	10.176	8.768	149.7E6	185.3E6	74.057	74.016
Target Compounds							
3)	L1 AR-1016-1	5.584	4.779	37666711	41500905	733.930	731.871
4)	L1 AR-1016-2	5.606	4.798	58982115	60138021	732.838	735.311
5)	L1 AR-1016-3	5.668	4.977	35063609	32884474	732.742	735.733
6)	L1 AR-1016-4	5.767	5.020	29878097	26869566	734.731	738.883
7)	L1 AR-1016-5	6.063	5.236	30505207	36661882	733.093	723.523
31)	L7 AR-1260-1	7.195	6.286	67231860	81297440	728.883	736.273
32)	L7 AR-1260-2	7.453	6.476	80418486	100.9E6	732.531	738.242
33)	L7 AR-1260-3	7.814	6.632	62686574	95205967	737.514	737.475
34)	L7 AR-1260-4	8.040	7.109	74599157	84938438	739.247	738.444
35)	L7 AR-1260-5	8.360	7.353	154.4E6	206.0E6	739.330	736.988



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012424\
Data File : PP063094.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jan 2024 01:30
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 02:56:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012424.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 25 02:56:18 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.409 min
Delta R.T.: 0.000 min
Response: 118642129
Conc: 73.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.674 min
Delta R.T.: 0.000 min
Response: 104350802
Conc: 73.08 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.176 min
Delta R.T.: 0.000 min
Response: 149674106
Conc: 74.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.768 min
Delta R.T.: 0.000 min
Response: 185255328
Conc: 74.02 ng/ml



#3 AR-1016-1


R.T.: 5.584 min
Delta R.T.: 0.000 min
Response: 37666711
Conc: 733.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#3 AR-1016-1

R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 41500905
Conc: 731.87 ng/ml




#4 AR-1016-2

R.T.: 5.606 min
Delta R.T.: 0.000 min
Response: 58982115
Conc: 732.84 ng/ml



#4 AR-1016-2


R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 60138021
Conc: 735.31 ng/ml



#5 AR-1016-3


R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 35063609
Conc: 732.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#5 AR-1016-3

R.T.: 4.977 min
Delta R.T.: 0.000 min
Response: 32884474
Conc: 735.73 ng/ml




#6 AR-1016-4

R.T.: 5.767 min
Delta R.T.: 0.000 min
Response: 29878097
Conc: 734.73 ng/ml



#6 AR-1016-4


R.T.: 5.020 min
Delta R.T.: 0.000 min
Response: 26869566
Conc: 738.88 ng/ml



#7 AR-1016-5


R.T.: 6.063 min
Delta R.T.: 0.000 min
Response: 30505207
Conc: 733.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#7 AR-1016-5

R.T.: 5.236 min
Delta R.T.: 0.000 min
Response: 36661882
Conc: 723.52 ng/ml




#31 AR-1260-1

R.T.: 7.195 min
Delta R.T.: 0.000 min
Response: 67231860
Conc: 728.88 ng/ml



#31 AR-1260-1


R.T.: 6.286 min
Delta R.T.: 0.000 min
Response: 81297440
Conc: 736.27 ng/ml



#32 AR-1260-2


R.T.: 7.453 min
Delta R.T.: 0.000 min
Response: 80418486
Conc: 732.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#32 AR-1260-2

R.T.: 6.476 min
Delta R.T.: 0.000 min
Response: 100924405
Conc: 738.24 ng/ml




#33 AR-1260-3

R.T.: 7.814 min
Delta R.T.: 0.000 min
Response: 62686574
Conc: 737.51 ng/ml



#33 AR-1260-3


R.T.: 6.632 min
Delta R.T.: 0.000 min
Response: 95205967
Conc: 737.47 ng/ml



#34 AR-1260-4


R.T.: 8.040 min
Delta R.T.: 0.000 min
Response: 74599157
Conc: 739.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#34 AR-1260-4

R.T.: 7.109 min
Delta R.T.: 0.000 min
Response: 84938438
Conc: 738.44 ng/ml



#35 AR-1260-5

R.T.: 8.360 min
Delta R.T.: 0.000 min
Response: 154382727
Conc: 739.33 ng/ml



#35 AR-1260-5

R.T.: 7.353 min
Delta R.T.: 0.000 min
Response: 206013528
Conc: 736.99 ng/ml