

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP012522\  
 Data File : PP043264.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 25 Jan 2022 19:03  
 Operator : AJ\MA  
 Sample : N1222-16 10X  
 Misc :  
 ALS Vial : 33 Sample Multiplier: 1

Instrument :  
 ECD\_P  
 ClientSampleId :  
 CLI-2022-8B

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jan 25 21:40:47 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP012422.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Tue Jan 25 03:09:04 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1 | Resp#2 | ng/ml  | ng/ml  |
|-----------------------------|--------|-------|--------|--------|--------|--------|
| -----                       |        |       |        |        |        |        |
| System Monitoring Compounds |        |       |        |        |        |        |
| 1) SA Tetrachlo...          | 4.868  | 4.033 | 44249  | 48897  | 1.925  | 2.194  |
| 2) SA Decachlor...          | 10.790 | 9.361 | 42126  | 43719  | 2.974  | 2.345  |
| Target Compounds            |        |       |        |        |        |        |
| 30) L6 AR-1254-5            | 8.415  | 0.000 | 9680   | 0      | 11.985 | N.D. # |
| -----                       |        |       |        |        |        |        |



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP012522\  
Data File : PP043264.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 25 Jan 2022 19:03  
Operator : AJ\MA  
Sample : N1222-16 10X  
Misc :  
ALS Vial : 33 Sample Multiplier: 1

Instrument :  
ECD\_P  
ClientSampleId :  
CLI-2022-8B

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jan 25 21:40:47 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP012422.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Tue Jan 25 03:09:04 2022  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.868 min  
Delta R.T.: -0.002 min  
Response: 44249  
Conc: 1.93 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :  
CLI-2022-8B




#1 Tetrachloro-m-xylene

R.T.: 4.033 min  
Delta R.T.: -0.002 min  
Response: 48897  
Conc: 2.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.790 min  
Delta R.T.: -0.008 min  
Response: 42126  
Conc: 2.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.361 min  
Delta R.T.: -0.006 min  
Response: 43719  
Conc: 2.35 ng/ml



#30 AR-1254-5

R.T.: 8.415 min  
Delta R.T.: 0.007 min  
Response: 9680  
Conc: 11.98 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :  
CLI-2022-8B



#30 AR-1254-5

R.T.: 0.000 min  
Exp R.T. : 7.430 min  
Response: 0  
Conc: N.D.