

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012523\
 Data File : PP054856.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2023 09:52
 Operator : YP\AJ
 Sample : 01235-04
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 15:53:42 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.330	3.577	16577536	14246365	9.534	9.490
2)	SA Decachlor...	10.086	8.593	19638128	15770042	11.341	11.677
Target Compounds							
3)	L1 AR-1016-1	5.501	4.663	1788333	248253	26.608	5.016 #
4)	L1 AR-1016-2	5.524	4.681	1563021	745746	16.310	10.774 #
5)	L1 AR-1016-3	5.589	4.858	1075503	300706	17.980	8.084 #
6)	L1 AR-1016-4	5.684	4.900	197699	418469	4.126	13.318 #
7)	L1 AR-1016-5	5.982	5.112	724741	407085	14.302	9.957 #
8)	L2 AR-1221-1	4.535	3.778	193233	117743	8.217	5.939 #
9)	L2 AR-1221-2	4.638	3.877	486530	275114	27.384	18.561 #
10)	L2 AR-1221-3	4.716	3.948	880260	1137153	16.359	25.206 #
11)	L3 AR-1232-1	4.716	3.948	880260	1137153	19.833	30.851 #
12)	L3 AR-1232-2	5.232	4.681	1956478	745746	86.513	23.346 #
13)	L3 AR-1232-3	5.524	4.858	1563021	300706	35.590	18.038 #
14)	L3 AR-1232-4	5.684	4.941	197699	395940	8.932	24.553 #
15)	L3 AR-1232-5	5.777	5.112	424133	407085	24.226	23.049
16)	L4 AR-1242-1	5.501	4.663	1788333	248253	31.941	6.061 #
17)	L4 AR-1242-2	5.524	4.681	1563021	745746	19.579	13.060 #
18)	L4 AR-1242-3	5.589	4.858	1075503	300706	21.657	9.742 #
19)	L4 AR-1242-4	5.684	4.941	197699	395940	4.933	12.188 #
20)	L4 AR-1242-5	6.427	5.466	740713	507270	16.659	13.574
21)	L5 AR-1248-1	5.501	4.663	1788333	248253	42.882	8.007 #
22)	L5 AR-1248-2	5.777	4.900	424133	418469	6.671	9.151 #
23)	L5 AR-1248-3	5.982	4.941	724741	395940	10.479	8.305
24)	L5 AR-1248-4	6.390	5.112	496141	407085	6.542	7.371
25)	L5 AR-1248-5	6.427	5.508	740713	298114	10.038	6.063 #
26)	L6 AR-1254-1	6.361	5.466	915412	507270	11.603	6.281 #
27)	L6 AR-1254-2	6.581	5.618	776007	336345	6.412	4.700 #
28)	L6 AR-1254-3	6.951	6.019	561502	766177	4.584	6.877 #
29)	L6 AR-1254-4	7.242	6.249	93615	356801	1.062	5.801 #
30)	L6 AR-1254-5	7.677f	6.666	3080536	628668	30.452	6.631 #
31)	L7 AR-1260-1	7.113	6.150	594264	557468	6.234	7.189
32)	L7 AR-1260-2	7.375	6.341	1529009	470331	13.586	5.232 #
33)	L7 AR-1260-3	7.733	6.496	37242	291196	0.463	3.429 #
34)	L7 AR-1260-4	7.966	6.969	453592	46236	4.673	0.646 #
35)	L7 AR-1260-5	8.277	7.208	821103	567254	4.617	3.690
36)	L8 AR-1262-1	7.733	6.762	37242	33028	0.316	0.769 #
37)	L8 AR-1262-2	8.277	6.969	821103	46236	4.077	0.499 #
38)	L8 AR-1262-3	8.568f	7.489	56443	1085350	0.406	14.896 #
39)	L8 AR-1262-4	8.665	7.552	628198	573245	8.171	4.383 #
40)	L8 AR-1262-5	0.000	8.052	0	5269	N.D.	0.089 #
41)	L9 AR-1268-1	8.568f	7.489	56443	1085350	0.231	5.020 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012523\
 Data File : PP054856.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2023 09:52
 Operator : YP\AJ
 Sample : 01235-04
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 15:53:42 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.691	7.552	243224	573245	1.071	2.938 #
43) L9	AR-1268-3	8.903	7.773	533657	103590	2.671	0.603 #
44) L9	AR-1268-4	0.000	8.052	0	5269	N.D.	0.083 #
45) L9	AR-1268-5	9.743	8.343	267573	181944	0.421	0.345



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012523\
Data File : PP054856.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2023 09:52
Operator : YP\AJ
Sample : 01235-04
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 24 15:53:42 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.330 min
Delta R.T.: -0.002 min
Response: 16577536
Conc: 9.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.577 min
Delta R.T.: -0.002 min
Response: 14246365
Conc: 9.49 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.086 min
Delta R.T.: -0.005 min
Response: 19638128
Conc: 11.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.593 min
Delta R.T.: -0.005 min
Response: 15770042
Conc: 11.68 ng/ml



#3 AR-1016-1


R.T.: 5.501 min
Delta R.T.: -0.002 min
Response: 1788333
Conc: 26.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.663 min
Delta R.T.: -0.001 min
Response: 248253
Conc: 5.02 ng/ml




#4 AR-1016-2

R.T.: 5.524 min
Delta R.T.: -0.002 min
Response: 1563021
Conc: 16.31 ng/ml



#4 AR-1016-2


R.T.: 4.681 min
Delta R.T.: -0.001 min
Response: 745746
Conc: 10.77 ng/ml



#5 AR-1016-3


R.T.: 5.589 min
Delta R.T.: 0.000 min
Response: 1075503
Conc: 17.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.858 min
Delta R.T.: -0.001 min
Response: 300706
Conc: 8.08 ng/ml




#6 AR-1016-4

R.T.: 5.684 min
Delta R.T.: -0.003 min
Response: 197699
Conc: 4.13 ng/ml



#6 AR-1016-4


R.T.: 4.900 min
Delta R.T.: -0.002 min
Response: 418469
Conc: 13.32 ng/ml



#7 AR-1016-5


R.T.: 5.982 min
Delta R.T.: 0.000 min
Response: 724741
Conc: 14.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.112 min
Delta R.T.: -0.003 min
Response: 407085
Conc: 9.96 ng/ml




#8 AR-1221-1

R.T.: 4.535 min
Delta R.T.: -0.005 min
Response: 193233
Conc: 8.22 ng/ml



#8 AR-1221-1


R.T.: 3.778 min
Delta R.T.: -0.013 min
Response: 117743
Conc: 5.94 ng/ml



#9 AR-1221-2


R.T.: 4.638 min
Delta R.T.: 0.012 min
Response: 486530
Conc: 27.38 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.877 min
Delta R.T.: 0.000 min
Response: 275114
Conc: 18.56 ng/ml




#10 AR-1221-3

R.T.: 4.716 min
Delta R.T.: 0.014 min
Response: 880260
Conc: 16.36 ng/ml



#10 AR-1221-3


R.T.: 3.948 min
Delta R.T.: -0.004 min
Response: 1137153
Conc: 25.21 ng/ml



#11 AR-1232-1


R.T.: 4.716 min
Delta R.T.: 0.013 min
Response: 880260
Conc: 19.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.948 min
Delta R.T.: -0.005 min
Response: 1137153
Conc: 30.85 ng/ml




#12 AR-1232-2

R.T.: 5.232 min
Delta R.T.: -0.003 min
Response: 1956478
Conc: 86.51 ng/ml



#12 AR-1232-2


R.T.: 4.681 min
Delta R.T.: -0.002 min
Response: 745746
Conc: 23.35 ng/ml



#13 AR-1232-3


R.T.: 5.524 min
Delta R.T.: -0.002 min
Response: 1563021
Conc: 35.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.858 min
Delta R.T.: 0.000 min
Response: 300706
Conc: 18.04 ng/ml




#14 AR-1232-4

R.T.: 5.684 min
Delta R.T.: -0.003 min
Response: 197699
Conc: 8.93 ng/ml



#14 AR-1232-4


R.T.: 4.941 min
Delta R.T.: -0.002 min
Response: 395940
Conc: 24.55 ng/ml



#15 AR-1232-5


R.T.: 5.777 min
Delta R.T.: -0.002 min
Response: 424133
Conc: 24.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.112 min
Delta R.T.: -0.003 min
Response: 407085
Conc: 23.05 ng/ml




#16 AR-1242-1

R.T.: 5.501 min
Delta R.T.: -0.002 min
Response: 1788333
Conc: 31.94 ng/ml



#16 AR-1242-1


R.T.: 4.663 min
Delta R.T.: -0.002 min
Response: 248253
Conc: 6.06 ng/ml



#17 AR-1242-2


R.T.: 5.524 min
Delta R.T.: -0.002 min
Response: 1563021
Conc: 19.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.681 min
Delta R.T.: -0.002 min
Response: 745746
Conc: 13.06 ng/ml




#18 AR-1242-3

R.T.: 5.589 min
Delta R.T.: 0.000 min
Response: 1075503
Conc: 21.66 ng/ml



#18 AR-1242-3


R.T.: 4.858 min
Delta R.T.: -0.002 min
Response: 300706
Conc: 9.74 ng/ml



#19 AR-1242-4


R.T.: 5.684 min
Delta R.T.: -0.002 min
Response: 197699
Conc: 4.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.941 min
Delta R.T.: -0.002 min
Response: 395940
Conc: 12.19 ng/ml




#20 AR-1242-5

R.T.: 6.427 min
Delta R.T.: 0.000 min
Response: 740713
Conc: 16.66 ng/ml



#20 AR-1242-5


R.T.: 5.466 min
Delta R.T.: -0.002 min
Response: 507270
Conc: 13.57 ng/ml



#21 AR-1248-1


R.T.: 5.501 min
Delta R.T.: 0.000 min
Response: 1788333
Conc: 42.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.663 min
Delta R.T.: 0.000 min
Response: 248253
Conc: 8.01 ng/ml




#22 AR-1248-2

R.T.: 5.777 min
Delta R.T.: 0.000 min
Response: 424133
Conc: 6.67 ng/ml



#22 AR-1248-2


R.T.: 4.900 min
Delta R.T.: -0.001 min
Response: 418469
Conc: 9.15 ng/ml



#23 AR-1248-3


R.T.: 5.982 min
Delta R.T.: 0.000 min
Response: 724741
Conc: 10.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.941 min
Delta R.T.: -0.001 min
Response: 395940
Conc: 8.31 ng/ml




#24 AR-1248-4

R.T.: 6.390 min
Delta R.T.: 0.002 min
Response: 496141
Conc: 6.54 ng/ml



#24 AR-1248-4


R.T.: 5.112 min
Delta R.T.: -0.002 min
Response: 407085
Conc: 7.37 ng/ml



#25 AR-1248-5


R.T.: 6.427 min
Delta R.T.: 0.000 min
Response: 740713
Conc: 10.04 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.508 min
Delta R.T.: 0.000 min
Response: 298114
Conc: 6.06 ng/ml




#26 AR-1254-1

R.T.: 6.361 min
Delta R.T.: 0.000 min
Response: 915412
Conc: 11.60 ng/ml



#26 AR-1254-1


R.T.: 5.466 min
Delta R.T.: -0.001 min
Response: 507270
Conc: 6.28 ng/ml



#27 AR-1254-2


R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 776007
Conc: 6.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.618 min
Delta R.T.: 0.002 min
Response: 336345
Conc: 4.70 ng/ml




#28 AR-1254-3

R.T.: 6.951 min
Delta R.T.: 0.003 min
Response: 561502
Conc: 4.58 ng/ml



#28 AR-1254-3


R.T.: 6.019 min
Delta R.T.: -0.001 min
Response: 766177
Conc: 6.88 ng/ml



#29 AR-1254-4


R.T.: 7.242 min
Delta R.T.: 0.006 min
Response: 93615
Conc: 1.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.249 min
Delta R.T.: 0.000 min
Response: 356801
Conc: 5.80 ng/ml




#30 AR-1254-5

R.T.: 7.677 min
Delta R.T.: 0.020 min
Response: 3080536
Conc: 30.45 ng/ml



#30 AR-1254-5


R.T.: 6.666 min
Delta R.T.: -0.003 min
Response: 628668
Conc: 6.63 ng/ml



#31 AR-1260-1


R.T.: 7.113 min
Delta R.T.: -0.003 min
Response: 594264
Conc: 6.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.150 min
Delta R.T.: -0.002 min
Response: 557468
Conc: 7.19 ng/ml




#32 AR-1260-2

R.T.: 7.375 min
Delta R.T.: 0.001 min
Response: 1529009
Conc: 13.59 ng/ml



#32 AR-1260-2


R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 470331
Conc: 5.23 ng/ml



#33 AR-1260-3


R.T.: 7.733 min
Delta R.T.: -0.002 min
Response: 37242
Conc: 0.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.496 min
Delta R.T.: 0.001 min
Response: 291196
Conc: 3.43 ng/ml




#34 AR-1260-4

R.T.: 7.966 min
Delta R.T.: 0.005 min
Response: 453592
Conc: 4.67 ng/ml



#34 AR-1260-4


R.T.: 6.969 min
Delta R.T.: 0.000 min
Response: 46236
Conc: 0.65 ng/ml



#35 AR-1260-5


R.T.: 8.277 min
Delta R.T.: 0.000 min
Response: 821103
Conc: 4.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.208 min
Delta R.T.: -0.004 min
Response: 567254
Conc: 3.69 ng/ml




#36 AR-1262-1

R.T.: 7.733 min
Delta R.T.: 0.000 min
Response: 37242
Conc: 0.32 ng/ml



#36 AR-1262-1


R.T.: 6.762 min
Delta R.T.: 0.002 min
Response: 33028
Conc: 0.77 ng/ml



#37 AR-1262-2


R.T.: 8.277 min
Delta R.T.: 0.000 min
Response: 821103
Conc: 4.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.969 min
Delta R.T.: 0.001 min
Response: 46236
Conc: 0.50 ng/ml




#38 AR-1262-3

R.T.: 8.568 min
Delta R.T.: -0.021 min
Response: 56443
Conc: 0.41 ng/ml



#38 AR-1262-3


R.T.: 7.489 min
Delta R.T.: -0.006 min
Response: 1085350
Conc: 14.90 ng/ml



#39 AR-1262-4


R.T.: 8.665 min
Delta R.T.: -0.013 min
Response: 628198
Conc: 8.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.552 min
Delta R.T.: -0.007 min
Response: 573245
Conc: 4.38 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 9.334 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 8.052 min
Delta R.T.: -0.002 min
Response: 5269
Conc: 0.09 ng/ml



#41 AR-1268-1


R.T.: 8.568 min
Delta R.T.: -0.017 min
Response: 56443
Conc: 0.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.489 min
Delta R.T.: -0.005 min
Response: 1085350
Conc: 5.02 ng/ml




#42 AR-1268-2

R.T.: 8.691 min
Delta R.T.: 0.012 min
Response: 243224
Conc: 1.07 ng/ml



#42 AR-1268-2


R.T.: 7.552 min
Delta R.T.: -0.007 min
Response: 573245
Conc: 2.94 ng/ml



#43 AR-1268-3


R.T.: 8.903 min
Delta R.T.: -0.004 min
Response: 533657
Conc: 2.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.773 min
Delta R.T.: 0.006 min
Response: 103590
Conc: 0.60 ng/ml




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T. : 9.333 min
Response: 0
Conc: N.D.



#44 AR-1268-4


R.T.: 8.052 min
Delta R.T.: -0.002 min
Response: 5269
Conc: 0.08 ng/ml



#45 AR-1268-5

R.T.: 9.743 min
Delta R.T.: -0.006 min
Response: 267573
Conc: 0.42 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.343 min
Delta R.T.: 0.000 min
Response: 181944
Conc: 0.35 ng/ml