

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012523\
 Data File : PP054875.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2023 15:20
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 17:18:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.332	3.577	87360609	76044802	50.244	50.656
2)	SA Decachlor...	10.087	8.593	93998667	74526765	54.282	55.182
Target Compounds							
26)	L6 AR-1254-1	6.362	5.466	37676586	44057281	477.569	545.546
27)	L6 AR-1254-2	6.582	5.614	61175131	37707633	505.483	526.918
28)	L6 AR-1254-3	6.949	6.019	63975114	57729780	522.288	518.185
29)	L6 AR-1254-4	7.236	6.248	50111131	31597600	568.721	513.720
30)	L6 AR-1254-5	7.656	6.667	52234335	49158891	516.349	518.532



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012523\
Data File : PP054875.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2023 15:20
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 24 17:18:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.332 min
Delta R.T.: 0.000 min
Response: 87360609
Conc: 50.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.577 min
Delta R.T.: 0.000 min
Response: 76044802
Conc: 50.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.087 min
Delta R.T.: -0.004 min
Response: 93998667
Conc: 54.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.593 min
Delta R.T.: -0.005 min
Response: 74526765
Conc: 55.18 ng/ml



#26 AR-1254-1


R.T.: 6.362 min
Delta R.T.: 0.001 min
Response: 37676586
Conc: 477.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.466 min
Delta R.T.: 0.000 min
Response: 44057281
Conc: 545.55 ng/ml




#27 AR-1254-2

R.T.: 6.582 min
Delta R.T.: 0.000 min
Response: 61175131
Conc: 505.48 ng/ml



#27 AR-1254-2


R.T.: 5.614 min
Delta R.T.: -0.002 min
Response: 37707633
Conc: 526.92 ng/ml



#28 AR-1254-3


R.T.: 6.949 min
Delta R.T.: 0.000 min
Response: 63975114
Conc: 522.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.019 min
Delta R.T.: 0.000 min
Response: 57729780
Conc: 518.19 ng/ml




#29 AR-1254-4

R.T.: 7.236 min
Delta R.T.: 0.000 min
Response: 50111131
Conc: 568.72 ng/ml



#29 AR-1254-4


R.T.: 6.248 min
Delta R.T.: -0.002 min
Response: 31597600
Conc: 513.72 ng/ml



#30 AR-1254-5

R.T.: 7.656 min
Delta R.T.: 0.000 min
Response: 52234335
Conc: 516.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.667 min
Delta R.T.: -0.002 min
Response: 49158891
Conc: 518.53 ng/ml