

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012721\
 Data File : PP033003.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Jan 2021 20:16
 Operator : DD\AJ
 Sample : PB134369BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 01:22:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 14 08:12:55 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.734	4.008	1191078	965079	20.507	20.516
2)	SA Decachlor...	10.564	9.281f	881453	844032	22.342	22.530
Target Compounds							
34)	L7 AR-1260-4	8.561f	0.000	29214	0	14.178	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012721\
Data File : PP033003.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Jan 2021 20:16
Operator : DD\AJ
Sample : PB134369BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 01:22:16 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011421.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 14 08:12:55 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.734 min
Delta R.T.: -0.007 min
Response: 1191078
Conc: 20.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.008 min
Delta R.T.: -0.005 min
Response: 965079
Conc: 20.52 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.564 min
Delta R.T.: -0.013 min
Response: 881453
Conc: 22.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.281 min
Delta R.T.: -0.015 min
Response: 844032
Conc: 22.53 ng/ml



#34 AR-1260-4

R.T.: 8.561 min
Delta R.T.: -0.017 min
Response: 29214
Conc: 14.18 ng/ml

Instrument :
ECD_P
ClientSampleId :



#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 7.680 min
Response: 0
Conc: N.D.