

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012723\
 Data File : PP054952.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jan 2023 09:38
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 26 20:39:50 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.332	3.578	82806994	72204669	47.625	48.098
2) SA Decachlor...	10.095	8.597	90274598	70243085	52.131	52.010
Target Compounds						
3) L1 AR-1016-1	5.504	4.663	31772445	24130520	472.733	487.545
4) L1 AR-1016-2	5.527	4.682	45231502	33804014	472.001	488.363
5) L1 AR-1016-3	5.589	4.858	28589407	18397366	477.953	494.598
6) L1 AR-1016-4	5.688	4.900	23144221	15301370	482.996	486.960
7) L1 AR-1016-5	5.984	5.114	24120552	21033755	475.993	514.468
31) L7 AR-1260-1	7.117	6.151	47576312	39904896	499.088	514.606
32) L7 AR-1260-2	7.375	6.340	58798019	45900572	522.434	510.585
33) L7 AR-1260-3	7.737	6.495	43049132	43129775	534.646	507.943
34) L7 AR-1260-4	7.963	6.968	48713991	36583761	501.809	510.927
35) L7 AR-1260-5	8.280	7.212	95902881	79022449	539.233	514.013



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012723\
Data File : PP054952.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jan 2023 09:38
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 26 20:39:50 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.332 min
Delta R.T.: 0.000 min
Response: 82806994
Conc: 47.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.578 min
Delta R.T.: 0.000 min
Response: 72204669
Conc: 48.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.095 min
Delta R.T.: 0.005 min
Response: 90274598
Conc: 52.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.597 min
Delta R.T.: -0.001 min
Response: 70243085
Conc: 52.01 ng/ml



#3 AR-1016-1


R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 31772445
Conc: 472.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.663 min
Delta R.T.: 0.000 min
Response: 24130520
Conc: 487.54 ng/ml




#4 AR-1016-2

R.T.: 5.527 min
Delta R.T.: 0.000 min
Response: 45231502
Conc: 472.00 ng/ml



#4 AR-1016-2


R.T.: 4.682 min
Delta R.T.: 0.000 min
Response: 33804014
Conc: 488.36 ng/ml



#5 AR-1016-3


R.T.: 5.589 min
Delta R.T.: 0.000 min
Response: 28589407
Conc: 477.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.858 min
Delta R.T.: -0.001 min
Response: 18397366
Conc: 494.60 ng/ml




#6 AR-1016-4

R.T.: 5.688 min
Delta R.T.: 0.000 min
Response: 23144221
Conc: 483.00 ng/ml



#6 AR-1016-4


R.T.: 4.900 min
Delta R.T.: 0.000 min
Response: 15301370
Conc: 486.96 ng/ml



#7 AR-1016-5


R.T.: 5.984 min
Delta R.T.: 0.000 min
Response: 24120552
Conc: 475.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.114 min
Delta R.T.: 0.000 min
Response: 21033755
Conc: 514.47 ng/ml




#31 AR-1260-1

R.T.: 7.117 min
Delta R.T.: 0.000 min
Response: 47576312
Conc: 499.09 ng/ml



#31 AR-1260-1


R.T.: 6.151 min
Delta R.T.: -0.001 min
Response: 39904896
Conc: 514.61 ng/ml



#32 AR-1260-2


R.T.: 7.375 min
Delta R.T.: 0.001 min
Response: 58798019
Conc: 522.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.340 min
Delta R.T.: -0.001 min
Response: 45900572
Conc: 510.58 ng/ml




#33 AR-1260-3

R.T.: 7.737 min
Delta R.T.: 0.002 min
Response: 43049132
Conc: 534.65 ng/ml



#33 AR-1260-3


R.T.: 6.495 min
Delta R.T.: 0.000 min
Response: 43129775
Conc: 507.94 ng/ml



#34 AR-1260-4


R.T.: 7.963 min
Delta R.T.: 0.001 min
Response: 48713991
Conc: 501.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.968 min
Delta R.T.: 0.000 min
Response: 36583761
Conc: 510.93 ng/ml



#35 AR-1260-5

R.T.: 8.280 min
Delta R.T.: 0.003 min
Response: 95902881
Conc: 539.23 ng/ml



#35 AR-1260-5

R.T.: 7.212 min
Delta R.T.: 0.000 min
Response: 79022449
Conc: 514.01 ng/ml