

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012821\
 Data File : PP033002.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jan 2021 11:51
 Operator : DD\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 12:50:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 28 12:49:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.736	4.007	2910855	2342209	50.000	50.000
2) SA Decachlor...	10.567	9.285	1841853	1660575	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	6.036	5.256	789133	653281	500.000	500.000
4) L1 AR-1016-2	6.059	5.275	1205552	957713	500.000	500.000
5) L1 AR-1016-3	6.125	5.466	750738	527799	500.000	500.000
6) L1 AR-1016-4	6.230	5.517	630200	405374	500.000	500.000
7) L1 AR-1016-5	6.543	5.745	577806	528162	500.000	500.000
31) L7 AR-1260-1	7.713	6.837	946334	909090	500.000	500.000
32) L7 AR-1260-2	7.978	7.035	1241378	1052552	500.000	500.000
33) L7 AR-1260-3	8.343	7.189	1088165	1056704	500.000	500.000
34) L7 AR-1260-4	8.571	7.670	1059944	857831	500.000	500.000
35) L7 AR-1260-5	8.884	7.920	2285202	1921226	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012821\
Data File : PP033002.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Jan 2021 11:51
Operator : DD\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 12:50:18 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 28 12:49:30 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.736 min
Delta R.T.: 0.000 min
Response: 2910855
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 4.007 min
Delta R.T.: 0.000 min
Response: 2342209
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.567 min
Delta R.T.: 0.000 min
Response: 1841853
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.285 min
Delta R.T.: 0.000 min
Response: 1660575
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 6.036 min
Delta R.T.: 0.000 min
Response: 789133
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 653281
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 6.059 min
Delta R.T.: 0.000 min
Response: 1205552
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 5.275 min
Delta R.T.: 0.000 min
Response: 957713
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 6.125 min
Delta R.T.: 0.000 min
Response: 750738
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 5.466 min
Delta R.T.: 0.000 min
Response: 527799
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 6.230 min
Delta R.T.: 0.000 min
Response: 630200
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 5.517 min
Delta R.T.: 0.000 min
Response: 405374
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 577806
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.745 min
Delta R.T.: 0.000 min
Response: 528162
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.713 min
Delta R.T.: 0.000 min
Response: 946334
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.837 min
Delta R.T.: 0.000 min
Response: 909090
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.978 min
Delta R.T.: 0.000 min
Response: 1241378
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 7.035 min
Delta R.T.: 0.000 min
Response: 1052552
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 8.343 min
Delta R.T.: 0.000 min
Response: 1088165
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 7.189 min
Delta R.T.: 0.000 min
Response: 1056704
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 8.571 min
Delta R.T.: 0.000 min
Response: 1059944
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 7.670 min
Delta R.T.: 0.000 min
Response: 857831
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.884 min
Delta R.T.: 0.000 min
Response: 2285202
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.920 min
Delta R.T.: 0.000 min
Response: 1921226
Conc: 500.00 ng/ml