

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012823\
 Data File : PP055038.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2023 03:52
 Operator : YP\AJ
 Sample : 01350-06
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 05:13:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.329	3.576	33309308	29896681	19.157	19.915
2)	SA Decachlor...	10.083	8.589	36949565	29497211	21.337	21.841
Target Compounds							
3)	L1 AR-1016-1	5.499	4.673	225905	341043	3.361	6.891 #
4)	L1 AR-1016-2	5.527	4.673	387348	341043	4.042	4.927
5)	L1 AR-1016-3	5.580	4.856	347710	153916	5.813	4.138 #
6)	L1 AR-1016-4	5.681	4.897	-5820	3219611	N.D.	102.463 #
7)	L1 AR-1016-5	5.979	5.110	2202775	1865628	43.469	45.632
8)	L2 AR-1221-1	4.537	3.785	156405	58174	6.651	2.934 #
9)	L2 AR-1221-2	4.636	3.873	816689	1001625	45.967	67.575 #
10)	L2 AR-1221-3	4.717	0.000	832145	0	15.465	N.D. #
11)	L3 AR-1232-1	4.717	0.000	832145	0	18.749	N.D. #
12)	L3 AR-1232-2	5.257f	4.673	183279	341043	8.104	10.677 #
13)	L3 AR-1232-3	5.527	4.856	387348	153916	8.820	9.233
14)	L3 AR-1232-4	5.681	4.937	-5820	1124984	N.D.	69.762 #
15)	L3 AR-1232-5	5.775	5.110	3890817	1865628	222.243	105.633 #
16)	L4 AR-1242-1	5.499	4.673	225905	341043	4.035	8.326 #
17)	L4 AR-1242-2	5.527	4.673	387348	341043	4.852	5.973
18)	L4 AR-1242-3	5.580	4.856	347710	153916	7.002	4.986 #
19)	L4 AR-1242-4	5.681	4.937	-5820	1124984	N.D.	34.629 #
20)	L4 AR-1242-5	6.426	5.463	3653336	9380793	82.165	251.013 #
21)	L5 AR-1248-1	5.499	4.673	225905	341043	5.417	11.000 #
22)	L5 AR-1248-2	5.775	4.897	3890817	3219611	61.201	70.409
23)	L5 AR-1248-3	5.979	4.937	2202775	1124984	31.851	23.598 #
24)	L5 AR-1248-4	6.384	5.110	3948549	1865628	52.068	33.781 #
25)	L5 AR-1248-5	6.426	5.503	3653336	1502548	49.508	30.558 #
26)	L6 AR-1254-1	6.358	5.463	9109796	9380793	115.471	116.159
27)	L6 AR-1254-2	6.578	5.611	14929478	9267935	123.361	129.508
28)	L6 AR-1254-3	6.945	6.016	17377227	16148587	141.867	144.951
29)	L6 AR-1254-4	7.233	6.245	14861221	10513358	168.663	170.928
30)	L6 AR-1254-5	7.654	6.665	17299106	16983253	171.006	179.141
31)	L7 AR-1260-1	7.111	6.147	8310413	8475043	87.178	109.293 #
32)	L7 AR-1260-2	7.369	6.335	10905883	8090047	96.901	89.991
33)	L7 AR-1260-3	7.713f	6.488	2697064	7203544	33.496	84.837 #
34)	L7 AR-1260-4	7.943f	6.963	8622323	1367090	88.820	19.093 #
35)	L7 AR-1260-5	8.273	7.204	4964083	3256510	27.912	21.182
36)	L8 AR-1262-1	7.713f	6.755	2697064	855278	22.860	19.908
37)	L8 AR-1262-2	8.273	6.963	4964083	1367090	24.646	14.760 #
38)	L8 AR-1262-3	8.597	7.486	2174898	1025945	15.639	14.081
39)	L8 AR-1262-4	8.664	7.552	841875	3563547	10.950	27.249 #
40)	L8 AR-1262-5	9.327	8.049	442365	405202	6.185	6.830
41)	L9 AR-1268-1	8.597	7.486	2174898	1025945	8.899	4.745 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012823\
 Data File : PP055038.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2023 03:52
 Operator : YP\AJ
 Sample : 01350-06
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 05:13:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.664	7.552	841875	3563547	3.707	18.262 #
43) L9	AR-1268-3	8.902	7.758	260686	47600	1.305	0.277 #
44) L9	AR-1268-4	9.327	8.049	442365	405202	5.632	6.353
45) L9	AR-1268-5	9.751	8.338	187836	262109	0.295	0.498 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012823\
Data File : PP055038.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jan 2023 03:52
Operator : YP\AJ
Sample : 01350-06
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 05:13:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.329 min
Delta R.T.: -0.003 min
Response: 33309308
Conc: 19.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.576 min
Delta R.T.: -0.003 min
Response: 29896681
Conc: 19.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.083 min
Delta R.T.: -0.008 min
Response: 36949565
Conc: 21.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.589 min
Delta R.T.: -0.010 min
Response: 29497211
Conc: 21.84 ng/ml



#3 AR-1016-1


R.T.: 5.499 min
Delta R.T.: -0.005 min
Response: 225905
Conc: 3.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.673 min
Delta R.T.: 0.009 min
Response: 341043
Conc: 6.89 ng/ml




#4 AR-1016-2

R.T.: 5.527 min
Delta R.T.: 0.001 min
Response: 387348
Conc: 4.04 ng/ml



#4 AR-1016-2


R.T.: 4.673 min
Delta R.T.: -0.009 min
Response: 341043
Conc: 4.93 ng/ml



#5 AR-1016-3


R.T.: 5.580 min
Delta R.T.: -0.008 min
Response: 347710
Conc: 5.81 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.856 min
Delta R.T.: -0.003 min
Response: 153916
Conc: 4.14 ng/ml




#6 AR-1016-4

R.T.: 5.681 min
Delta R.T.: -0.006 min
Response: -5820
Conc: N.D.



#6 AR-1016-4


R.T.: 4.897 min
Delta R.T.: -0.004 min
Response: 3219611
Conc: 102.46 ng/ml



#7 AR-1016-5


R.T.: 5.979 min
Delta R.T.: -0.004 min
Response: 2202775
Conc: 43.47 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.110 min
Delta R.T.: -0.005 min
Response: 1865628
Conc: 45.63 ng/ml




#8 AR-1221-1

R.T.: 4.537 min
Delta R.T.: -0.002 min
Response: 156405
Conc: 6.65 ng/ml



#8 AR-1221-1


R.T.: 3.785 min
Delta R.T.: -0.006 min
Response: 58174
Conc: 2.93 ng/ml



#9 AR-1221-2


R.T.: 4.636 min
Delta R.T.: 0.010 min
Response: 816689
Conc: 45.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.873 min
Delta R.T.: -0.003 min
Response: 1001625
Conc: 67.58 ng/ml




#10 AR-1221-3

R.T.: 4.717 min
Delta R.T.: 0.015 min
Response: 832145
Conc: 15.46 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 3.952 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.717 min
Delta R.T.: 0.013 min
Response: 832145
Conc: 18.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.953 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.257 min
Delta R.T.: 0.022 min
Response: 183279
Conc: 8.10 ng/ml



#12 AR-1232-2


R.T.: 4.673 min
Delta R.T.: -0.010 min
Response: 341043
Conc: 10.68 ng/ml



#13 AR-1232-3


R.T.: 5.527 min
Delta R.T.: 0.001 min
Response: 387348
Conc: 8.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.856 min
Delta R.T.: -0.003 min
Response: 153916
Conc: 9.23 ng/ml




#14 AR-1232-4

R.T.: 5.681 min
Delta R.T.: -0.006 min
Response: -5820
Conc: N.D.



#14 AR-1232-4


R.T.: 4.937 min
Delta R.T.: -0.006 min
Response: 1124984
Conc: 69.76 ng/ml



#15 AR-1232-5


R.T.: 5.775 min
Delta R.T.: -0.005 min
Response: 3890817
Conc: 222.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.110 min
Delta R.T.: -0.006 min
Response: 1865628
Conc: 105.63 ng/ml




#16 AR-1242-1

R.T.: 5.499 min
Delta R.T.: -0.005 min
Response: 225905
Conc: 4.03 ng/ml



#16 AR-1242-1


R.T.: 4.673 min
Delta R.T.: 0.009 min
Response: 341043
Conc: 8.33 ng/ml



#17 AR-1242-2


R.T.: 5.527 min
Delta R.T.: 0.001 min
Response: 387348
Conc: 4.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.673 min
Delta R.T.: -0.009 min
Response: 341043
Conc: 5.97 ng/ml




#18 AR-1242-3

R.T.: 5.580 min
Delta R.T.: -0.008 min
Response: 347710
Conc: 7.00 ng/ml



#18 AR-1242-3


R.T.: 4.856 min
Delta R.T.: -0.003 min
Response: 153916
Conc: 4.99 ng/ml



#19 AR-1242-4


R.T.: 5.681 min
Delta R.T.: -0.005 min
Response: -5820
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.937 min
Delta R.T.: -0.006 min
Response: 1124984
Conc: 34.63 ng/ml




#20 AR-1242-5

R.T.: 6.426 min
Delta R.T.: -0.001 min
Response: 3653336
Conc: 82.16 ng/ml



#20 AR-1242-5


R.T.: 5.463 min
Delta R.T.: -0.005 min
Response: 9380793
Conc: 251.01 ng/ml



#21 AR-1248-1


R.T.: 5.499 min
Delta R.T.: -0.003 min
Response: 225905
Conc: 5.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.673 min
Delta R.T.: 0.010 min
Response: 341043
Conc: 11.00 ng/ml




#22 AR-1248-2

R.T.: 5.775 min
Delta R.T.: -0.003 min
Response: 3890817
Conc: 61.20 ng/ml



#22 AR-1248-2


R.T.: 4.897 min
Delta R.T.: -0.004 min
Response: 3219611
Conc: 70.41 ng/ml



#23 AR-1248-3


R.T.: 5.979 min
Delta R.T.: -0.003 min
Response: 2202775
Conc: 31.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.937 min
Delta R.T.: -0.005 min
Response: 1124984
Conc: 23.60 ng/ml




#24 AR-1248-4

R.T.: 6.384 min
Delta R.T.: -0.003 min
Response: 3948549
Conc: 52.07 ng/ml



#24 AR-1248-4


R.T.: 5.110 min
Delta R.T.: -0.004 min
Response: 1865628
Conc: 33.78 ng/ml



#25 AR-1248-5


R.T.: 6.426 min
Delta R.T.: 0.000 min
Response: 3653336
Conc: 49.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.503 min
Delta R.T.: -0.005 min
Response: 1502548
Conc: 30.56 ng/ml




#26 AR-1254-1

R.T.: 6.358 min
Delta R.T.: -0.003 min
Response: 9109796
Conc: 115.47 ng/ml



#26 AR-1254-1


R.T.: 5.463 min
Delta R.T.: -0.004 min
Response: 9380793
Conc: 116.16 ng/ml



#27 AR-1254-2


R.T.: 6.578 min
Delta R.T.: -0.003 min
Response: 14929478
Conc: 123.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.611 min
Delta R.T.: -0.004 min
Response: 9267935
Conc: 129.51 ng/ml




#28 AR-1254-3

R.T.: 6.945 min
Delta R.T.: -0.003 min
Response: 17377227
Conc: 141.87 ng/ml



#28 AR-1254-3


R.T.: 6.016 min
Delta R.T.: -0.004 min
Response: 16148587
Conc: 144.95 ng/ml



#29 AR-1254-4


R.T.: 7.233 min
Delta R.T.: -0.004 min
Response: 14861221
Conc: 168.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.245 min
Delta R.T.: -0.004 min
Response: 10513358
Conc: 170.93 ng/ml




#30 AR-1254-5

R.T.: 7.654 min
Delta R.T.: -0.003 min
Response: 17299106
Conc: 171.01 ng/ml



#30 AR-1254-5


R.T.: 6.665 min
Delta R.T.: -0.004 min
Response: 16983253
Conc: 179.14 ng/ml



#31 AR-1260-1


R.T.: 7.111 min
Delta R.T.: -0.005 min
Response: 8310413
Conc: 87.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.147 min
Delta R.T.: -0.005 min
Response: 8475043
Conc: 109.29 ng/ml




#32 AR-1260-2

R.T.: 7.369 min
Delta R.T.: -0.005 min
Response: 10905883
Conc: 96.90 ng/ml



#32 AR-1260-2


R.T.: 6.335 min
Delta R.T.: -0.006 min
Response: 8090047
Conc: 89.99 ng/ml



#33 AR-1260-3


R.T.: 7.713 min
Delta R.T.: -0.021 min
Response: 2697064
Conc: 33.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.488 min
Delta R.T.: -0.007 min
Response: 7203544
Conc: 84.84 ng/ml




#34 AR-1260-4

R.T.: 7.943 min
Delta R.T.: -0.018 min
Response: 8622323
Conc: 88.82 ng/ml



#34 AR-1260-4


R.T.: 6.963 min
Delta R.T.: -0.006 min
Response: 1367090
Conc: 19.09 ng/ml



#35 AR-1260-5


R.T.: 8.273 min
Delta R.T.: -0.005 min
Response: 4964083
Conc: 27.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.204 min
Delta R.T.: -0.008 min
Response: 3256510
Conc: 21.18 ng/ml




#36 AR-1262-1

R.T.: 7.713 min
Delta R.T.: -0.021 min
Response: 2697064
Conc: 22.86 ng/ml



#36 AR-1262-1


R.T.: 6.755 min
Delta R.T.: -0.006 min
Response: 855278
Conc: 19.91 ng/ml



#37 AR-1262-2


R.T.: 8.273 min
Delta R.T.: -0.004 min
Response: 4964083
Conc: 24.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.963 min
Delta R.T.: -0.004 min
Response: 1367090
Conc: 14.76 ng/ml




#38 AR-1262-3

R.T.: 8.597 min
Delta R.T.: 0.008 min
Response: 2174898
Conc: 15.64 ng/ml



#38 AR-1262-3


R.T.: 7.486 min
Delta R.T.: -0.009 min
Response: 1025945
Conc: 14.08 ng/ml



#39 AR-1262-4


R.T.: 8.664 min
Delta R.T.: -0.013 min
Response: 841875
Conc: 10.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.552 min
Delta R.T.: -0.006 min
Response: 3563547
Conc: 27.25 ng/ml




#40 AR-1262-5

R.T.: 9.327 min
Delta R.T.: -0.007 min
Response: 442365
Conc: 6.19 ng/ml



#40 AR-1262-5


R.T.: 8.049 min
Delta R.T.: -0.005 min
Response: 405202
Conc: 6.83 ng/ml



#41 AR-1268-1


R.T.: 8.597 min
Delta R.T.: 0.012 min
Response: 2174898
Conc: 8.90 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.486 min
Delta R.T.: -0.008 min
Response: 1025945
Conc: 4.74 ng/ml




#42 AR-1268-2

R.T.: 8.664 min
Delta R.T.: -0.015 min
Response: 841875
Conc: 3.71 ng/ml



#42 AR-1268-2


R.T.: 7.552 min
Delta R.T.: -0.006 min
Response: 3563547
Conc: 18.26 ng/ml



#43 AR-1268-3


R.T.: 8.902 min
Delta R.T.: -0.006 min
Response: 260686
Conc: 1.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.758 min
Delta R.T.: -0.008 min
Response: 47600
Conc: 0.28 ng/ml




#44 AR-1268-4

R.T.: 9.327 min
Delta R.T.: -0.006 min
Response: 442365
Conc: 5.63 ng/ml



#44 AR-1268-4


R.T.: 8.049 min
Delta R.T.: -0.005 min
Response: 405202
Conc: 6.35 ng/ml



#45 AR-1268-5

R.T.: 9.751 min
Delta R.T.: 0.002 min
Response: 187836
Conc: 0.30 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.338 min
Delta R.T.: -0.005 min
Response: 262109
Conc: 0.50 ng/ml