

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012823\
 Data File : PP055020.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jan 2023 22:57
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 05:00:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.329	3.576	82383325	73760054	47.382	49.134
2) SA Decachlor...	10.083	8.589	90934191	71220250	52.512	52.734
Target Compounds						
26) L6 AR-1254-1	6.359	5.463	36577489	40481917	463.638	501.274
27) L6 AR-1254-2	6.579	5.612	57956298	35486207	478.887	495.876
28) L6 AR-1254-3	6.946	6.016	60120445	56578538	490.819	507.852
29) L6 AR-1254-4	7.233	6.245	44439274	31336405	504.350	509.474
30) L6 AR-1254-5	7.654	6.665	49159938	48327551	485.958	509.763



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012823\
Data File : PP055020.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jan 2023 22:57
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 05:00:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.329 min
Delta R.T.: -0.003 min
Response: 82383325
Conc: 47.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.576 min
Delta R.T.: -0.002 min
Response: 73760054
Conc: 49.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.083 min
Delta R.T.: -0.008 min
Response: 90934191
Conc: 52.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.589 min
Delta R.T.: -0.009 min
Response: 71220250
Conc: 52.73 ng/ml



#26 AR-1254-1


R.T.: 6.359 min
Delta R.T.: -0.003 min
Response: 36577489
Conc: 463.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.463 min
Delta R.T.: -0.004 min
Response: 40481917
Conc: 501.27 ng/ml




#27 AR-1254-2

R.T.: 6.579 min
Delta R.T.: -0.002 min
Response: 57956298
Conc: 478.89 ng/ml



#27 AR-1254-2


R.T.: 5.612 min
Delta R.T.: -0.004 min
Response: 35486207
Conc: 495.88 ng/ml



#28 AR-1254-3


R.T.: 6.946 min
Delta R.T.: -0.003 min
Response: 60120445
Conc: 490.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.016 min
Delta R.T.: -0.004 min
Response: 56578538
Conc: 507.85 ng/ml




#29 AR-1254-4

R.T.: 7.233 min
Delta R.T.: -0.003 min
Response: 44439274
Conc: 504.35 ng/ml



#29 AR-1254-4


R.T.: 6.245 min
Delta R.T.: -0.004 min
Response: 31336405
Conc: 509.47 ng/ml



#30 AR-1254-5

R.T.: 7.654 min
Delta R.T.: -0.003 min
Response: 49159938
Conc: 485.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.665 min
Delta R.T.: -0.004 min
Response: 48327551
Conc: 509.76 ng/ml