

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012823\
 Data File : PP055028.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2023 01:08
 Operator : YP\AJ
 Sample : 01328-21
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 COMP-5-BORING-2-0-7

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 05:06:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.329	3.575	32737074	28940846	18.828	19.278
2) SA Decachlor...	10.081	8.588	31308361	23640323	18.080	17.504

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012823\
Data File : PP055028.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jan 2023 01:08
Operator : YP\AJ
Sample : 01328-21
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
COMP-5-BORING-2-0-7

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 05:06:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.329 min
Delta R.T.: -0.003 min
Response: 32737074
Conc: 18.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
COMP-5-BORING-2-0-7




#1 Tetrachloro-m-xylene

R.T.: 3.575 min
Delta R.T.: -0.003 min
Response: 28940846
Conc: 19.28 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.081 min
Delta R.T.: -0.010 min
Response: 31308361
Conc: 18.08 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.588 min
Delta R.T.: -0.010 min
Response: 23640323
Conc: 17.50 ng/ml