

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012921\
 Data File : PP033071.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Jan 2021 15:05
 Operator : DD\AJ
 Sample : AR1262CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1262CCC500

Manual Integrations
 APPROVED

Ankita
 2/3/2021 4:32:59 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 30 00:24:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 04:07:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.734	4.004	2404743	2024991	42.350	44.050
2) SA Decachlor...	10.562	9.277	1848505	1639511	51.522	49.894
Target Compounds						
36) L8 AR-1262-1	8.340	7.363	1509548	601138	487.692	508.842
37) L8 AR-1262-2	8.881	7.914	2516184	2026291	527.839	519.073
38) L8 AR-1262-3	9.174	8.198	1724722	843562	512.725	488.978
39) L8 AR-1262-4	9.261	8.263	1326374	1567504	477.415m	500.479
40) L8 AR-1262-5	9.877	8.760	883044	717600	563.356	556.785

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012921\
Data File : PP033071.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Jan 2021 15:05
Operator : DD\AJ
Sample : AR1262CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampled :
AR1262CCC500


Manual Integrations
APPROVED

Ankita
2/3/2021 4:32:59 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 00:24:59 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 29 04:07:51 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.734 min
Delta R.T.: 0.000 min
Response: 2404743
Conc: 42.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262CCC500


Manual Integrations
APPROVED

Ankita
2/3/2021 4:32:59 PM




#1 Tetrachloro-m-xylene

R.T.: 4.004 min
Delta R.T.: -0.003 min
Response: 2024991
Conc: 44.05 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.562 min
Delta R.T.: -0.003 min
Response: 1848505
Conc: 51.52 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.277 min
Delta R.T.: -0.006 min
Response: 1639511
Conc: 49.89 ng/ml




#36 AR-1262-1

R.T.: 8.340 min
Delta R.T.: 0.000 min
Response: 1509548
Conc: 487.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262CCC500


Manual Integrations
APPROVED

Ankita
2/3/2021 4:32:59 PM




#36 AR-1262-1

R.T.: 7.363 min
Delta R.T.: -0.003 min
Response: 601138
Conc: 508.84 ng/ml




#37 AR-1262-2

R.T.: 8.881 min
Delta R.T.: 0.000 min
Response: 2516184
Conc: 527.84 ng/ml



#37 AR-1262-2

R.T.: 7.914 min
Delta R.T.: -0.003 min
Response: 2026291
Conc: 519.07 ng/ml




#38 AR-1262-3

R.T.: 9.174 min
Delta R.T.: 0.000 min
Response: 1724722
Conc: 512.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262CCC500


Manual Integrations
APPROVED

Ankita
2/3/2021 4:32:59 PM




#38 AR-1262-3

R.T.: 8.198 min
Delta R.T.: -0.003 min
Response: 843562
Conc: 488.98 ng/ml




#39 AR-1262-4

R.T.: 9.261 min
Delta R.T.: -0.003 min
Response: 1326374
Conc: 477.42 ng/ml m



#39 AR-1262-4

R.T.: 8.263 min
Delta R.T.: -0.003 min
Response: 1567504
Conc: 500.48 ng/ml




#40 AR-1262-5

R.T.: 9.877 min
Delta R.T.: -0.002 min
Response: 883044
Conc: 563.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262CCC500

Manual Integrations
APPROVED

Ankita
2/3/2021 4:32:59 PM



#40 AR-1262-5

R.T.: 8.760 min
Delta R.T.: -0.004 min
Response: 717600
Conc: 556.78 ng/ml