

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012921\
 Data File : PP033102.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Jan 2021 00:15
 Operator : DD\AJ
 Sample : M1254-01
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 CRT-15

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 30 03:52:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 04:07:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.731	4.005	1061574	1019582	18.695	22.179
2)	SA Decachlor...	10.557	9.274	686178	3055165	19.125	92.976 #
Target Compounds							
26)	L6 AR-1254-1	6.933	6.120	883214	931437	524.205	464.723
27)	L6 AR-1254-2	7.163	6.278	1455894	752119	563.851	440.213
28)	L6 AR-1254-3	7.541	6.698	1572611	1664419	601.606	612.206
29)	L6 AR-1254-4	7.834	6.937	1048800	811999	594.440	569.249
30)	L6 AR-1254-5	8.261	7.363	1463329	1532370	739.142	678.469



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012921\
Data File : PP033102.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Jan 2021 00:15
Operator : DD\AJ
Sample : M1254-01
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
CRT-15

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 03:52:38 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 29 04:07:51 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.731 min
Delta R.T.: -0.004 min
Response: 1061574
Conc: 18.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
CRT-15




#1 Tetrachloro-m-xylene

R.T.: 4.005 min
Delta R.T.: -0.001 min
Response: 1019582
Conc: 22.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.557 min
Delta R.T.: -0.009 min
Response: 686178
Conc: 19.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.274 min
Delta R.T.: -0.010 min
Response: 3055165
Conc: 92.98 ng/ml



#26 AR-1254-1


R.T.: 6.933 min
Delta R.T.: -0.005 min
Response: 883214
Conc: 524.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
CRT-15




#26 AR-1254-1

R.T.: 6.120 min
Delta R.T.: -0.003 min
Response: 931437
Conc: 464.72 ng/ml




#27 AR-1254-2

R.T.: 7.163 min
Delta R.T.: -0.004 min
Response: 1455894
Conc: 563.85 ng/ml



#27 AR-1254-2


R.T.: 6.278 min
Delta R.T.: -0.003 min
Response: 752119
Conc: 440.21 ng/ml



#28 AR-1254-3


R.T.: 7.541 min
Delta R.T.: -0.004 min
Response: 1572611
Conc: 601.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
CRT-15




#28 AR-1254-3

R.T.: 6.698 min
Delta R.T.: -0.003 min
Response: 1664419
Conc: 612.21 ng/ml




#29 AR-1254-4

R.T.: 7.834 min
Delta R.T.: -0.005 min
Response: 1048800
Conc: 594.44 ng/ml



#29 AR-1254-4


R.T.: 6.937 min
Delta R.T.: -0.003 min
Response: 811999
Conc: 569.25 ng/ml



#30 AR-1254-5

R.T.: 8.261 min
Delta R.T.: -0.005 min
Response: 1463329
Conc: 739.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
CRT-15



#30 AR-1254-5

R.T.: 7.363 min
Delta R.T.: -0.004 min
Response: 1532370
Conc: 678.47 ng/ml