

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP013023\
 Data File : PP055099.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2023 00:37
 Operator : YP\AJ
 Sample : 01334-03 10X
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ECA2960-1-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/31/2023
 Supervised By :Ankita Jodhani 01/31/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 31 03:49:11 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.329	3.576	3413782	2943028	1.963	1.960
2) SA Decachlor...	10.084	8.589	4831708	3329233	2.790	2.465m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP013023\
Data File : PP055099.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2023 00:37
Operator : YP\AJ
Sample : 01334-03 10X
Misc :
ALS Vial : 41 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ECA2960-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/31/2023
Supervised By :Ankita Jodhani 01/31/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 03:49:11 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.329 min
Delta R.T.: -0.003 min
Response: 3413782
Conc: 1.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
ECA2960-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/31/2023
Supervised By :Ankita Jodhani 01/31/2023




#1 Tetrachloro-m-xylene

R.T.: 3.576 min
Delta R.T.: -0.003 min
Response: 2943028
Conc: 1.96 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.084 min
Delta R.T.: -0.007 min
Response: 4831708
Conc: 2.79 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.589 min
Delta R.T.: -0.009 min
Response: 3329233
Conc: 2.47 ng/ml m