

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP013122\
 Data File : PP043384.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Feb 2022 1:11
 Operator : AJ\MA
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 03:44:37 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 25 03:09:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.864	4.030	1146406	1139430	49.874	51.136
2) SA Decachlor...	10.776	9.354	829129	928408	58.540m	49.799
Target Compounds						
21) L5 AR-1248-1	6.165	5.293	187180	246983	477.434	451.212
22) L5 AR-1248-2	6.460	5.558	278441	363043	511.698	477.039
23) L5 AR-1248-3	6.674	5.602	312984	377750	481.709	494.730
24) L5 AR-1248-4	7.098	5.789	334442	410360	537.717	447.693
25) L5 AR-1248-5	7.138	6.212	323241	445190	531.713	503.945



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP013122\
Data File : PP043384.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Feb 2022 1:11
Operator : AJ\MA
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 03:44:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012422.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 25 03:09:04 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.864 min
Delta R.T.: -0.006 min
Response: 1146406
Conc: 49.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.030 min
Delta R.T.: -0.005 min
Response: 1139430
Conc: 51.14 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.776 min
Delta R.T.: -0.011 min
Response: 829129
Conc: 58.54 ng/ml m



#2 Decachlorobiphenyl


R.T.: 9.354 min
Delta R.T.: -0.013 min
Response: 928408
Conc: 49.80 ng/ml



#21 AR-1248-1


R.T.: 6.165 min
Delta R.T.: -0.008 min
Response: 187180
Conc: 477.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.293 min
Delta R.T.: -0.007 min
Response: 246983
Conc: 451.21 ng/ml




#22 AR-1248-2

R.T.: 6.460 min
Delta R.T.: -0.008 min
Response: 278441
Conc: 511.70 ng/ml



#22 AR-1248-2


R.T.: 5.558 min
Delta R.T.: -0.007 min
Response: 363043
Conc: 477.04 ng/ml



#23 AR-1248-3


R.T.: 6.674 min
Delta R.T.: -0.009 min
Response: 312984
Conc: 481.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.602 min
Delta R.T.: -0.007 min
Response: 377750
Conc: 494.73 ng/ml




#24 AR-1248-4

R.T.: 7.098 min
Delta R.T.: -0.009 min
Response: 334442
Conc: 537.72 ng/ml



#24 AR-1248-4


R.T.: 5.789 min
Delta R.T.: -0.007 min
Response: 410360
Conc: 447.69 ng/ml



#25 AR-1248-5

R.T.: 7.138 min
Delta R.T.: -0.009 min
Response: 323241
Conc: 531.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.212 min
Delta R.T.: -0.008 min
Response: 445190
Conc: 503.94 ng/ml