

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP013123\
 Data File : PP055171.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2023 22:29
 Operator : YP\AJ
 Sample : 01348-05
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MID2-H5(32-34)

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/01/2023
 Supervised By :Ankita Jodhani 02/01/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 06:26:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.328	3.574	26164040	22936285	15.048	15.279
2) SA Decachlor...	10.081	8.586	28786162	24307002	16.623	17.998m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP013123\
Data File : PP055171.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2023 22:29
Operator : YP\AJ
Sample : 01348-05
Misc :
ALS Vial : 32 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
MID2-H5(32-34)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/01/2023
Supervised By :Ankita Jodhani 02/01/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 06:26:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.328 min
Delta R.T.: -0.004 min
Response: 26164040
Conc: 15.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
MID2-H5(32-34)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/01/2023
Supervised By :Ankita Jodhani 02/01/2023




#1 Tetrachloro-m-xylene

R.T.: 3.574 min
Delta R.T.: -0.004 min
Response: 22936285
Conc: 15.28 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.081 min
Delta R.T.: -0.010 min
Response: 28786162
Conc: 16.62 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.586 min
Delta R.T.: -0.013 min
Response: 24307002
Conc: 18.00 ng/ml m