

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP013123\
 Data File : PP055179.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2023 00:41
 Operator : YP\AJ
 Sample : 01365-02
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 149TH-ST-DRUM-PPE-10

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 10:05:55 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.328	3.574	31006464	21863666	17.833	14.564m
2) SA Decachlor...	10.082	8.586	33944736	26776495	19.602	19.826
Target Compounds						
16) L4 AR-1242-1	5.499	4.658	16894441	11127325	301.747	271.648
17) L4 AR-1242-2	5.521	4.677	26699588	20291332	334.456	355.358
18) L4 AR-1242-3	5.583	4.853	15012676	10214736	302.300	330.921
19) L4 AR-1242-4	5.682	4.937	11719498	10704330	292.426m	329.495
20) L4 AR-1242-5	6.422	5.461	2673810	4558969	60.135m	121.990m#



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP013123\
Data File : PP055179.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Feb 2023 00:41
Operator : YP\AJ
Sample : 01365-02
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
149TH-ST-DRUM-PPE-10

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 02 10:05:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.328 min
Delta R.T.: -0.004 min
Response: 31006464
Conc: 17.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
149TH-ST-DRUM-PPE-10




#1 Tetrachloro-m-xylene

R.T.: 3.574 min
Delta R.T.: -0.005 min
Response: 21863666
Conc: 14.56 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.082 min
Delta R.T.: -0.009 min
Response: 33944736
Conc: 19.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.586 min
Delta R.T.: -0.012 min
Response: 26776495
Conc: 19.83 ng/ml



#16 AR-1242-1


R.T.: 5.499 min
Delta R.T.: -0.005 min
Response: 16894441
Conc: 301.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
149TH-ST-DRUM-PPE-10




#16 AR-1242-1

R.T.: 4.658 min
Delta R.T.: -0.006 min
Response: 11127325
Conc: 271.65 ng/ml




#17 AR-1242-2

R.T.: 5.521 min
Delta R.T.: -0.004 min
Response: 26699588
Conc: 334.46 ng/ml



#17 AR-1242-2


R.T.: 4.677 min
Delta R.T.: -0.006 min
Response: 20291332
Conc: 355.36 ng/ml



#18 AR-1242-3


R.T.: 5.583 min
Delta R.T.: -0.005 min
Response: 15012676
Conc: 302.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
149TH-ST-DRUM-PPE-10




#18 AR-1242-3

R.T.: 4.853 min
Delta R.T.: -0.007 min
Response: 10214736
Conc: 330.92 ng/ml




#19 AR-1242-4

R.T.: 5.682 min
Delta R.T.: -0.005 min
Response: 11719498
Conc: 292.43 ng/ml m



#19 AR-1242-4


R.T.: 4.937 min
Delta R.T.: -0.007 min
Response: 10704330
Conc: 329.49 ng/ml



#20 AR-1242-5

R.T.: 6.422 min
Delta R.T.: -0.005 min
Response: 2673810
Conc: 60.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
149TH-ST-DRUM-PPE-10



#20 AR-1242-5

R.T.: 5.461 min
Delta R.T.: -0.007 min
Response: 4558969
Conc: 121.99 ng/ml m