

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020122\
 Data File : PP043415.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Feb 2022 19:00
 Operator : AJ\MA
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 00:18:36 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 25 03:09:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.861	4.028	970452	1052870	42.220	47.252m
2)	SA Decachlor...	10.773	9.350	749613	885493	52.926	47.497
Target Compounds							
21)	L5 AR-1248-1	6.163	5.290	167377	231819	426.922	423.508
22)	L5 AR-1248-2	6.458	5.555	237497	322655	436.454	423.968
23)	L5 AR-1248-3	6.672	5.600	281582	336397	433.377	440.571
24)	L5 AR-1248-4	7.096	5.786	291765	400798	469.099	437.261
25)	L5 AR-1248-5	7.135	6.210	285649	395757	469.877	447.987



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020122\
Data File : PP043415.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Feb 2022 19:00
Operator : AJ\MA
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 02 00:18:36 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012422.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 25 03:09:04 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.861 min
Delta R.T.: -0.009 min
Response: 970452
Conc: 42.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.028 min
Delta R.T.: -0.006 min
Response: 1052870
Conc: 47.25 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.773 min
Delta R.T.: -0.012 min
Response: 749613
Conc: 52.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.350 min
Delta R.T.: -0.010 min
Response: 885493
Conc: 47.50 ng/ml



#21 AR-1248-1


R.T.: 6.163 min
Delta R.T.: -0.011 min
Response: 167377
Conc: 426.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.290 min
Delta R.T.: -0.009 min
Response: 231819
Conc: 423.51 ng/ml




#22 AR-1248-2

R.T.: 6.458 min
Delta R.T.: -0.011 min
Response: 237497
Conc: 436.45 ng/ml



#22 AR-1248-2


R.T.: 5.555 min
Delta R.T.: -0.009 min
Response: 322655
Conc: 423.97 ng/ml



#23 AR-1248-3


R.T.: 6.672 min
Delta R.T.: -0.012 min
Response: 281582
Conc: 433.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.600 min
Delta R.T.: -0.009 min
Response: 336397
Conc: 440.57 ng/ml




#24 AR-1248-4

R.T.: 7.096 min
Delta R.T.: -0.012 min
Response: 291765
Conc: 469.10 ng/ml



#24 AR-1248-4


R.T.: 5.786 min
Delta R.T.: -0.009 min
Response: 400798
Conc: 437.26 ng/ml



#25 AR-1248-5

R.T.: 7.135 min
Delta R.T.: -0.012 min
Response: 285649
Conc: 469.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.210 min
Delta R.T.: -0.011 min
Response: 395757
Conc: 447.99 ng/ml