

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020122\
 Data File : PP043416.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Feb 2022 19:17
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 00:19:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 25 03:09:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.862	4.029	1018393	1050251	44.305	47.134
2)	SA Decachlor...	10.775	9.351	780145	918862	55.082	49.287
Target Compounds							
26)	L6 AR-1254-1	7.069	6.167	327174	634554	457.155	447.497
27)	L6 AR-1254-2	7.297	6.326	502826	556181	478.216	449.832
28)	L6 AR-1254-3	7.674	6.749	535441	860747	497.848	447.893
29)	L6 AR-1254-4	7.968	6.989	363829	520572	516.148	470.896
30)	L6 AR-1254-5	8.396	7.419	408502	721926	505.741	467.342



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020122\
Data File : PP043416.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Feb 2022 19:17
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 02 00:19:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012422.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 25 03:09:04 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.862 min
Delta R.T.: -0.008 min
Response: 1018393
Conc: 44.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.029 min
Delta R.T.: -0.006 min
Response: 1050251
Conc: 47.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.775 min
Delta R.T.: -0.010 min
Response: 780145
Conc: 55.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.351 min
Delta R.T.: -0.009 min
Response: 918862
Conc: 49.29 ng/ml



#26 AR-1254-1


R.T.: 7.069 min
Delta R.T.: -0.011 min
Response: 327174
Conc: 457.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.167 min
Delta R.T.: -0.010 min
Response: 634554
Conc: 447.50 ng/ml




#27 AR-1254-2

R.T.: 7.297 min
Delta R.T.: -0.011 min
Response: 502826
Conc: 478.22 ng/ml



#27 AR-1254-2


R.T.: 6.326 min
Delta R.T.: -0.010 min
Response: 556181
Conc: 449.83 ng/ml



#28 AR-1254-3


R.T.: 7.674 min
Delta R.T.: -0.011 min
Response: 535441
Conc: 497.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.749 min
Delta R.T.: -0.010 min
Response: 860747
Conc: 447.89 ng/ml




#29 AR-1254-4

R.T.: 7.968 min
Delta R.T.: -0.012 min
Response: 363829
Conc: 516.15 ng/ml



#29 AR-1254-4


R.T.: 6.989 min
Delta R.T.: -0.011 min
Response: 520572
Conc: 470.90 ng/ml



#30 AR-1254-5

R.T.: 8.396 min
Delta R.T.: -0.012 min
Response: 408502
Conc: 505.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.419 min
Delta R.T.: -0.011 min
Response: 721926
Conc: 467.34 ng/ml