

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020122\
 Data File : PP043429.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Feb 2022 23:47
 Operator : AJ\MA
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 00:25:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 25 03:09:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.862	4.029	1041613	1015368	45.315	45.569
2) SA Decachlor...	10.774	9.351	755974	897062	53.375	48.118
Target Compounds						
16) L4 AR-1242-1	6.164	5.292	232261	327009	454.776	442.305
17) L4 AR-1242-2	6.187	5.312	357822	462752	448.990	445.851
18) L4 AR-1242-3	6.253	5.506	213174	241908	438.886	420.164
19) L4 AR-1242-4	6.358	5.601	174134	234227	444.045	468.555
20) L4 AR-1242-5	7.136	6.167	170317	301411	466.694	471.431



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020122\
Data File : PP043429.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Feb 2022 23:47
Operator : AJ\MA
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 02 00:25:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012422.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 25 03:09:04 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.862 min
Delta R.T.: -0.008 min
Response: 1041613
Conc: 45.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.029 min
Delta R.T.: -0.006 min
Response: 1015368
Conc: 45.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.774 min
Delta R.T.: -0.011 min
Response: 755974
Conc: 53.37 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.351 min
Delta R.T.: -0.009 min
Response: 897062
Conc: 48.12 ng/ml



#16 AR-1242-1


R.T.: 6.164 min
Delta R.T.: -0.010 min
Response: 232261
Conc: 454.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.292 min
Delta R.T.: -0.009 min
Response: 327009
Conc: 442.30 ng/ml




#17 AR-1242-2

R.T.: 6.187 min
Delta R.T.: -0.011 min
Response: 357822
Conc: 448.99 ng/ml



#17 AR-1242-2


R.T.: 5.312 min
Delta R.T.: -0.009 min
Response: 462752
Conc: 445.85 ng/ml



#18 AR-1242-3


R.T.: 6.253 min
Delta R.T.: -0.011 min
Response: 213174
Conc: 438.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.506 min
Delta R.T.: -0.009 min
Response: 241908
Conc: 420.16 ng/ml




#19 AR-1242-4

R.T.: 6.358 min
Delta R.T.: -0.010 min
Response: 174134
Conc: 444.05 ng/ml



#19 AR-1242-4


R.T.: 5.601 min
Delta R.T.: -0.009 min
Response: 234227
Conc: 468.55 ng/ml



#20 AR-1242-5

R.T.: 7.136 min
Delta R.T.: -0.011 min
Response: 170317
Conc: 466.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 6.167 min
Delta R.T.: -0.011 min
Response: 301411
Conc: 471.43 ng/ml