

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020123\
 Data File : PP055212.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2023 16:55
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 20:56:18 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.328	3.574	79538741	71308625	45.746	47.501
2) SA Decachlor...	10.081	8.585	85629004	71089049	49.449	52.637
Target Compounds						
26) L6 AR-1254-1	6.358	5.461	35425072	39861461	449.030	493.591
27) L6 AR-1254-2	6.578	5.610	55139770	34962279	455.614	488.555
28) L6 AR-1254-3	6.945	6.014	57441614	55328121	468.949	496.628
29) L6 AR-1254-4	7.232	6.244	41244412	30177622	468.091	490.634
30) L6 AR-1254-5	7.653	6.663	45912246	47131282	453.854	497.144



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020123\
Data File : PP055212.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Feb 2023 16:55
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 20:56:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.328 min
Delta R.T.: -0.004 min
Response: 79538741
Conc: 45.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.574 min
Delta R.T.: -0.004 min
Response: 71308625
Conc: 47.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.081 min
Delta R.T.: -0.010 min
Response: 85629004
Conc: 49.45 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.585 min
Delta R.T.: -0.013 min
Response: 71089049
Conc: 52.64 ng/ml



#26 AR-1254-1


R.T.: 6.358 min
Delta R.T.: -0.004 min
Response: 35425072
Conc: 449.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.461 min
Delta R.T.: -0.006 min
Response: 39861461
Conc: 493.59 ng/ml




#27 AR-1254-2

R.T.: 6.578 min
Delta R.T.: -0.004 min
Response: 55139770
Conc: 455.61 ng/ml



#27 AR-1254-2


R.T.: 5.610 min
Delta R.T.: -0.006 min
Response: 34962279
Conc: 488.56 ng/ml



#28 AR-1254-3


R.T.: 6.945 min
Delta R.T.: -0.004 min
Response: 57441614
Conc: 468.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.014 min
Delta R.T.: -0.006 min
Response: 55328121
Conc: 496.63 ng/ml




#29 AR-1254-4

R.T.: 7.232 min
Delta R.T.: -0.004 min
Response: 41244412
Conc: 468.09 ng/ml



#29 AR-1254-4


R.T.: 6.244 min
Delta R.T.: -0.006 min
Response: 30177622
Conc: 490.63 ng/ml



#30 AR-1254-5

R.T.: 7.653 min
Delta R.T.: -0.005 min
Response: 45912246
Conc: 453.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.663 min
Delta R.T.: -0.006 min
Response: 47131282
Conc: 497.14 ng/ml