

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP020222.M
 Title : GC EXTRACTABLES
 Last Update : Thu Feb 03 09:16:35 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP043460.D 1000 =PP043458.D 750 =PP043459.D
 250 =PP043461.D 50 =PP043462.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.891	2.785	2.822	2.875	2.440	2.763 E4	6.71
2) SA	Decachlorobiphenyl	2.619	2.456	2.523	2.730	2.924	2.650 E4	6.97
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	2.177	2.046	2.093	2.250	2.449	2.203 E3	7.19
42) L9	AR-1268-2	2.081	1.947	1.991	2.142	2.343	2.101 E3	7.39
43) L9	AR-1268-3	1.801	1.708	1.732	1.838	1.943	1.804 E3	5.16
44) L9	AR-1268-4	6.665	6.405	6.492	6.663	6.133	6.472 E2	3.40
45) L9	AR-1268-5	5.540	5.254	5.373	5.669	7.233	5.814 E3	13.92

Signal #2 Calibration Files

500 =PP043460.D 1000 =PP043458.D 750 =PP043459.D
 250 =PP043461.D 50 =PP043462.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.413	2.269	2.322	2.494	2.772	2.454 E4	8.05
2) SA	Decachlorobiphenyl	3.333	3.078	3.166	3.552	3.916	3.409 E4	9.85
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP020222.M
 Title : GC EXTRACTABLES
 Last Update : Thu Feb 03 09:16:35 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP043460.D 1000 =PP043458.D 750 =PP043459.D
 250 =PP043461.D 50 =PP043462.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	2.819	2.618	2.670	2.979	3.082	2.834 E3	6.98
42) L9	AR-1268-2	2.715	2.530	2.594	2.846	3.100	2.757 E3	8.22
43) L9	AR-1268-3	2.342	2.164	2.212	2.453	4.126	2.659 E3	31.13
44) L9	AR-1268-4	0.990	0.922	0.936	1.019	1.066	0.987 E3	6.04
45) L9	AR-1268-5	6.952	6.624	6.732	7.175	7.844	7.065 E3	6.85

(#) = Out of Range