

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020222\
 Data File : PP043453.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Feb 2022 19:17
 Operator : AJ\MA
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 09:25:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Feb 03 09:25:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.867	4.025	2275474	1858342	77.244	76.707
2) SA Decachlor...	10.781	9.349	1127471	1457279	75.231	75.606
Target Compounds						
26) L6 AR-1254-1	7.074	6.165	638269	938030	764.193	761.903
27) L6 AR-1254-2	7.302	6.324	921241	816704	759.114	760.298
28) L6 AR-1254-3	7.680	6.747	943109	1314048	760.934	760.082
29) L6 AR-1254-4	7.973	6.987	623373	790006	759.192	759.857
30) L6 AR-1254-5	8.400	7.417	682572	1116690	756.817	762.321



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020222\
Data File : PP043453.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Feb 2022 19:17
Operator : AJ\MA
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 03 09:25:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Feb 03 09:25:11 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.867 min
Delta R.T.: 0.000 min
Response: 2275474
Conc: 77.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#1 Tetrachloro-m-xylene

R.T.: 4.025 min
Delta R.T.: 0.000 min
Response: 1858342
Conc: 76.71 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.781 min
Delta R.T.: 0.000 min
Response: 1127471
Conc: 75.23 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.349 min
Delta R.T.: 0.000 min
Response: 1457279
Conc: 75.61 ng/ml



#26 AR-1254-1


R.T.: 7.074 min
Delta R.T.: 0.000 min
Response: 638269
Conc: 764.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254\ICC750




#26 AR-1254-1

R.T.: 6.165 min
Delta R.T.: 0.000 min
Response: 938030
Conc: 761.90 ng/ml




#27 AR-1254-2

R.T.: 7.302 min
Delta R.T.: 0.000 min
Response: 921241
Conc: 759.11 ng/ml



#27 AR-1254-2


R.T.: 6.324 min
Delta R.T.: 0.000 min
Response: 816704
Conc: 760.30 ng/ml



#28 AR-1254-3


R.T.: 7.680 min
Delta R.T.: 0.000 min
Response: 943109
Conc: 760.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#28 AR-1254-3

R.T.: 6.747 min
Delta R.T.: 0.000 min
Response: 1314048
Conc: 760.08 ng/ml




#29 AR-1254-4

R.T.: 7.973 min
Delta R.T.: 0.000 min
Response: 623373
Conc: 759.19 ng/ml



#29 AR-1254-4


R.T.: 6.987 min
Delta R.T.: 0.000 min
Response: 790006
Conc: 759.86 ng/ml



#30 AR-1254-5

R.T.: 8.400 min
Delta R.T.: 0.000 min
Response: 682572
Conc: 756.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254\ICC750



#30 AR-1254-5

R.T.: 7.417 min
Delta R.T.: 0.000 min
Response: 1116690
Conc: 762.32 ng/ml