

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020223\
 Data File : PP055283.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Feb 2023 22:04
 Operator : YP\AJ
 Sample : PB150622BS
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 04:39:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.328	3.575	32752583	26653339	18.837	17.755
2)	SA Decachlor...	10.085	8.589	34416251	27485223	19.875	20.351
Target Compounds							
3)	L1 AR-1016-1	5.499	4.658	24869349	20004392	370.024	404.179
4)	L1 AR-1016-2	5.522	4.677	35693381	28356454	372.469	409.662
5)	L1 AR-1016-3	5.584	4.853	22233426	15213422	371.695	409.000
6)	L1 AR-1016-4	5.683	4.895	17768469	13098528	370.809	416.855
7)	L1 AR-1016-5	5.979	5.109	17950977	15295786	354.243	374.122
8)	L2 AR-1221-1	4.532	3.787	2422386	2111934	103.014	106.519
9)	L2 AR-1221-2	4.623	3.873	3121069	2740458	175.667	184.887
10)	L2 AR-1221-3	4.700	3.949	13763515	11782892	255.782	261.178
11)	L3 AR-1232-1	4.700	3.949	13763515	11782892	310.109	319.671
12)	L3 AR-1232-2	5.230	4.677	18297606	28356454	809.099	887.733
13)	L3 AR-1232-3	5.522	4.853	35693381	15213422	812.741	912.569
14)	L3 AR-1232-4	5.683	4.937	17768469	15976752	802.760	990.737
15)	L3 AR-1232-5	5.775	5.109	16392360	15295786	936.330	866.054
16)	L4 AR-1242-1	5.499	4.658	24869349	20004392	444.184	488.362
17)	L4 AR-1242-2	5.522	4.677	35693381	28356454	447.119	496.601
18)	L4 AR-1242-3	5.584	4.853	22233426	15213422	447.700	492.860
19)	L4 AR-1242-4	5.683	4.937	17768469	15976752	443.361	491.788
20)	L4 AR-1242-5	6.423	5.461	2374298	12929100	53.399	345.960 #
21)	L5 AR-1248-1	5.499	4.658	24869349	20004392	596.334	645.238
22)	L5 AR-1248-2	5.775	4.895	16392360	13098528	257.846	286.448
23)	L5 AR-1248-3	5.979	4.937	17950977	15976752	259.564	335.130 #
24)	L5 AR-1248-4	6.383	5.109	2555702	15295786	33.701	276.961 #
25)	L5 AR-1248-5	6.423	5.502	2374298	1733762	32.175	35.260
26)	L6 AR-1254-1	6.358	5.461	13815781	12929100	175.122	160.097
27)	L6 AR-1254-2	6.578	5.609	15066140	12372807	124.490	172.895 #
28)	L6 AR-1254-3	6.945	6.029	8105708	21251225	66.174	190.752 #
29)	L6 AR-1254-4	7.233	6.243	5409693	1884410	61.396	30.637 #
30)	L6 AR-1254-5	7.653	6.662	47014250	39693317	464.748	418.688
31)	L7 AR-1260-1	7.111	6.145	33668732	30484589	353.194	393.124
32)	L7 AR-1260-2	7.370	6.333	39112308	35172995	347.522	391.254
33)	L7 AR-1260-3	7.731	6.487	25740809	33432478	319.687	393.737
34)	L7 AR-1260-4	7.956	6.961	32160413	24731346	331.288	345.396
35)	L7 AR-1260-5	8.274	7.204	55595534	52774631	312.597	343.280
36)	L8 AR-1262-1	7.731	6.755	25740809	14379331	218.175	334.703 #
37)	L8 AR-1262-2	8.274	6.961	55595534	24731346	276.027	267.010
38)	L8 AR-1262-3	8.597	7.488	38433933	11291959	276.362	154.982 #
39)	L8 AR-1262-4	8.656f	7.551	23184749	40104068	301.565	306.654
40)	L8 AR-1262-5	9.331	8.049	14646570	10799968	204.791	182.039
41)	L9 AR-1268-1	8.597	7.488	38433933	11291959	157.254	52.223 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020223\
 Data File : PP055283.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Feb 2023 22:04
 Operator : YP\AJ
 Sample : PB150622BS
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 04:39:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.656f	7.551	23184749	40104068	102.080	205.521 #
43) L9	AR-1268-3	8.904	7.760	1371396	929300	6.864	5.410
44) L9	AR-1268-4	9.331	8.049	14646570	10799968	186.470	169.325
45) L9	AR-1268-5	9.746	8.337	5231997	4336152	8.225	8.234



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020223\
Data File : PP055283.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Feb 2023 22:04
Operator : YP\AJ
Sample : PB150622BS
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 03 04:39:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.328 min
Delta R.T.: -0.004 min
Response: 32752583
Conc: 18.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.575 min
Delta R.T.: -0.003 min
Response: 26653339
Conc: 17.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.085 min
Delta R.T.: -0.006 min
Response: 34416251
Conc: 19.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.589 min
Delta R.T.: -0.010 min
Response: 27485223
Conc: 20.35 ng/ml



#3 AR-1016-1


R.T.: 5.499 min
Delta R.T.: -0.004 min
Response: 24869349
Conc: 370.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.658 min
Delta R.T.: -0.005 min
Response: 20004392
Conc: 404.18 ng/ml




#4 AR-1016-2

R.T.: 5.522 min
Delta R.T.: -0.005 min
Response: 35693381
Conc: 372.47 ng/ml



#4 AR-1016-2


R.T.: 4.677 min
Delta R.T.: -0.006 min
Response: 28356454
Conc: 409.66 ng/ml



#5 AR-1016-3


R.T.: 5.584 min
Delta R.T.: -0.004 min
Response: 22233426
Conc: 371.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.853 min
Delta R.T.: -0.007 min
Response: 15213422
Conc: 409.00 ng/ml




#6 AR-1016-4

R.T.: 5.683 min
Delta R.T.: -0.004 min
Response: 17768469
Conc: 370.81 ng/ml



#6 AR-1016-4


R.T.: 4.895 min
Delta R.T.: -0.006 min
Response: 13098528
Conc: 416.86 ng/ml



#7 AR-1016-5


R.T.: 5.979 min
Delta R.T.: -0.004 min
Response: 17950977
Conc: 354.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.109 min
Delta R.T.: -0.006 min
Response: 15295786
Conc: 374.12 ng/ml




#8 AR-1221-1

R.T.: 4.532 min
Delta R.T.: -0.008 min
Response: 2422386
Conc: 103.01 ng/ml



#8 AR-1221-1


R.T.: 3.787 min
Delta R.T.: -0.005 min
Response: 2111934
Conc: 106.52 ng/ml



#9 AR-1221-2


R.T.: 4.623 min
Delta R.T.: -0.003 min
Response: 3121069
Conc: 175.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.873 min
Delta R.T.: -0.003 min
Response: 2740458
Conc: 184.89 ng/ml




#10 AR-1221-3

R.T.: 4.700 min
Delta R.T.: -0.003 min
Response: 13763515
Conc: 255.78 ng/ml



#10 AR-1221-3


R.T.: 3.949 min
Delta R.T.: -0.004 min
Response: 11782892
Conc: 261.18 ng/ml



#11 AR-1232-1


R.T.: 4.700 min
Delta R.T.: -0.004 min
Response: 13763515
Conc: 310.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.949 min
Delta R.T.: -0.005 min
Response: 11782892
Conc: 319.67 ng/ml




#12 AR-1232-2

R.T.: 5.230 min
Delta R.T.: -0.005 min
Response: 18297606
Conc: 809.10 ng/ml



#12 AR-1232-2


R.T.: 4.677 min
Delta R.T.: -0.006 min
Response: 28356454
Conc: 887.73 ng/ml



#13 AR-1232-3


R.T.: 5.522 min
Delta R.T.: -0.005 min
Response: 35693381
Conc: 812.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.853 min
Delta R.T.: -0.006 min
Response: 15213422
Conc: 912.57 ng/ml




#14 AR-1232-4

R.T.: 5.683 min
Delta R.T.: -0.005 min
Response: 17768469
Conc: 802.76 ng/ml



#14 AR-1232-4


R.T.: 4.937 min
Delta R.T.: -0.007 min
Response: 15976752
Conc: 990.74 ng/ml



#15 AR-1232-5


R.T.: 5.775 min
Delta R.T.: -0.005 min
Response: 16392360
Conc: 936.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.109 min
Delta R.T.: -0.007 min
Response: 15295786
Conc: 866.05 ng/ml




#16 AR-1242-1

R.T.: 5.499 min
Delta R.T.: -0.004 min
Response: 24869349
Conc: 444.18 ng/ml



#16 AR-1242-1


R.T.: 4.658 min
Delta R.T.: -0.006 min
Response: 20004392
Conc: 488.36 ng/ml



#17 AR-1242-2


R.T.: 5.522 min
Delta R.T.: -0.004 min
Response: 35693381
Conc: 447.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.677 min
Delta R.T.: -0.006 min
Response: 28356454
Conc: 496.60 ng/ml




#18 AR-1242-3

R.T.: 5.584 min
Delta R.T.: -0.004 min
Response: 22233426
Conc: 447.70 ng/ml



#18 AR-1242-3


R.T.: 4.853 min
Delta R.T.: -0.007 min
Response: 15213422
Conc: 492.86 ng/ml



#19 AR-1242-4


R.T.: 5.683 min
Delta R.T.: -0.004 min
Response: 17768469
Conc: 443.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.937 min
Delta R.T.: -0.006 min
Response: 15976752
Conc: 491.79 ng/ml




#20 AR-1242-5

R.T.: 6.423 min
Delta R.T.: -0.003 min
Response: 2374298
Conc: 53.40 ng/ml



#20 AR-1242-5


R.T.: 5.461 min
Delta R.T.: -0.006 min
Response: 12929100
Conc: 345.96 ng/ml



#21 AR-1248-1


R.T.: 5.499 min
Delta R.T.: -0.003 min
Response: 24869349
Conc: 596.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.658 min
Delta R.T.: -0.005 min
Response: 20004392
Conc: 645.24 ng/ml




#22 AR-1248-2

R.T.: 5.775 min
Delta R.T.: -0.003 min
Response: 16392360
Conc: 257.85 ng/ml



#22 AR-1248-2


R.T.: 4.895 min
Delta R.T.: -0.005 min
Response: 13098528
Conc: 286.45 ng/ml



#23 AR-1248-3


R.T.: 5.979 min
Delta R.T.: -0.003 min
Response: 17950977
Conc: 259.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.937 min
Delta R.T.: -0.005 min
Response: 15976752
Conc: 335.13 ng/ml




#24 AR-1248-4

R.T.: 6.383 min
Delta R.T.: -0.004 min
Response: 2555702
Conc: 33.70 ng/ml



#24 AR-1248-4


R.T.: 5.109 min
Delta R.T.: -0.005 min
Response: 15295786
Conc: 276.96 ng/ml



#25 AR-1248-5


R.T.: 6.423 min
Delta R.T.: -0.003 min
Response: 2374298
Conc: 32.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.502 min
Delta R.T.: -0.006 min
Response: 1733762
Conc: 35.26 ng/ml




#26 AR-1254-1

R.T.: 6.358 min
Delta R.T.: -0.003 min
Response: 13815781
Conc: 175.12 ng/ml



#26 AR-1254-1


R.T.: 5.461 min
Delta R.T.: -0.006 min
Response: 12929100
Conc: 160.10 ng/ml



#27 AR-1254-2


R.T.: 6.578 min
Delta R.T.: -0.004 min
Response: 15066140
Conc: 124.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.609 min
Delta R.T.: -0.006 min
Response: 12372807
Conc: 172.89 ng/ml




#28 AR-1254-3

R.T.: 6.945 min
Delta R.T.: -0.003 min
Response: 8105708
Conc: 66.17 ng/ml



#28 AR-1254-3


R.T.: 6.029 min
Delta R.T.: 0.009 min
Response: 21251225
Conc: 190.75 ng/ml



#29 AR-1254-4


R.T.: 7.233 min
Delta R.T.: -0.003 min
Response: 5409693
Conc: 61.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.243 min
Delta R.T.: -0.006 min
Response: 1884410
Conc: 30.64 ng/ml




#30 AR-1254-5

R.T.: 7.653 min
Delta R.T.: -0.005 min
Response: 47014250
Conc: 464.75 ng/ml



#30 AR-1254-5


R.T.: 6.662 min
Delta R.T.: -0.007 min
Response: 39693317
Conc: 418.69 ng/ml



#31 AR-1260-1


R.T.: 7.111 min
Delta R.T.: -0.004 min
Response: 33668732
Conc: 353.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.145 min
Delta R.T.: -0.007 min
Response: 30484589
Conc: 393.12 ng/ml




#32 AR-1260-2

R.T.: 7.370 min
Delta R.T.: -0.004 min
Response: 39112308
Conc: 347.52 ng/ml



#32 AR-1260-2


R.T.: 6.333 min
Delta R.T.: -0.008 min
Response: 35172995
Conc: 391.25 ng/ml



#33 AR-1260-3


R.T.: 7.731 min
Delta R.T.: -0.004 min
Response: 25740809
Conc: 319.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.487 min
Delta R.T.: -0.008 min
Response: 33432478
Conc: 393.74 ng/ml




#34 AR-1260-4

R.T.: 7.956 min
Delta R.T.: -0.005 min
Response: 32160413
Conc: 331.29 ng/ml



#34 AR-1260-4


R.T.: 6.961 min
Delta R.T.: -0.008 min
Response: 24731346
Conc: 345.40 ng/ml



#35 AR-1260-5


R.T.: 8.274 min
Delta R.T.: -0.004 min
Response: 55595534
Conc: 312.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.204 min
Delta R.T.: -0.008 min
Response: 52774631
Conc: 343.28 ng/ml




#36 AR-1262-1

R.T.: 7.731 min
Delta R.T.: -0.003 min
Response: 25740809
Conc: 218.18 ng/ml



#36 AR-1262-1


R.T.: 6.755 min
Delta R.T.: -0.006 min
Response: 14379331
Conc: 334.70 ng/ml



#37 AR-1262-2


R.T.: 8.274 min
Delta R.T.: -0.003 min
Response: 55595534
Conc: 276.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.961 min
Delta R.T.: -0.006 min
Response: 24731346
Conc: 267.01 ng/ml




#38 AR-1262-3

R.T.: 8.597 min
Delta R.T.: 0.008 min
Response: 38433933
Conc: 276.36 ng/ml



#38 AR-1262-3


R.T.: 7.488 min
Delta R.T.: -0.006 min
Response: 11291959
Conc: 154.98 ng/ml



#39 AR-1262-4


R.T.: 8.656 min
Delta R.T.: -0.022 min
Response: 23184749
Conc: 301.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.551 min
Delta R.T.: -0.008 min
Response: 40104068
Conc: 306.65 ng/ml




#40 AR-1262-5

R.T.: 9.331 min
Delta R.T.: -0.003 min
Response: 14646570
Conc: 204.79 ng/ml



#40 AR-1262-5


R.T.: 8.049 min
Delta R.T.: -0.005 min
Response: 10799968
Conc: 182.04 ng/ml



#41 AR-1268-1


R.T.: 8.597 min
Delta R.T.: 0.012 min
Response: 38433933
Conc: 157.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.488 min
Delta R.T.: -0.005 min
Response: 11291959
Conc: 52.22 ng/ml




#42 AR-1268-2

R.T.: 8.656 min
Delta R.T.: -0.023 min
Response: 23184749
Conc: 102.08 ng/ml



#42 AR-1268-2


R.T.: 7.551 min
Delta R.T.: -0.007 min
Response: 40104068
Conc: 205.52 ng/ml



#43 AR-1268-3


R.T.: 8.904 min
Delta R.T.: -0.003 min
Response: 1371396
Conc: 6.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.760 min
Delta R.T.: -0.006 min
Response: 929300
Conc: 5.41 ng/ml




#44 AR-1268-4

R.T.: 9.331 min
Delta R.T.: -0.002 min
Response: 14646570
Conc: 186.47 ng/ml



#44 AR-1268-4


R.T.: 8.049 min
Delta R.T.: -0.005 min
Response: 10799968
Conc: 169.33 ng/ml



#45 AR-1268-5

R.T.: 9.746 min
Delta R.T.: -0.003 min
Response: 5231997
Conc: 8.22 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.337 min
Delta R.T.: -0.006 min
Response: 4336152
Conc: 8.23 ng/ml