

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020223\
 Data File : PP055293.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Feb 2023 01:51
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 04:45:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.328	3.573	87537963	76450196	50.346	50.926
2) SA Decachlor...	10.086	8.587	87107568	69270500	50.303	51.290
Target Compounds						
26) L6 AR-1254-1	6.358	5.460	35113552	42798239	445.082	529.956
27) L6 AR-1254-2	6.578	5.609	55676030	37300796	460.045	521.233
28) L6 AR-1254-3	6.945	6.013	57368792	58074180	468.355	521.277
29) L6 AR-1254-4	7.232	6.243	42364361	31812146	480.801	517.208
30) L6 AR-1254-5	7.654	6.662	45293220	48713128	447.735	513.830



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020223\
Data File : PP055293.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Feb 2023 01:51
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 03 04:45:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.328 min
Delta R.T.: -0.004 min
Response: 87537963
Conc: 50.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.573 min
Delta R.T.: -0.005 min
Response: 76450196
Conc: 50.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.086 min
Delta R.T.: -0.005 min
Response: 87107568
Conc: 50.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.587 min
Delta R.T.: -0.011 min
Response: 69270500
Conc: 51.29 ng/ml



#26 AR-1254-1


R.T.: 6.358 min
Delta R.T.: -0.003 min
Response: 35113552
Conc: 445.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.460 min
Delta R.T.: -0.007 min
Response: 42798239
Conc: 529.96 ng/ml




#27 AR-1254-2

R.T.: 6.578 min
Delta R.T.: -0.004 min
Response: 55676030
Conc: 460.04 ng/ml



#27 AR-1254-2


R.T.: 5.609 min
Delta R.T.: -0.007 min
Response: 37300796
Conc: 521.23 ng/ml



#28 AR-1254-3


R.T.: 6.945 min
Delta R.T.: -0.003 min
Response: 57368792
Conc: 468.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.013 min
Delta R.T.: -0.007 min
Response: 58074180
Conc: 521.28 ng/ml




#29 AR-1254-4

R.T.: 7.232 min
Delta R.T.: -0.004 min
Response: 42364361
Conc: 480.80 ng/ml



#29 AR-1254-4


R.T.: 6.243 min
Delta R.T.: -0.007 min
Response: 31812146
Conc: 517.21 ng/ml



#30 AR-1254-5

R.T.: 7.654 min
Delta R.T.: -0.003 min
Response: 45293220
Conc: 447.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.662 min
Delta R.T.: -0.007 min
Response: 48713128
Conc: 513.83 ng/ml