

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020322\
 Data File : PP043471.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Feb 2022 11:35
 Operator : AJ\MA
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 00:48:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Feb 03 10:01:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.867	4.024	1472095	1245869	53.639	55.009
2) SA Decachlor...	10.782	9.348	758673	1003653	49.457	53.091
Target Compounds						
21) L5 AR-1248-1	6.170	5.288	229942	242248	477.434	484.249
22) L5 AR-1248-2	6.465	5.553	323146	344351	467.337	481.019
23) L5 AR-1248-3	6.679	5.598	384819	362109	479.734	486.859
24) L5 AR-1248-4	7.103	5.784	376711	428387	486.604	478.085
25) L5 AR-1248-5	7.143	6.208	366805	426411	485.406	488.600



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020322\
Data File : PP043471.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Feb 2022 11:35
Operator : AJ\MA
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 00:48:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Feb 03 10:01:05 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.867 min
Delta R.T.: 0.002 min
Response: 1472095
Conc: 53.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.024 min
Delta R.T.: 0.000 min
Response: 1245869
Conc: 55.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.782 min
Delta R.T.: 0.003 min
Response: 758673
Conc: 49.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.348 min
Delta R.T.: 0.000 min
Response: 1003653
Conc: 53.09 ng/ml



#21 AR-1248-1


R.T.: 6.170 min
Delta R.T.: 0.002 min
Response: 229942
Conc: 477.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.288 min
Delta R.T.: 0.000 min
Response: 242248
Conc: 484.25 ng/ml




#22 AR-1248-2

R.T.: 6.465 min
Delta R.T.: 0.002 min
Response: 323146
Conc: 467.34 ng/ml



#22 AR-1248-2


R.T.: 5.553 min
Delta R.T.: 0.000 min
Response: 344351
Conc: 481.02 ng/ml



#23 AR-1248-3


R.T.: 6.679 min
Delta R.T.: 0.002 min
Response: 384819
Conc: 479.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.598 min
Delta R.T.: 0.000 min
Response: 362109
Conc: 486.86 ng/ml




#24 AR-1248-4

R.T.: 7.103 min
Delta R.T.: 0.002 min
Response: 376711
Conc: 486.60 ng/ml



#24 AR-1248-4


R.T.: 5.784 min
Delta R.T.: 0.000 min
Response: 428387
Conc: 478.09 ng/ml



#25 AR-1248-5

R.T.: 7.143 min
Delta R.T.: 0.002 min
Response: 366805
Conc: 485.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.208 min
Delta R.T.: 0.000 min
Response: 426411
Conc: 488.60 ng/ml