

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020323\
 Data File : PP055331.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Feb 2023 23:10
 Operator : YP\AJ
 Sample : 01418-03
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 06:23:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.326	3.573	25914888	22863316	14.905	15.230
2)	SA Decachlor...	10.078	8.582	24625313	20146270	14.221	14.917
Target Compounds							
3)	L1 AR-1016-1	5.497	4.654	803140	1504985	11.950	30.407 #
4)	L1 AR-1016-2	5.518	4.654f	883759	1504985	9.222	21.742 #
5)	L1 AR-1016-3	5.577	4.850	443987	617180	7.422	16.592 #
6)	L1 AR-1016-4	5.686	4.896	723938	1337711	15.108	42.572 #
7)	L1 AR-1016-5	5.970	0.000	2508957	0	49.512	N.D. #
8)	L2 AR-1221-1	4.535	3.791	161202	582783	6.855	29.394 #
9)	L2 AR-1221-2	4.633	3.872	1117370	2870072	62.890	193.631 #
10)	L2 AR-1221-3	4.711	3.948	925893	649635	17.207	14.400
11)	L3 AR-1232-1	4.711	3.948	925893	649635	20.862	17.625
12)	L3 AR-1232-2	5.234	4.654f	2383748	1504985	105.407	47.115 #
13)	L3 AR-1232-3	5.518	4.850	883759	617180	20.123	37.021 #
14)	L3 AR-1232-4	5.686	4.956	723938	1342161	32.707	83.229 #
15)	L3 AR-1232-5	5.775	0.000	863004	0	49.295	N.D. #
16)	L4 AR-1242-1	5.497	4.654	803140	1504985	14.345	36.741 #
17)	L4 AR-1242-2	5.518	4.654f	883759	1504985	11.071	26.357 #
18)	L4 AR-1242-3	5.577	4.850	443987	617180	8.940	19.994 #
19)	L4 AR-1242-4	5.686	4.956	723938	1342161	18.064	41.314 #
20)	L4 AR-1242-5	6.428	5.462	351853	1399982	7.913	37.461 #
21)	L5 AR-1248-1	5.497	4.654	803140	1504985	19.258	48.543 #
22)	L5 AR-1248-2	5.775	4.896	863004	1337711	13.575	29.254 #
23)	L5 AR-1248-3	5.970	4.956	2508957	1342161	36.279	28.153
24)	L5 AR-1248-4	6.388	0.000	369917	0	4.878	N.D. #
25)	L5 AR-1248-5	6.428	5.505	351853	1376946	4.768	28.003 #
26)	L6 AR-1254-1	6.350	5.462	697215	1399982	8.838	17.335 #
27)	L6 AR-1254-2	6.575	5.608	2461063	1438277	20.335	20.098
28)	L6 AR-1254-3	6.973f	6.022	7314341	1274336	59.714	11.439 #
29)	L6 AR-1254-4	7.227	6.233f	500431	1042630	5.679	16.951 #
30)	L6 AR-1254-5	7.671	6.658	2471905	883054	24.435	9.315 #
31)	L7 AR-1260-1	7.104	6.144	1939531	2303071	20.346	29.700 #
32)	L7 AR-1260-2	7.351f	6.347	1187762	2026322	10.554	22.540 #
33)	L7 AR-1260-3	7.728	6.484	282953	508612	3.514	5.990 #
34)	L7 AR-1260-4	7.957	6.955	1527871	217661	15.739	3.040 #
35)	L7 AR-1260-5	8.271	7.198	1226360	668264	6.895	4.347 #
36)	L8 AR-1262-1	7.728	6.753	282953	203236	2.398	4.731 #
37)	L8 AR-1262-2	8.271	6.955	1226360	217661	6.089	2.350 #
38)	L8 AR-1262-3	8.570f	7.482	84299	808808	0.606	11.101 #
39)	L8 AR-1262-4	8.673	7.550	803721	880132	10.454	6.730 #
40)	L8 AR-1262-5	9.328	8.077f	-18252	77614	N.D.	1.308 #
41)	L9 AR-1268-1	8.570f	7.482	84299	808808	0.345	3.741 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020323\
Data File : PP055331.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Feb 2023 23:10
Operator : YP\AJ
Sample : 01418-03
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 06:23:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.673	7.550	803721	880132	3.539	4.510 #
43)	L9 AR-1268-3	8.884f	0.000	139381	0	0.698	N.D. #
44)	L9 AR-1268-4	9.328	8.077f	-18252	77614	N.D.	1.217 #
45)	L9 AR-1268-5	9.740	8.335	367740	122449	0.578	0.233 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020323\
Data File : PP055331.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Feb 2023 23:10
Operator : YP\AJ
Sample : 01418-03
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 06:23:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.326 min
Delta R.T.: -0.006 min
Response: 25914888
Conc: 14.90 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.573 min
Delta R.T.: -0.005 min
Response: 22863316
Conc: 15.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.078 min
Delta R.T.: -0.013 min
Response: 24625313
Conc: 14.22 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.582 min
Delta R.T.: -0.016 min
Response: 20146270
Conc: 14.92 ng/ml



#3 AR-1016-1


R.T.: 5.497 min
Delta R.T.: -0.006 min
Response: 803140
Conc: 11.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.654 min
Delta R.T.: -0.009 min
Response: 1504985
Conc: 30.41 ng/ml




#4 AR-1016-2

R.T.: 5.518 min
Delta R.T.: -0.008 min
Response: 883759
Conc: 9.22 ng/ml



#4 AR-1016-2


R.T.: 4.654 min
Delta R.T.: -0.028 min
Response: 1504985
Conc: 21.74 ng/ml



#5 AR-1016-3


R.T.: 5.577 min
Delta R.T.: -0.011 min
Response: 443987
Conc: 7.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.850 min
Delta R.T.: -0.009 min
Response: 617180
Conc: 16.59 ng/ml




#6 AR-1016-4

R.T.: 5.686 min
Delta R.T.: 0.000 min
Response: 723938
Conc: 15.11 ng/ml



#6 AR-1016-4


R.T.: 4.896 min
Delta R.T.: -0.005 min
Response: 1337711
Conc: 42.57 ng/ml



#7 AR-1016-5


R.T.: 5.970 min
Delta R.T.: -0.013 min
Response: 2508957
Conc: 49.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.115 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 4.535 min
Delta R.T.: -0.004 min
Response: 161202
Conc: 6.86 ng/ml



#8 AR-1221-1


R.T.: 3.791 min
Delta R.T.: 0.000 min
Response: 582783
Conc: 29.39 ng/ml



#9 AR-1221-2


R.T.: 4.633 min
Delta R.T.: 0.008 min
Response: 1117370
Conc: 62.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.872 min
Delta R.T.: -0.005 min
Response: 2870072
Conc: 193.63 ng/ml




#10 AR-1221-3

R.T.: 4.711 min
Delta R.T.: 0.008 min
Response: 925893
Conc: 17.21 ng/ml



#10 AR-1221-3


R.T.: 3.948 min
Delta R.T.: -0.005 min
Response: 649635
Conc: 14.40 ng/ml



#11 AR-1232-1


R.T.: 4.711 min
Delta R.T.: 0.007 min
Response: 925893
Conc: 20.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.948 min
Delta R.T.: -0.006 min
Response: 649635
Conc: 17.62 ng/ml




#12 AR-1232-2

R.T.: 5.234 min
Delta R.T.: -0.001 min
Response: 2383748
Conc: 105.41 ng/ml



#12 AR-1232-2


R.T.: 4.654 min
Delta R.T.: -0.029 min
Response: 1504985
Conc: 47.12 ng/ml



#13 AR-1232-3


R.T.: 5.518 min
Delta R.T.: -0.008 min
Response: 883759
Conc: 20.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.850 min
Delta R.T.: -0.009 min
Response: 617180
Conc: 37.02 ng/ml




#14 AR-1232-4

R.T.: 5.686 min
Delta R.T.: -0.001 min
Response: 723938
Conc: 32.71 ng/ml



#14 AR-1232-4


R.T.: 4.956 min
Delta R.T.: 0.012 min
Response: 1342161
Conc: 83.23 ng/ml



#15 AR-1232-5


R.T.: 5.775 min
Delta R.T.: -0.004 min
Response: 863004
Conc: 49.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 5.115 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.497 min
Delta R.T.: -0.006 min
Response: 803140
Conc: 14.34 ng/ml



#16 AR-1242-1


R.T.: 4.654 min
Delta R.T.: -0.010 min
Response: 1504985
Conc: 36.74 ng/ml



#17 AR-1242-2


R.T.: 5.518 min
Delta R.T.: -0.008 min
Response: 883759
Conc: 11.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.654 min
Delta R.T.: -0.028 min
Response: 1504985
Conc: 26.36 ng/ml




#18 AR-1242-3

R.T.: 5.577 min
Delta R.T.: -0.011 min
Response: 443987
Conc: 8.94 ng/ml



#18 AR-1242-3


R.T.: 4.850 min
Delta R.T.: -0.010 min
Response: 617180
Conc: 19.99 ng/ml



#19 AR-1242-4


R.T.: 5.686 min
Delta R.T.: 0.000 min
Response: 723938
Conc: 18.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.956 min
Delta R.T.: 0.012 min
Response: 1342161
Conc: 41.31 ng/ml




#20 AR-1242-5

R.T.: 6.428 min
Delta R.T.: 0.000 min
Response: 351853
Conc: 7.91 ng/ml



#20 AR-1242-5


R.T.: 5.462 min
Delta R.T.: -0.005 min
Response: 1399982
Conc: 37.46 ng/ml



#21 AR-1248-1


R.T.: 5.497 min
Delta R.T.: -0.005 min
Response: 803140
Conc: 19.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.654 min
Delta R.T.: -0.009 min
Response: 1504985
Conc: 48.54 ng/ml




#22 AR-1248-2

R.T.: 5.775 min
Delta R.T.: -0.002 min
Response: 863004
Conc: 13.57 ng/ml



#22 AR-1248-2


R.T.: 4.896 min
Delta R.T.: -0.005 min
Response: 1337711
Conc: 29.25 ng/ml



#23 AR-1248-3


R.T.: 5.970 min
Delta R.T.: -0.011 min
Response: 2508957
Conc: 36.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.956 min
Delta R.T.: 0.013 min
Response: 1342161
Conc: 28.15 ng/ml




#24 AR-1248-4

R.T.: 6.388 min
Delta R.T.: 0.000 min
Response: 369917
Conc: 4.88 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.114 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.428 min
Delta R.T.: 0.000 min
Response: 351853
Conc: 4.77 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.505 min
Delta R.T.: -0.003 min
Response: 1376946
Conc: 28.00 ng/ml




#26 AR-1254-1

R.T.: 6.350 min
Delta R.T.: -0.011 min
Response: 697215
Conc: 8.84 ng/ml



#26 AR-1254-1


R.T.: 5.462 min
Delta R.T.: -0.005 min
Response: 1399982
Conc: 17.34 ng/ml



#27 AR-1254-2


R.T.: 6.575 min
Delta R.T.: -0.006 min
Response: 2461063
Conc: 20.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.608 min
Delta R.T.: -0.007 min
Response: 1438277
Conc: 20.10 ng/ml




#28 AR-1254-3

R.T.: 6.973 min
Delta R.T.: 0.024 min
Response: 7314341
Conc: 59.71 ng/ml



#28 AR-1254-3


R.T.: 6.022 min
Delta R.T.: 0.002 min
Response: 1274336
Conc: 11.44 ng/ml



#29 AR-1254-4


R.T.: 7.227 min
Delta R.T.: -0.010 min
Response: 500431
Conc: 5.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.233 min
Delta R.T.: -0.017 min
Response: 1042630
Conc: 16.95 ng/ml




#30 AR-1254-5

R.T.: 7.671 min
Delta R.T.: 0.014 min
Response: 2471905
Conc: 24.44 ng/ml



#30 AR-1254-5


R.T.: 6.658 min
Delta R.T.: -0.010 min
Response: 883054
Conc: 9.31 ng/ml



#31 AR-1260-1


R.T.: 7.104 min
Delta R.T.: -0.012 min
Response: 1939531
Conc: 20.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.144 min
Delta R.T.: -0.008 min
Response: 2303071
Conc: 29.70 ng/ml




#32 AR-1260-2

R.T.: 7.351 min
Delta R.T.: -0.023 min
Response: 1187762
Conc: 10.55 ng/ml



#32 AR-1260-2


R.T.: 6.347 min
Delta R.T.: 0.005 min
Response: 2026322
Conc: 22.54 ng/ml



#33 AR-1260-3


R.T.: 7.728 min
Delta R.T.: -0.007 min
Response: 282953
Conc: 3.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.484 min
Delta R.T.: -0.011 min
Response: 508612
Conc: 5.99 ng/ml




#34 AR-1260-4

R.T.: 7.957 min
Delta R.T.: -0.004 min
Response: 1527871
Conc: 15.74 ng/ml



#34 AR-1260-4


R.T.: 6.955 min
Delta R.T.: -0.014 min
Response: 217661
Conc: 3.04 ng/ml



#35 AR-1260-5


R.T.: 8.271 min
Delta R.T.: -0.006 min
Response: 1226360
Conc: 6.90 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.198 min
Delta R.T.: -0.014 min
Response: 668264
Conc: 4.35 ng/ml




#36 AR-1262-1

R.T.: 7.728 min
Delta R.T.: -0.006 min
Response: 282953
Conc: 2.40 ng/ml



#36 AR-1262-1


R.T.: 6.753 min
Delta R.T.: -0.007 min
Response: 203236
Conc: 4.73 ng/ml



#37 AR-1262-2


R.T.: 8.271 min
Delta R.T.: -0.005 min
Response: 1226360
Conc: 6.09 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.955 min
Delta R.T.: -0.012 min
Response: 217661
Conc: 2.35 ng/ml




#38 AR-1262-3

R.T.: 8.570 min
Delta R.T.: -0.020 min
Response: 84299
Conc: 0.61 ng/ml



#38 AR-1262-3


R.T.: 7.482 min
Delta R.T.: -0.013 min
Response: 808808
Conc: 11.10 ng/ml



#39 AR-1262-4


R.T.: 8.673 min
Delta R.T.: -0.005 min
Response: 803721
Conc: 10.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.550 min
Delta R.T.: -0.009 min
Response: 880132
Conc: 6.73 ng/ml




#40 AR-1262-5

R.T.: 9.328 min
Delta R.T.: -0.006 min
Response: -18252
Conc: N.D.



#40 AR-1262-5


R.T.: 8.077 min
Delta R.T.: 0.023 min
Response: 77614
Conc: 1.31 ng/ml



#41 AR-1268-1


R.T.: 8.570 min
Delta R.T.: -0.015 min
Response: 84299
Conc: 0.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.482 min
Delta R.T.: -0.011 min
Response: 808808
Conc: 3.74 ng/ml




#42 AR-1268-2

R.T.: 8.673 min
Delta R.T.: -0.006 min
Response: 803721
Conc: 3.54 ng/ml



#42 AR-1268-2


R.T.: 7.550 min
Delta R.T.: -0.009 min
Response: 880132
Conc: 4.51 ng/ml



#43 AR-1268-3


R.T.: 8.884 min
Delta R.T.: -0.023 min
Response: 139381
Conc: 0.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.766 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.328 min
Delta R.T.: -0.005 min
Response: -18252
Conc: N.D.



#44 AR-1268-4


R.T.: 8.077 min
Delta R.T.: 0.023 min
Response: 77614
Conc: 1.22 ng/ml



#45 AR-1268-5

R.T.: 9.740 min
Delta R.T.: -0.009 min
Response: 367740
Conc: 0.58 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.335 min
Delta R.T.: -0.008 min
Response: 122449
Conc: 0.23 ng/ml