

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020323\
 Data File : PP055343.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Feb 2023 03:25
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 06:37:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.326	3.573	32938785	28728433	18.944	19.137
2)	SA Decachlor...	10.078	8.581	36707705	30225136	21.198	22.380
Target Compounds							
3)	L1 AR-1016-1	5.491	4.665	60681	278905	0.903	5.635 #
4)	L1 AR-1016-2	5.518	4.665f	96713	278905	1.009	4.029 #
5)	L1 AR-1016-3	5.572f	0.000	277132	0	4.633	N.D. #
6)	L1 AR-1016-4	5.690	4.905	18435	761105	0.385	24.222 #
7)	L1 AR-1016-5	5.969	5.131f	27923	684509	0.551	16.742 #
8)	L2 AR-1221-1	4.535	3.775f	85360	46481	3.630	2.344 #
9)	L2 AR-1221-2	4.633	3.871	556822	473518	31.340	31.946
10)	L2 AR-1221-3	4.693	3.944	335965	140200	6.244	3.108 #
11)	L3 AR-1232-1	4.693	3.944	335965	140200	7.570	3.804 #
12)	L3 AR-1232-2	5.244	4.665f	36449	278905	1.612	8.731 #
13)	L3 AR-1232-3	5.518	0.000	96713	0	2.202	N.D. #
14)	L3 AR-1232-4	5.690	4.943	18435	598081	0.833	37.088 #
15)	L3 AR-1232-5	5.765	5.131f	44716	684509	2.554	38.757 #
16)	L4 AR-1242-1	5.491	4.665	60681	278905	1.084	6.809 #
17)	L4 AR-1242-2	5.518	4.665f	96713	278905	1.211	4.884 #
18)	L4 AR-1242-3	5.572f	0.000	277132	0	5.580	N.D. #
19)	L4 AR-1242-4	5.690	4.943	18435	598081	0.460	18.410 #
20)	L4 AR-1242-5	6.418	5.481	1250774	3649735	28.130	97.660 #
21)	L5 AR-1248-1	5.491	4.665	60681	278905	1.455	8.996 #
22)	L5 AR-1248-2	5.765	4.905	44716	761105	0.703	16.644 #
23)	L5 AR-1248-3	5.969	4.943	27923	598081	0.404	12.545 #
24)	L5 AR-1248-4	6.359f	5.131f	2728162	684509	35.975	12.394 #
25)	L5 AR-1248-5	6.418	5.481f	1250774	3649735	16.950	74.226 #
26)	L6 AR-1254-1	6.359	5.481	2728162	3649735	34.581	45.193 #
27)	L6 AR-1254-2	6.582	5.622	15429	2770203	0.127	38.710 #
28)	L6 AR-1254-3	6.942	6.015	262071	1134710	2.140	10.185 #
29)	L6 AR-1254-4	7.233	6.241	20991	270341	0.238	4.395 #
30)	L6 AR-1254-5	0.000	6.698f	0	66612	N.D.	0.703 #
31)	L7 AR-1260-1	7.111	6.151	105475	651919	1.106	8.407 #
32)	L7 AR-1260-2	7.381	6.366f	970348	73295	8.622	0.815 #
33)	L7 AR-1260-3	0.000	6.470f	0	82386	N.D.	0.970 #
34)	L7 AR-1260-4	7.961	6.965	130624	26631	1.346	0.372 #
35)	L7 AR-1260-5	8.274	7.199	299938	53148	1.686	0.346 #
36)	L8 AR-1262-1	0.000	6.773	0	60327	N.D.	1.404 #
37)	L8 AR-1262-2	8.274	6.965	299938	26631	1.489	0.288 #
38)	L8 AR-1262-3	0.000	7.517f	0	321264	N.D.	4.409 #
39)	L8 AR-1262-4	8.681	7.541f	186048	132263	2.420	1.011 #
40)	L8 AR-1262-5	9.329	0.000	14580	0	0.204	N.D. #
41)	L9 AR-1268-1	0.000	7.517f	0	321264	N.D.	1.486 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020323\
 Data File : PP055343.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Feb 2023 03:25
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 06:37:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.681	7.541f	186048	132263	0.819	0.678
43) L9	AR-1268-3	8.911	0.000	113606	0	0.569	N.D. #
44) L9	AR-1268-4	9.329	0.000	14580	0	0.186	N.D. #
45) L9	AR-1268-5	9.739	8.333	161561	97962	0.254	0.186 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020323\
Data File : PP055343.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Feb 2023 03:25
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 06:37:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.326 min
Delta R.T.: -0.006 min
Response: 32938785
Conc: 18.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.573 min
Delta R.T.: -0.006 min
Response: 28728433
Conc: 19.14 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.078 min
Delta R.T.: -0.013 min
Response: 36707705
Conc: 21.20 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.581 min
Delta R.T.: -0.017 min
Response: 30225136
Conc: 22.38 ng/ml



#3 AR-1016-1


R.T.: 5.491 min
Delta R.T.: -0.012 min
Response: 60681
Conc: 0.90 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.665 min
Delta R.T.: 0.001 min
Response: 278905
Conc: 5.64 ng/ml




#4 AR-1016-2

R.T.: 5.518 min
Delta R.T.: -0.008 min
Response: 96713
Conc: 1.01 ng/ml



#4 AR-1016-2


R.T.: 4.665 min
Delta R.T.: -0.018 min
Response: 278905
Conc: 4.03 ng/ml



#5 AR-1016-3


R.T.: 5.572 min
Delta R.T.: -0.017 min
Response: 277132
Conc: 4.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.859 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.690 min
Delta R.T.: 0.003 min
Response: 18435
Conc: 0.38 ng/ml



#6 AR-1016-4


R.T.: 4.905 min
Delta R.T.: 0.003 min
Response: 761105
Conc: 24.22 ng/ml



#7 AR-1016-5


R.T.: 5.969 min
Delta R.T.: -0.015 min
Response: 27923
Conc: 0.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.131 min
Delta R.T.: 0.016 min
Response: 684509
Conc: 16.74 ng/ml




#8 AR-1221-1

R.T.: 4.535 min
Delta R.T.: -0.004 min
Response: 85360
Conc: 3.63 ng/ml



#8 AR-1221-1


R.T.: 3.775 min
Delta R.T.: -0.016 min
Response: 46481
Conc: 2.34 ng/ml



#9 AR-1221-2


R.T.: 4.633 min
Delta R.T.: 0.007 min
Response: 556822
Conc: 31.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.871 min
Delta R.T.: -0.006 min
Response: 473518
Conc: 31.95 ng/ml




#10 AR-1221-3

R.T.: 4.693 min
Delta R.T.: -0.009 min
Response: 335965
Conc: 6.24 ng/ml



#10 AR-1221-3


R.T.: 3.944 min
Delta R.T.: -0.008 min
Response: 140200
Conc: 3.11 ng/ml



#11 AR-1232-1


R.T.: 4.693 min
Delta R.T.: -0.011 min
Response: 335965
Conc: 7.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.944 min
Delta R.T.: -0.009 min
Response: 140200
Conc: 3.80 ng/ml




#12 AR-1232-2

R.T.: 5.244 min
Delta R.T.: 0.009 min
Response: 36449
Conc: 1.61 ng/ml



#12 AR-1232-2


R.T.: 4.665 min
Delta R.T.: -0.018 min
Response: 278905
Conc: 8.73 ng/ml



#13 AR-1232-3


R.T.: 5.518 min
Delta R.T.: -0.008 min
Response: 96713
Conc: 2.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.859 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.690 min
Delta R.T.: 0.002 min
Response: 18435
Conc: 0.83 ng/ml



#14 AR-1232-4


R.T.: 4.943 min
Delta R.T.: 0.000 min
Response: 598081
Conc: 37.09 ng/ml



#15 AR-1232-5


R.T.: 5.765 min
Delta R.T.: -0.015 min
Response: 44716
Conc: 2.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.131 min
Delta R.T.: 0.016 min
Response: 684509
Conc: 38.76 ng/ml




#16 AR-1242-1

R.T.: 5.491 min
Delta R.T.: -0.012 min
Response: 60681
Conc: 1.08 ng/ml



#16 AR-1242-1


R.T.: 4.665 min
Delta R.T.: 0.000 min
Response: 278905
Conc: 6.81 ng/ml



#17 AR-1242-2


R.T.: 5.518 min
Delta R.T.: -0.008 min
Response: 96713
Conc: 1.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.665 min
Delta R.T.: -0.018 min
Response: 278905
Conc: 4.88 ng/ml




#18 AR-1242-3

R.T.: 5.572 min
Delta R.T.: -0.017 min
Response: 277132
Conc: 5.58 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T.: 4.860 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.690 min
Delta R.T.: 0.003 min
Response: 18435
Conc: 0.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.943 min
Delta R.T.: 0.000 min
Response: 598081
Conc: 18.41 ng/ml




#20 AR-1242-5

R.T.: 6.418 min
Delta R.T.: -0.009 min
Response: 1250774
Conc: 28.13 ng/ml



#20 AR-1242-5


R.T.: 5.481 min
Delta R.T.: 0.013 min
Response: 3649735
Conc: 97.66 ng/ml



#21 AR-1248-1


R.T.: 5.491 min
Delta R.T.: -0.011 min
Response: 60681
Conc: 1.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.665 min
Delta R.T.: 0.001 min
Response: 278905
Conc: 9.00 ng/ml




#22 AR-1248-2

R.T.: 5.765 min
Delta R.T.: -0.013 min
Response: 44716
Conc: 0.70 ng/ml



#22 AR-1248-2


R.T.: 4.905 min
Delta R.T.: 0.004 min
Response: 761105
Conc: 16.64 ng/ml



#23 AR-1248-3


R.T.: 5.969 min
Delta R.T.: -0.013 min
Response: 27923
Conc: 0.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.943 min
Delta R.T.: 0.000 min
Response: 598081
Conc: 12.55 ng/ml




#24 AR-1248-4

R.T.: 6.359 min
Delta R.T.: -0.029 min
Response: 2728162
Conc: 35.98 ng/ml



#24 AR-1248-4


R.T.: 5.131 min
Delta R.T.: 0.017 min
Response: 684509
Conc: 12.39 ng/ml



#25 AR-1248-5


R.T.: 6.418 min
Delta R.T.: -0.009 min
Response: 1250774
Conc: 16.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.481 min
Delta R.T.: -0.027 min
Response: 3649735
Conc: 74.23 ng/ml




#26 AR-1254-1

R.T.: 6.359 min
Delta R.T.: -0.003 min
Response: 2728162
Conc: 34.58 ng/ml



#26 AR-1254-1


R.T.: 5.481 min
Delta R.T.: 0.014 min
Response: 3649735
Conc: 45.19 ng/ml



#27 AR-1254-2


R.T.: 6.582 min
Delta R.T.: 0.000 min
Response: 15429
Conc: 0.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.622 min
Delta R.T.: 0.006 min
Response: 2770203
Conc: 38.71 ng/ml




#28 AR-1254-3

R.T.: 6.942 min
Delta R.T.: -0.007 min
Response: 262071
Conc: 2.14 ng/ml



#28 AR-1254-3


R.T.: 6.015 min
Delta R.T.: -0.005 min
Response: 1134710
Conc: 10.19 ng/ml



#29 AR-1254-4


R.T.: 7.233 min
Delta R.T.: -0.004 min
Response: 20991
Conc: 0.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.241 min
Delta R.T.: -0.008 min
Response: 270341
Conc: 4.40 ng/ml




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.657 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 6.698 min
Delta R.T.: 0.029 min
Response: 66612
Conc: 0.70 ng/ml



#31 AR-1260-1


R.T.: 7.111 min
Delta R.T.: -0.005 min
Response: 105475
Conc: 1.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.151 min
Delta R.T.: -0.001 min
Response: 651919
Conc: 8.41 ng/ml




#32 AR-1260-2

R.T.: 7.381 min
Delta R.T.: 0.008 min
Response: 970348
Conc: 8.62 ng/ml



#32 AR-1260-2


R.T.: 6.366 min
Delta R.T.: 0.025 min
Response: 73295
Conc: 0.82 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T.: 7.735 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.470 min
Delta R.T.: -0.025 min
Response: 82386
Conc: 0.97 ng/ml




#34 AR-1260-4

R.T.: 7.961 min
Delta R.T.: 0.000 min
Response: 130624
Conc: 1.35 ng/ml



#34 AR-1260-4


R.T.: 6.965 min
Delta R.T.: -0.004 min
Response: 26631
Conc: 0.37 ng/ml



#35 AR-1260-5


R.T.: 8.274 min
Delta R.T.: -0.004 min
Response: 299938
Conc: 1.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.199 min
Delta R.T.: -0.013 min
Response: 53148
Conc: 0.35 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 7.734 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 6.773 min
Delta R.T.: 0.013 min
Response: 60327
Conc: 1.40 ng/ml



#37 AR-1262-2


R.T.: 8.274 min
Delta R.T.: -0.003 min
Response: 299938
Conc: 1.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.965 min
Delta R.T.: -0.002 min
Response: 26631
Conc: 0.29 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 8.589 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.517 min
Delta R.T.: 0.022 min
Response: 321264
Conc: 4.41 ng/ml



#39 AR-1262-4


R.T.: 8.681 min
Delta R.T.: 0.003 min
Response: 186048
Conc: 2.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.541 min
Delta R.T.: -0.018 min
Response: 132263
Conc: 1.01 ng/ml




#40 AR-1262-5

R.T.: 9.329 min
Delta R.T.: -0.005 min
Response: 14580
Conc: 0.20 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T.: 8.054 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 8.585 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.517 min
Delta R.T.: 0.024 min
Response: 321264
Conc: 1.49 ng/ml




#42 AR-1268-2

R.T.: 8.681 min
Delta R.T.: 0.002 min
Response: 186048
Conc: 0.82 ng/ml



#42 AR-1268-2


R.T.: 7.541 min
Delta R.T.: -0.018 min
Response: 132263
Conc: 0.68 ng/ml



#43 AR-1268-3


R.T.: 8.911 min
Delta R.T.: 0.004 min
Response: 113606
Conc: 0.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.766 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.329 min
Delta R.T.: -0.004 min
Response: 14580
Conc: 0.19 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 8.054 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.739 min
Delta R.T.: -0.010 min
Response: 161561
Conc: 0.25 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.333 min
Delta R.T.: -0.010 min
Response: 97962
Conc: 0.19 ng/ml