

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020325\
 Data File : PP069412.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Feb 2025 17:01
 Operator : YP\AJ
 Sample : Q1267-01
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TRE-25-0003

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/04/2025
 Supervised By :Ankita Jodhani 02/04/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 00:37:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 29 00:13:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.535	3.840	34410745	23592467	24.039	26.433
2) SA Decachlor...	10.271	8.906	27948271	26575312	25.573	23.760
Target Compounds						
16) L4 AR-1242-1	5.688	4.933	11352957	9566148	272.219	331.789
17) L4 AR-1242-2	5.710	4.950	21368437	11913588	362.233	302.393
18) L4 AR-1242-3	5.772	5.128	12303906	6505457	331.654	289.079m
19) L4 AR-1242-4	5.870	5.212	13063904	7911309	398.355	366.566m
20) L4 AR-1242-5	6.601	5.739	13826472	16171708	424.329	562.135 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020325\
Data File : PP069412.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Feb 2025 17:01
Operator : YP\AJ
Sample : Q1267-01
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
TRE-25-0003


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/04/2025
Supervised By :Ankita Jodhani 02/04/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 00:37:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 29 00:13:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.535 min
Delta R.T.: -0.003 min
Response: 34410745
Conc: 24.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0003


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/04/2025
Supervised By :Ankita Jodhani 02/04/2025




#1 Tetrachloro-m-xylene

R.T.: 3.840 min
Delta R.T.: 0.000 min
Response: 23592467
Conc: 26.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.271 min
Delta R.T.: -0.008 min
Response: 27948271
Conc: 25.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.906 min
Delta R.T.: -0.005 min
Response: 26575312
Conc: 23.76 ng/ml




#16 AR-1242-1

R.T.: 5.688 min
Delta R.T.: -0.004 min
Response: 11352957
Conc: 272.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0003


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/04/2025
Supervised By :Ankita Jodhani 02/04/2025




#16 AR-1242-1

R.T.: 4.933 min
Delta R.T.: 0.000 min
Response: 9566148
Conc: 331.79 ng/ml




#17 AR-1242-2

R.T.: 5.710 min
Delta R.T.: -0.004 min
Response: 21368437
Conc: 362.23 ng/ml



#17 AR-1242-2

R.T.: 4.950 min
Delta R.T.: -0.003 min
Response: 11913588
Conc: 302.39 ng/ml




#18 AR-1242-3

R.T.: 5.772 min
Delta R.T.: -0.004 min
Response: 12303906
Conc: 331.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0003


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/04/2025
Supervised By :Ankita Jodhani 02/04/2025




#18 AR-1242-3

R.T.: 5.128 min
Delta R.T.: -0.002 min
Response: 6505457
Conc: 289.08 ng/ml m




#19 AR-1242-4

R.T.: 5.870 min
Delta R.T.: -0.004 min
Response: 13063904
Conc: 398.35 ng/ml



#19 AR-1242-4

R.T.: 5.212 min
Delta R.T.: -0.003 min
Response: 7911309
Conc: 366.57 ng/ml m




#20 AR-1242-5

R.T.: 6.601 min
Delta R.T.: -0.003 min
Response: 13826472
Conc: 424.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-25-0003

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/04/2025
Supervised By :Ankita Jodhani 02/04/2025



#20 AR-1242-5

R.T.: 5.739 min
Delta R.T.: -0.002 min
Response: 16171708
Conc: 562.14 ng/ml