

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020325\
 Data File : PP069427.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Feb 2025 21:37
 Operator : YP\AJ
 Sample : Q1261-01
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 CHRT-20430

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/05/2025
 Supervised By :Ankita Jodhani 02/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 00:42:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 29 00:13:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.538	3.872f	39307245	20764610	27.459m	23.265m
2) SA Decachlor...	10.270	8.906	18045198	16707154	16.511	14.937m
Target Compounds						
26) L6 AR-1254-1	6.550	5.754	17917211	17817248	309.645m	323.622m
27) L6 AR-1254-2	6.766	5.903	26053760	14720033	324.690m	304.329m
28) L6 AR-1254-3	7.126	6.305	27590112	26144096	327.330m	344.472
29) L6 AR-1254-4	7.409	6.532	23160499	17430963	339.657	340.118
30) L6 AR-1254-5	7.821	6.947	32054183	27637828	471.065m	415.229
31) L7 AR-1260-1	7.292	6.435	16662286	16026626	285.136	332.299
32) L7 AR-1260-2	7.543	6.621	21644254	15943056	278.158	260.067
33) L7 AR-1260-3	7.899	6.773	8944610	15221022	143.070	253.024 #
34) L7 AR-1260-4	8.111	7.244	18149970	5614521	275.096	120.826 #
35) L7 AR-1260-5	8.443	7.483	18796267	20720194	143.564	186.425 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020325\
Data File : PP069427.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Feb 2025 21:37
Operator : YP\AJ
Sample : Q1261-01
Misc :
ALS Vial : 33 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
CHRT-20430


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/05/2025
Supervised By :Ankita Jodhani 02/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 00:42:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 29 00:13:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.538 min
Delta R.T.: 0.000 min
Response: 39307245
Conc: 27.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
CHRT-20430


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/05/2025
Supervised By :Ankita Jodhani 02/05/2025




#1 Tetrachloro-m-xylene

R.T.: 3.872 min
Delta R.T.: 0.032 min
Response: 20764610
Conc: 23.26 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.270 min
Delta R.T.: -0.009 min
Response: 18045198
Conc: 16.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.906 min
Delta R.T.: -0.005 min
Response: 16707154
Conc: 14.94 ng/ml




#26 AR-1254-1

R.T.: 6.550 min
Delta R.T.: 0.010 min
Response: 17917211
Conc: 309.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
CHRT-20430


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/05/2025
Supervised By :Ankita Jodhani 02/05/2025




#26 AR-1254-1

R.T.: 5.754 min
Delta R.T.: 0.013 min
Response: 17817248
Conc: 323.62 ng/ml m




#27 AR-1254-2

R.T.: 6.766 min
Delta R.T.: 0.009 min
Response: 26053760
Conc: 324.69 ng/ml m



#27 AR-1254-2

R.T.: 5.903 min
Delta R.T.: 0.013 min
Response: 14720033
Conc: 304.33 ng/ml m




#28 AR-1254-3

R.T.: 7.126 min
Delta R.T.: 0.006 min
Response: 27590112
Conc: 327.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
CHRT-20430


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/05/2025
Supervised By :Ankita Jodhani 02/05/2025




#28 AR-1254-3

R.T.: 6.305 min
Delta R.T.: 0.010 min
Response: 26144096
Conc: 344.47 ng/ml




#29 AR-1254-4

R.T.: 7.409 min
Delta R.T.: 0.006 min
Response: 23160499
Conc: 339.66 ng/ml



#29 AR-1254-4

R.T.: 6.532 min
Delta R.T.: 0.008 min
Response: 17430963
Conc: 340.12 ng/ml




#30 AR-1254-5

R.T.: 7.821 min
Delta R.T.: 0.001 min
Response: 32054183
Conc: 471.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
CHRT-20430


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/05/2025
Supervised By :Ankita Jodhani 02/05/2025




#30 AR-1254-5

R.T.: 6.947 min
Delta R.T.: 0.004 min
Response: 27637828
Conc: 415.23 ng/ml




#31 AR-1260-1

R.T.: 7.292 min
Delta R.T.: 0.005 min
Response: 16662286
Conc: 285.14 ng/ml



#31 AR-1260-1

R.T.: 6.435 min
Delta R.T.: 0.009 min
Response: 16026626
Conc: 332.30 ng/ml




#32 AR-1260-2

R.T.: 7.543 min
Delta R.T.: 0.002 min
Response: 21644254
Conc: 278.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
CHRT-20430


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/05/2025
Supervised By :Ankita Jodhani 02/05/2025




#32 AR-1260-2

R.T.: 6.621 min
Delta R.T.: 0.006 min
Response: 15943056
Conc: 260.07 ng/ml




#33 AR-1260-3

R.T.: 7.899 min
Delta R.T.: 0.000 min
Response: 8944610
Conc: 143.07 ng/ml



#33 AR-1260-3

R.T.: 6.773 min
Delta R.T.: 0.003 min
Response: 15221022
Conc: 253.02 ng/ml




#34 AR-1260-4

R.T.: 8.111 min
Delta R.T.: -0.015 min
Response: 18149970
Conc: 275.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
CHRT-20430


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/05/2025
Supervised By :Ankita Jodhani 02/05/2025




#34 AR-1260-4

R.T.: 7.244 min
Delta R.T.: 0.002 min
Response: 5614521
Conc: 120.83 ng/ml



#35 AR-1260-5

R.T.: 8.443 min
Delta R.T.: -0.004 min
Response: 18796267
Conc: 143.56 ng/ml



#35 AR-1260-5

R.T.: 7.483 min
Delta R.T.: 0.000 min
Response: 20720194
Conc: 186.42 ng/ml