

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020325\
 Data File : PP069432.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Feb 2025 23:30
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 00:44:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 29 00:13:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.536	3.840	73266597	46614150	51.182	52.226
2) SA Decachlor...	10.271	8.907	48790358	46683420	44.643	41.738
Target Compounds						
16) L4 AR-1242-1	5.689	4.932	20257784	14779963	485.737	512.623
17) L4 AR-1242-2	5.711	4.951	30060415	20385735	509.577	517.434
18) L4 AR-1242-3	5.773	5.129	19336352	11474631	521.215	509.891
19) L4 AR-1242-4	5.870	5.213	15751204	10929958	480.298	506.433
20) L4 AR-1242-5	6.601	5.739	14839206	14669309	455.409	509.911



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020325\
Data File : PP069432.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Feb 2025 23:30
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 00:44:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 29 00:13:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.536 min
Delta R.T.: -0.002 min
Response: 73266597
Conc: 51.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.840 min
Delta R.T.: 0.000 min
Response: 46614150
Conc: 52.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.271 min
Delta R.T.: -0.008 min
Response: 48790358
Conc: 44.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.907 min
Delta R.T.: -0.004 min
Response: 46683420
Conc: 41.74 ng/ml



#16 AR-1242-1


R.T.: 5.689 min
Delta R.T.: -0.004 min
Response: 20257784
Conc: 485.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.932 min
Delta R.T.: 0.000 min
Response: 14779963
Conc: 512.62 ng/ml




#17 AR-1242-2

R.T.: 5.711 min
Delta R.T.: -0.003 min
Response: 30060415
Conc: 509.58 ng/ml



#17 AR-1242-2


R.T.: 4.951 min
Delta R.T.: -0.001 min
Response: 20385735
Conc: 517.43 ng/ml



#18 AR-1242-3


R.T.: 5.773 min
Delta R.T.: -0.003 min
Response: 19336352
Conc: 521.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.129 min
Delta R.T.: 0.000 min
Response: 11474631
Conc: 509.89 ng/ml




#19 AR-1242-4

R.T.: 5.870 min
Delta R.T.: -0.004 min
Response: 15751204
Conc: 480.30 ng/ml



#19 AR-1242-4


R.T.: 5.213 min
Delta R.T.: -0.001 min
Response: 10929958
Conc: 506.43 ng/ml



#20 AR-1242-5

R.T.: 6.601 min
Delta R.T.: -0.004 min
Response: 14839206
Conc: 455.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.739 min
Delta R.T.: -0.002 min
Response: 14669309
Conc: 509.91 ng/ml