

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020422\
 Data File : PP043494.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Feb 2022 10:50
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 23:46:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Feb 03 10:01:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.876	4.027	1298250	1112814	47.304	49.134
2) SA Decachlor...	10.798f	9.354	769548	987867	50.166	52.256
Target Compounds						
26) L6 AR-1254-1	7.085	6.168	379179	611843	438.109	456.744
27) L6 AR-1254-2	7.313	6.327	574754	533265	447.832	457.401
28) L6 AR-1254-3	7.691	6.751	596855	851146	464.924	462.343
29) L6 AR-1254-4	7.984	6.991	404830	523190	471.174	473.524
30) L6 AR-1254-5	8.412	7.421	443562	734754	477.821	475.957



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020422\
Data File : PP043494.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Feb 2022 10:50
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 23:46:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Feb 03 10:01:05 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.876 min
Delta R.T.: 0.011 min
Response: 1298250
Conc: 47.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.027 min
Delta R.T.: 0.003 min
Response: 1112814
Conc: 49.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.798 min
Delta R.T.: 0.018 min
Response: 769548
Conc: 50.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.354 min
Delta R.T.: 0.008 min
Response: 987867
Conc: 52.26 ng/ml



#26 AR-1254-1


R.T.: 7.085 min
Delta R.T.: 0.011 min
Response: 379179
Conc: 438.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.168 min
Delta R.T.: 0.004 min
Response: 611843
Conc: 456.74 ng/ml




#27 AR-1254-2

R.T.: 7.313 min
Delta R.T.: 0.011 min
Response: 574754
Conc: 447.83 ng/ml



#27 AR-1254-2


R.T.: 6.327 min
Delta R.T.: 0.004 min
Response: 533265
Conc: 457.40 ng/ml



#28 AR-1254-3


R.T.: 7.691 min
Delta R.T.: 0.012 min
Response: 596855
Conc: 464.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.751 min
Delta R.T.: 0.004 min
Response: 851146
Conc: 462.34 ng/ml




#29 AR-1254-4

R.T.: 7.984 min
Delta R.T.: 0.011 min
Response: 404830
Conc: 471.17 ng/ml



#29 AR-1254-4


R.T.: 6.991 min
Delta R.T.: 0.004 min
Response: 523190
Conc: 473.52 ng/ml



#30 AR-1254-5

R.T.: 8.412 min
Delta R.T.: 0.012 min
Response: 443562
Conc: 477.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.421 min
Delta R.T.: 0.004 min
Response: 734754
Conc: 475.96 ng/ml