

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020422\
 Data File : PP043521.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Feb 2022 19:24
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 23:56:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Feb 03 10:01:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.869	4.027	1274925	1101446	46.454	48.632
2) SA Decachlor...	10.786	9.352	794056	996306	51.764	52.702
Target Compounds						
26) L6 AR-1254-1	7.077	6.166	383392	613268	442.977	457.808
27) L6 AR-1254-2	7.305	6.326	582366	533209	453.763	457.353
28) L6 AR-1254-3	7.683	6.749	614107	856448	478.363	465.223
29) L6 AR-1254-4	7.976	6.989	418636	529180	487.242	478.944
30) L6 AR-1254-5	8.405	7.420	459052	744831	494.507	482.484



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020422\
Data File : PP043521.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Feb 2022 19:24
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 23:56:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Feb 03 10:01:05 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.869 min
Delta R.T.: 0.004 min
Response: 1274925
Conc: 46.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.027 min
Delta R.T.: 0.003 min
Response: 1101446
Conc: 48.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.786 min
Delta R.T.: 0.007 min
Response: 794056
Conc: 51.76 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.352 min
Delta R.T.: 0.005 min
Response: 996306
Conc: 52.70 ng/ml



#26 AR-1254-1


R.T.: 7.077 min
Delta R.T.: 0.004 min
Response: 383392
Conc: 442.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.166 min
Delta R.T.: 0.002 min
Response: 613268
Conc: 457.81 ng/ml




#27 AR-1254-2

R.T.: 7.305 min
Delta R.T.: 0.004 min
Response: 582366
Conc: 453.76 ng/ml



#27 AR-1254-2


R.T.: 6.326 min
Delta R.T.: 0.002 min
Response: 533209
Conc: 457.35 ng/ml



#28 AR-1254-3


R.T.: 7.683 min
Delta R.T.: 0.004 min
Response: 614107
Conc: 478.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.749 min
Delta R.T.: 0.003 min
Response: 856448
Conc: 465.22 ng/ml




#29 AR-1254-4

R.T.: 7.976 min
Delta R.T.: 0.004 min
Response: 418636
Conc: 487.24 ng/ml



#29 AR-1254-4


R.T.: 6.989 min
Delta R.T.: 0.002 min
Response: 529180
Conc: 478.94 ng/ml



#30 AR-1254-5

R.T.: 8.405 min
Delta R.T.: 0.005 min
Response: 459052
Conc: 494.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.420 min
Delta R.T.: 0.003 min
Response: 744831
Conc: 482.48 ng/ml