

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020425\
 Data File : PP069483.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Feb 2025 17:48
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 01:31:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 29 00:13:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.535	3.838	75871148	46853879	53.002	52.495
2) SA Decachlor...	10.268	8.903	52991236	53052750	48.487	47.433
Target Compounds						
16) L4 AR-1242-1	5.687	4.930	20769788	14802753	498.013	513.414
17) L4 AR-1242-2	5.709	4.949	30288864	20560600	513.449	521.873
18) L4 AR-1242-3	5.772	5.126	18815093	11453704	507.164	508.961
19) L4 AR-1242-4	5.869	5.211	16070297	11147873	490.028	516.530
20) L4 AR-1242-5	6.600	5.736	16578776	14747772	508.796	512.638



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020425\
Data File : PP069483.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Feb 2025 17:48
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 01:31:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 29 00:13:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.535 min
Delta R.T.: -0.003 min
Response: 75871148
Conc: 53.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.838 min
Delta R.T.: -0.002 min
Response: 46853879
Conc: 52.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.268 min
Delta R.T.: -0.011 min
Response: 52991236
Conc: 48.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.903 min
Delta R.T.: -0.009 min
Response: 53052750
Conc: 47.43 ng/ml



#16 AR-1242-1


R.T.: 5.687 min
Delta R.T.: -0.005 min
Response: 20769788
Conc: 498.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.930 min
Delta R.T.: -0.004 min
Response: 14802753
Conc: 513.41 ng/ml




#17 AR-1242-2

R.T.: 5.709 min
Delta R.T.: -0.005 min
Response: 30288864
Conc: 513.45 ng/ml



#17 AR-1242-2


R.T.: 4.949 min
Delta R.T.: -0.004 min
Response: 20560600
Conc: 521.87 ng/ml



#18 AR-1242-3


R.T.: 5.772 min
Delta R.T.: -0.005 min
Response: 18815093
Conc: 507.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.126 min
Delta R.T.: -0.003 min
Response: 11453704
Conc: 508.96 ng/ml




#19 AR-1242-4

R.T.: 5.869 min
Delta R.T.: -0.005 min
Response: 16070297
Conc: 490.03 ng/ml



#19 AR-1242-4


R.T.: 5.211 min
Delta R.T.: -0.004 min
Response: 11147873
Conc: 516.53 ng/ml



#20 AR-1242-5

R.T.: 6.600 min
Delta R.T.: -0.005 min
Response: 16578776
Conc: 508.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.736 min
Delta R.T.: -0.004 min
Response: 14747772
Conc: 512.64 ng/ml