

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020524\
 Data File : PP063262.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Feb 2024 12:05
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP020524

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 12:28:29 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 05 12:13:11 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.491	3.646	110.4E6	110.9E6	46.175	46.653
2) SA Decachlor...	10.363	8.695	109.9E6	116.5E6	48.855	47.028
Target Compounds						
26) L6 AR-1254-1	6.537	5.547	44703649	61234272	500.956	506.291
27) L6 AR-1254-2	6.758	5.695	66217805	54974619	502.127	508.589
28) L6 AR-1254-3	7.126	6.102	69395787	86945434	504.813	514.093
29) L6 AR-1254-4	7.415	6.330	41831299	41514233	509.080	520.201
30) L6 AR-1254-5	7.837	6.751	51965668	73209727	504.792	519.651



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020524\
Data File : PP063262.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Feb 2024 12:05
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP020524

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 12:28:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020124.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 05 12:13:11 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.491 min
Delta R.T.: 0.001 min
Response: 110448004
Conc: 46.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP020524




#1 Tetrachloro-m-xylene

R.T.: 3.646 min
Delta R.T.: 0.000 min
Response: 110888402
Conc: 46.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.363 min
Delta R.T.: 0.004 min
Response: 109865807
Conc: 48.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: -0.009 min
Response: 116493487
Conc: 47.03 ng/ml



#26 AR-1254-1


R.T.: 6.537 min
Delta R.T.: 0.000 min
Response: 44703649
Conc: 500.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP020524




#26 AR-1254-1

R.T.: 5.547 min
Delta R.T.: 0.000 min
Response: 61234272
Conc: 506.29 ng/ml




#27 AR-1254-2

R.T.: 6.758 min
Delta R.T.: 0.000 min
Response: 66217805
Conc: 502.13 ng/ml



#27 AR-1254-2


R.T.: 5.695 min
Delta R.T.: 0.000 min
Response: 54974619
Conc: 508.59 ng/ml



#28 AR-1254-3


R.T.: 7.126 min
Delta R.T.: 0.000 min
Response: 69395787
Conc: 504.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP020524




#28 AR-1254-3

R.T.: 6.102 min
Delta R.T.: 0.000 min
Response: 86945434
Conc: 514.09 ng/ml




#29 AR-1254-4

R.T.: 7.415 min
Delta R.T.: 0.000 min
Response: 41831299
Conc: 509.08 ng/ml



#29 AR-1254-4


R.T.: 6.330 min
Delta R.T.: 0.000 min
Response: 41514233
Conc: 520.20 ng/ml



#30 AR-1254-5

R.T.: 7.837 min
Delta R.T.: 0.000 min
Response: 51965668
Conc: 504.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP020524



#30 AR-1254-5

R.T.: 6.751 min
Delta R.T.: 0.000 min
Response: 73209727
Conc: 519.65 ng/ml