

Method Path : P:\HPCHEM1\ECD_P\methods\
 Method File : PP020618.M
 Title : 8080.M
 Last Update : Tue Feb 06 02:07:55 2018
 Response Via : Initial Calibration

Calibration Files

750 =PP023464.D 1000 =PP023465.D 1500 =PP023466.D
 500 =PP023463.D 200 =PP023462.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	8.218	7.848	7.651	7.800	8.388	7.981	E5	3.87
2) T	3,5-DICHLOROBENZOI...	7.877	8.790	7.165	7.541	8.442	7.963	E5	8.27
3) T	4-Nitrophenol	3.504	7.517	3.247	3.377	3.854	4.300	E5	42.16
4) S	2,4-DCAA	5.179	5.110	4.707	4.963	5.664	5.125	E5	6.86
5) T	DICAMBA	2.227	2.121	2.000	2.118	2.286	2.150	E6	5.14
6) T	MCPP	1.428	1.954	1.359	1.312	1.278	1.466	E6	18.98
7) T	MCPA	2.149	2.749	2.018	2.004	2.094	2.203	E6	14.12
8) T	DICHLORPROP	5.369	5.326	4.957	5.133	5.774	5.312	E5	5.77
9) T	2,4-D	6.799	6.639	6.205	6.467	7.366	6.695	E5	6.49
10) T	Pentachlorophenol	8.342	7.715	6.878	8.340	9.371	8.129	E6	11.29
11) T	2,4,5-TP (SILVEX)	3.443	3.194	3.001	3.315	3.609	3.312	E6	7.02
12) T	2,4,5-T	3.251	3.018	2.844	3.119	3.390	3.124	E6	6.72
13) T	2,4-DB	4.040	4.006	3.710	3.808	4.095	3.932	E5	4.18
14) T	DINOSEB	2.385	2.246	2.133	2.267	2.445	2.295	E6	5.34
15) T	Picloram	5.095	4.694	4.393	4.885	5.276	4.869	E6	7.07
16) T	DCPA	3.994	4.142	3.463	3.844	4.164	3.921	E6	7.31

Signal #2 Calibration Files

750 =PP023464.D 1000 =PP023465.D 1500 =PP023466.D
 500 =PP023463.D 200 =PP023462.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	8.449	8.035	7.848	8.049	9.005	8.277	E5	5.58
2) T	3,5-DICHLOROBENZOI...	8.159	9.226	7.450	7.883	8.985	8.341	E5	8.96
3) T	4-Nitrophenol	3.724	8.168	3.482	3.621	4.209	4.641	E5	42.90
4) S	2,4-DCAA	6.100	6.033	5.578	5.867	6.742	6.064	E5	7.08
5) T	DICAMBA	2.339	2.235	2.107	2.241	2.434	2.271	E6	5.41
6) T	MCPP	1.486	2.204	1.418	1.421	1.516	1.609	E6	20.85
7) T	MCPA	2.221	2.915	2.093	2.054	2.360	2.329	E6	14.99
8) T	DICHLORPROP	5.840	5.917	5.492	5.631	6.476	5.871	E5	6.43
9) T	2,4-D	7.744	7.684	7.055	7.465	8.439	7.677	E5	6.57
10) T	Pentachlorophenol	0.874	0.815	0.721	0.880	1.014	0.861	E7	12.39
11) T	2,4,5-TP (SILVEX)	3.671	3.452	3.198	3.572	3.995	3.578	E6	8.18
12) T	2,4,5-T	3.125	3.171	2.738	3.017	3.352	3.081	E6	7.35
13) T	2,4-DB	4.365	4.413	4.022	4.155	4.660	4.323	E5	5.69
14) T	DINOSEB	2.521	2.417	2.248	2.427	2.686	2.460	E6	6.51
15) T	Picloram	5.138	4.796	4.417	5.018	5.537	4.981	E6	8.33
16) T	DCPA	4.371	4.545	3.761	4.281	4.798	4.351	E6	8.83

(#) = Out of Range