

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020623\
 Data File : PP055348.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Feb 2023 10:01
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 06 21:26:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.328	3.574	80854490	71908932	46.502	47.901
2) SA Decachlor...	10.083	8.584	84653144	70929683	48.885	52.519
Target Compounds						
26) L6 AR-1254-1	6.357	5.460	35022049	40918015	443.922	506.674
27) L6 AR-1254-2	6.577	5.608	53735686	35941248	444.012	502.235
28) L6 AR-1254-3	6.945	6.012	55093225	57411750	449.777	515.331
29) L6 AR-1254-4	7.232	6.242	39924439	31579070	453.110	513.419
30) L6 AR-1254-5	7.653	6.661	44359223	48302726	438.502	509.501



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020623\
Data File : PP055348.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Feb 2023 10:01
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 06 21:26:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.328 min
Delta R.T.: -0.004 min
Response: 80854490
Conc: 46.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.574 min
Delta R.T.: -0.004 min
Response: 71908932
Conc: 47.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.083 min
Delta R.T.: -0.008 min
Response: 84653144
Conc: 48.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.584 min
Delta R.T.: -0.014 min
Response: 70929683
Conc: 52.52 ng/ml



#26 AR-1254-1


R.T.: 6.357 min
Delta R.T.: -0.004 min
Response: 35022049
Conc: 443.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.460 min
Delta R.T.: -0.007 min
Response: 40918015
Conc: 506.67 ng/ml




#27 AR-1254-2

R.T.: 6.577 min
Delta R.T.: -0.004 min
Response: 53735686
Conc: 444.01 ng/ml



#27 AR-1254-2


R.T.: 5.608 min
Delta R.T.: -0.007 min
Response: 35941248
Conc: 502.23 ng/ml



#28 AR-1254-3


R.T.: 6.945 min
Delta R.T.: -0.004 min
Response: 55093225
Conc: 449.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.012 min
Delta R.T.: -0.008 min
Response: 57411750
Conc: 515.33 ng/ml




#29 AR-1254-4

R.T.: 7.232 min
Delta R.T.: -0.004 min
Response: 39924439
Conc: 453.11 ng/ml



#29 AR-1254-4


R.T.: 6.242 min
Delta R.T.: -0.008 min
Response: 31579070
Conc: 513.42 ng/ml



#30 AR-1254-5

R.T.: 7.653 min
Delta R.T.: -0.004 min
Response: 44359223
Conc: 438.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.661 min
Delta R.T.: -0.008 min
Response: 48302726
Conc: 509.50 ng/ml