

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020623\
 Data File : PP055351.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Feb 2023 10:50
 Operator : YP\AJ
 Sample : 01391-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 COMP2-0201

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 02/07/2023
 Supervised By :mohammad ahmed 02/07/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 06 21:32:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.328	3.577	14867320	20167312	8.551m	13.434 #
2) SA Decachlor...	10.083	8.586	16995526	14499592	9.815	10.736

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020623\
Data File : PP055351.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Feb 2023 10:50
Operator : YP\AJ
Sample : 01391-04
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
COMP2-0201


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/07/2023
Supervised By :mohammad ahmed 02/07/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 06 21:32:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.328 min
Delta R.T.: -0.004 min
Response: 14867320
Conc: 8.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
COMP2-0201


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/07/2023
Supervised By :mohammad ahmed 02/07/2023




#1 Tetrachloro-m-xylene

R.T.: 3.577 min
Delta R.T.: -0.002 min
Response: 20167312
Conc: 13.43 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.083 min
Delta R.T.: -0.008 min
Response: 16995526
Conc: 9.81 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.586 min
Delta R.T.: -0.012 min
Response: 14499592
Conc: 10.74 ng/ml