

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020625\
 Data File : PP069566.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Feb 2025 16:54
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 07 02:03:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 29 00:13:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.536	3.838	74697377	48263061	52.182	54.074
2) SA Decachlor...	10.272	8.902	53703081	53878153	49.138	48.171
Target Compounds						
21) L5 AR-1248-1	5.689	4.929	15767353	11173997	486.061	507.665
22) L5 AR-1248-2	5.962	5.167	21569802	14645498	508.166	487.135
23) L5 AR-1248-3	6.164	5.210	23474019	15391195	499.584	496.742
24) L5 AR-1248-4	6.563	5.383	27899150	19390727	506.141	524.872
25) L5 AR-1248-5	6.602	5.778	26469178	19224519	495.828	514.170



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020625\
Data File : PP069566.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Feb 2025 16:54
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 07 02:03:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 29 00:13:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.536 min
Delta R.T.: -0.002 min
Response: 74697377
Conc: 52.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.838 min
Delta R.T.: -0.002 min
Response: 48263061
Conc: 54.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.272 min
Delta R.T.: -0.007 min
Response: 53703081
Conc: 49.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.902 min
Delta R.T.: -0.009 min
Response: 53878153
Conc: 48.17 ng/ml



#21 AR-1248-1


R.T.: 5.689 min
Delta R.T.: 0.000 min
Response: 15767353
Conc: 486.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.929 min
Delta R.T.: -0.005 min
Response: 11173997
Conc: 507.67 ng/ml




#22 AR-1248-2

R.T.: 5.962 min
Delta R.T.: 0.000 min
Response: 21569802
Conc: 508.17 ng/ml



#22 AR-1248-2


R.T.: 5.167 min
Delta R.T.: -0.005 min
Response: 14645498
Conc: 487.13 ng/ml



#23 AR-1248-3


R.T.: 6.164 min
Delta R.T.: -0.001 min
Response: 23474019
Conc: 499.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.210 min
Delta R.T.: -0.005 min
Response: 15391195
Conc: 496.74 ng/ml




#24 AR-1248-4

R.T.: 6.563 min
Delta R.T.: -0.001 min
Response: 27899150
Conc: 506.14 ng/ml



#24 AR-1248-4


R.T.: 5.383 min
Delta R.T.: -0.005 min
Response: 19390727
Conc: 524.87 ng/ml



#25 AR-1248-5

R.T.: 6.602 min
Delta R.T.: -0.001 min
Response: 26469178
Conc: 495.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.778 min
Delta R.T.: -0.006 min
Response: 19224519
Conc: 514.17 ng/ml