

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020725\
 Data File : PP069600.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Feb 2025 10:18
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/10/2025
 Supervised By :Ankita Jodhani 02/10/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 07 13:52:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 29 00:13:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.535	3.837	80692391	51803654	56.370	58.041
2) SA Decachlor...	10.269	8.901	58805925	55487193	53.807	49.609
Target Compounds						
26) L6 AR-1254-1	6.537	5.736	32837460	29746575	567.497m	540.299
27) L6 AR-1254-2	6.755	5.884	43669337	25879466	544.221	535.045
28) L6 AR-1254-3	7.117	6.289	45517821	39975694	540.025	526.716
29) L6 AR-1254-4	7.400	6.518	36884288	26547344	540.922	518.000
30) L6 AR-1254-5	7.817	6.937	36073289	35991987	530.130	540.741

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020725\
Data File : PP069600.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Feb 2025 10:18
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/10/2025
Supervised By :Ankita Jodhani 02/10/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 07 13:52:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 29 00:13:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.535 min
Delta R.T.: -0.003 min
Response: 80692391
Conc: 56.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/10/2025
Supervised By :Ankita Jodhani 02/10/2025




#1 Tetrachloro-m-xylene

R.T.: 3.837 min
Delta R.T.: -0.003 min
Response: 51803654
Conc: 58.04 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.269 min
Delta R.T.: -0.009 min
Response: 58805925
Conc: 53.81 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.901 min
Delta R.T.: -0.010 min
Response: 55487193
Conc: 49.61 ng/ml




#26 AR-1254-1

R.T.: 6.537 min
Delta R.T.: -0.003 min
Response: 32837460
Conc: 567.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/10/2025
Supervised By :Ankita Jodhani 02/10/2025




#26 AR-1254-1

R.T.: 5.736 min
Delta R.T.: -0.005 min
Response: 29746575
Conc: 540.30 ng/ml




#27 AR-1254-2

R.T.: 6.755 min
Delta R.T.: -0.002 min
Response: 43669337
Conc: 544.22 ng/ml



#27 AR-1254-2

R.T.: 5.884 min
Delta R.T.: -0.005 min
Response: 25879466
Conc: 535.04 ng/ml




#28 AR-1254-3

R.T.: 7.117 min
Delta R.T.: -0.003 min
Response: 45517821
Conc: 540.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/10/2025
Supervised By :Ankita Jodhani 02/10/2025




#28 AR-1254-3

R.T.: 6.289 min
Delta R.T.: -0.006 min
Response: 39975694
Conc: 526.72 ng/ml




#29 AR-1254-4

R.T.: 7.400 min
Delta R.T.: -0.003 min
Response: 36884288
Conc: 540.92 ng/ml



#29 AR-1254-4

R.T.: 6.518 min
Delta R.T.: -0.006 min
Response: 26547344
Conc: 518.00 ng/ml




#30 AR-1254-5

R.T.: 7.817 min
Delta R.T.: -0.002 min
Response: 36073289
Conc: 530.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/10/2025
Supervised By :Ankita Jodhani 02/10/2025



#30 AR-1254-5

R.T.: 6.937 min
Delta R.T.: -0.007 min
Response: 35991987
Conc: 540.74 ng/ml