

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020822\
 Data File : PP043542.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Feb 2022 10:31
 Operator : AJ\MA
 Sample : N1422-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 09 00:49:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Feb 03 10:01:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.868	4.027	549470	462649	20.021	20.427
2)	SA Decachlor...	10.783	9.349	401175	510602	26.152	27.010
Target Compounds							
3)	L1 AR-1016-1	6.169	5.288	265813	322814	338.682	392.129
4)	L1 AR-1016-2	6.193	5.308	433115	471043	360.772	394.768
5)	L1 AR-1016-3	6.259	5.502	284967	268395	381.895	416.494
6)	L1 AR-1016-4	6.364	5.553	224971	474146	380.368	934.591 #
7)	L1 AR-1016-5	6.678	5.784	404814	444395	719.204	687.671
8)	L2 AR-1221-1	5.103	4.283	55481	28084	192.283	94.916 #
9)	L2 AR-1221-2	5.207	4.385	46639	45143	209.513	198.736
10)	L2 AR-1221-3	5.293	4.474	197225	168901	272.963	249.449
11)	L3 AR-1232-1	5.293	4.474	197225	168901	329.663	301.096
12)	L3 AR-1232-2	5.878	5.308	347836	471043	1321.633	936.722 #
13)	L3 AR-1232-3	6.193	5.502	433115	268395	884.063	1006.623
14)	L3 AR-1232-4	6.364	5.597	224971	303180	977.933	1318.717 #
15)	L3 AR-1232-5	6.464	5.784	437170	444395	2715.294	1711.448 #
16)	L4 AR-1242-1	6.169	5.288	265813	322814	421.267	475.790
17)	L4 AR-1242-2	6.193	5.308	433115	471043	445.952	478.031
18)	L4 AR-1242-3	6.259	5.502	284967	268395	484.278	494.611
19)	L4 AR-1242-4	6.364	5.597	224971	303180	469.598	598.575 #
20)	L4 AR-1242-5	7.144	6.164	455609	1329732	1037.000	2142.367 #
21)	L5 AR-1248-1	6.169	5.288	265813	322814	551.913	645.299
22)	L5 AR-1248-2	6.464	5.553	437170	474146	632.240	662.329
23)	L5 AR-1248-3	6.678	5.597	404814	303180	504.661	407.628
24)	L5 AR-1248-4	7.103	5.784	508373	444395	656.674	495.950
25)	L5 AR-1248-5	7.144	6.207	455609	315423	602.923	361.425 #
26)	L6 AR-1254-1	7.075	6.164	898775	1329732	1038.459	992.651
27)	L6 AR-1254-2	7.303	6.323	1267438	1115070	987.551	956.438
28)	L6 AR-1254-3	7.681	6.747	1065134	1513478	829.694	822.122
29)	L6 AR-1254-4	7.975	6.986	532996	616664	620.343	558.124
30)	L6 AR-1254-5	8.402	7.417	1044659	1062071	1125.343	687.985 #
31)	L7 AR-1260-1	7.851	6.883	704089	903354	781.234	791.021
32)	L7 AR-1260-2	8.114	7.080	653592	827827	646.057	606.287
33)	L7 AR-1260-3	8.480	7.236	301584	804841	392.102	628.559 #
34)	L7 AR-1260-4	8.706	7.721	601432	356365	664.129	346.040 #
35)	L7 AR-1260-5	9.023	7.969	577884	798166	365.614	362.689
36)	L8 AR-1262-1	8.480	7.417	301584	1062071	278.727	1365.187 #
37)	L8 AR-1262-2	9.023	7.721	577884	356365	339.992	265.624
38)	L8 AR-1262-3	9.342	0.000	143014	0	118.257	N.D. #
39)	L8 AR-1262-4	9.418	8.318	106408	561590	202.328	303.568 #
40)	L8 AR-1262-5	10.057	8.819	173684	168676	313.413	196.032 #
41)	L9 AR-1268-1	9.342f	0.000	143014	0	65.679	N.D. #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020822\
Data File : PP043542.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Feb 2022 10:31
Operator : AJ\MA
Sample : N1422-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 00:49:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Feb 03 10:01:05 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	9.418	8.318	106408	561590	51.125	206.879 #
43)	L9 AR-1268-3	9.641	8.529	12212	36822	6.781	15.721 #
44)	L9 AR-1268-4	10.057	8.819	173684	168676	260.594	170.457 #
45)	L9 AR-1268-5	10.459	9.102	93170	125461	16.817	18.047



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020822\
Data File : PP043542.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Feb 2022 10:31
Operator : AJ\MA
Sample : N1422-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 00:49:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Feb 03 10:01:05 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.868 min
Delta R.T.: 0.003 min
Response: 549470
Conc: 20.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.027 min
Delta R.T.: 0.003 min
Response: 462649
Conc: 20.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.783 min
Delta R.T.: 0.003 min
Response: 401175
Conc: 26.15 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.349 min
Delta R.T.: 0.002 min
Response: 510602
Conc: 27.01 ng/ml



#3 AR-1016-1


R.T.: 6.169 min
Delta R.T.: 0.002 min
Response: 265813
Conc: 338.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.288 min
Delta R.T.: 0.001 min
Response: 322814
Conc: 392.13 ng/ml




#4 AR-1016-2

R.T.: 6.193 min
Delta R.T.: 0.002 min
Response: 433115
Conc: 360.77 ng/ml



#4 AR-1016-2


R.T.: 5.308 min
Delta R.T.: 0.000 min
Response: 471043
Conc: 394.77 ng/ml



#5 AR-1016-3


R.T.: 6.259 min
Delta R.T.: 0.002 min
Response: 284967
Conc: 381.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.502 min
Delta R.T.: 0.001 min
Response: 268395
Conc: 416.49 ng/ml




#6 AR-1016-4

R.T.: 6.364 min
Delta R.T.: 0.002 min
Response: 224971
Conc: 380.37 ng/ml



#6 AR-1016-4


R.T.: 5.553 min
Delta R.T.: 0.000 min
Response: 474146
Conc: 934.59 ng/ml



#7 AR-1016-5


R.T.: 6.678 min
Delta R.T.: 0.002 min
Response: 404814
Conc: 719.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.784 min
Delta R.T.: 0.001 min
Response: 444395
Conc: 687.67 ng/ml




#8 AR-1221-1

R.T.: 5.103 min
Delta R.T.: -0.009 min
Response: 55481
Conc: 192.28 ng/ml



#8 AR-1221-1


R.T.: 4.283 min
Delta R.T.: -0.001 min
Response: 28084
Conc: 94.92 ng/ml



#9 AR-1221-2


R.T.: 5.207 min
Delta R.T.: -0.002 min
Response: 46639
Conc: 209.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.385 min
Delta R.T.: 0.000 min
Response: 45143
Conc: 198.74 ng/ml




#10 AR-1221-3

R.T.: 5.293 min
Delta R.T.: 0.000 min
Response: 197225
Conc: 272.96 ng/ml



#10 AR-1221-3


R.T.: 4.474 min
Delta R.T.: 0.000 min
Response: 168901
Conc: 249.45 ng/ml



#11 AR-1232-1


R.T.: 5.293 min
Delta R.T.: 0.001 min
Response: 197225
Conc: 329.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.474 min
Delta R.T.: 0.001 min
Response: 168901
Conc: 301.10 ng/ml




#12 AR-1232-2

R.T.: 5.878 min
Delta R.T.: 0.000 min
Response: 347836
Conc: 1321.63 ng/ml



#12 AR-1232-2


R.T.: 5.308 min
Delta R.T.: 0.000 min
Response: 471043
Conc: 936.72 ng/ml



#13 AR-1232-3


R.T.: 6.193 min
Delta R.T.: 0.001 min
Response: 433115
Conc: 884.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.502 min
Delta R.T.: 0.000 min
Response: 268395
Conc: 1006.62 ng/ml




#14 AR-1232-4

R.T.: 6.364 min
Delta R.T.: 0.000 min
Response: 224971
Conc: 977.93 ng/ml



#14 AR-1232-4


R.T.: 5.597 min
Delta R.T.: 0.000 min
Response: 303180
Conc: 1318.72 ng/ml



#15 AR-1232-5


R.T.: 6.464 min
Delta R.T.: 0.000 min
Response: 437170
Conc: 2715.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.784 min
Delta R.T.: 0.000 min
Response: 444395
Conc: 1711.45 ng/ml




#16 AR-1242-1

R.T.: 6.169 min
Delta R.T.: 0.001 min
Response: 265813
Conc: 421.27 ng/ml



#16 AR-1242-1


R.T.: 5.288 min
Delta R.T.: 0.000 min
Response: 322814
Conc: 475.79 ng/ml



#17 AR-1242-2


R.T.: 6.193 min
Delta R.T.: 0.002 min
Response: 433115
Conc: 445.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.308 min
Delta R.T.: 0.000 min
Response: 471043
Conc: 478.03 ng/ml




#18 AR-1242-3

R.T.: 6.259 min
Delta R.T.: 0.000 min
Response: 284967
Conc: 484.28 ng/ml



#18 AR-1242-3


R.T.: 5.502 min
Delta R.T.: 0.000 min
Response: 268395
Conc: 494.61 ng/ml



#19 AR-1242-4


R.T.: 6.364 min
Delta R.T.: 0.001 min
Response: 224971
Conc: 469.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.597 min
Delta R.T.: 0.000 min
Response: 303180
Conc: 598.57 ng/ml




#20 AR-1242-5

R.T.: 7.144 min
Delta R.T.: 0.003 min
Response: 455609
Conc: 1037.00 ng/ml



#20 AR-1242-5


R.T.: 6.164 min
Delta R.T.: 0.000 min
Response: 1329732
Conc: 2142.37 ng/ml



#21 AR-1248-1


R.T.: 6.169 min
Delta R.T.: 0.001 min
Response: 265813
Conc: 551.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.288 min
Delta R.T.: 0.000 min
Response: 322814
Conc: 645.30 ng/ml




#22 AR-1248-2

R.T.: 6.464 min
Delta R.T.: 0.001 min
Response: 437170
Conc: 632.24 ng/ml



#22 AR-1248-2


R.T.: 5.553 min
Delta R.T.: 0.000 min
Response: 474146
Conc: 662.33 ng/ml



#23 AR-1248-3


R.T.: 6.678 min
Delta R.T.: 0.000 min
Response: 404814
Conc: 504.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.597 min
Delta R.T.: 0.000 min
Response: 303180
Conc: 407.63 ng/ml




#24 AR-1248-4

R.T.: 7.103 min
Delta R.T.: 0.002 min
Response: 508373
Conc: 656.67 ng/ml



#24 AR-1248-4


R.T.: 5.784 min
Delta R.T.: 0.000 min
Response: 444395
Conc: 495.95 ng/ml



#25 AR-1248-5


R.T.: 7.144 min
Delta R.T.: 0.003 min
Response: 455609
Conc: 602.92 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.207 min
Delta R.T.: 0.000 min
Response: 315423
Conc: 361.42 ng/ml




#26 AR-1254-1

R.T.: 7.075 min
Delta R.T.: 0.002 min
Response: 898775
Conc: 1038.46 ng/ml



#26 AR-1254-1


R.T.: 6.164 min
Delta R.T.: 0.000 min
Response: 1329732
Conc: 992.65 ng/ml



#27 AR-1254-2


R.T.: 7.303 min
Delta R.T.: 0.002 min
Response: 1267438
Conc: 987.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.323 min
Delta R.T.: 0.000 min
Response: 1115070
Conc: 956.44 ng/ml




#28 AR-1254-3

R.T.: 7.681 min
Delta R.T.: 0.002 min
Response: 1065134
Conc: 829.69 ng/ml



#28 AR-1254-3


R.T.: 6.747 min
Delta R.T.: 0.000 min
Response: 1513478
Conc: 822.12 ng/ml



#29 AR-1254-4


R.T.: 7.975 min
Delta R.T.: 0.002 min
Response: 532996
Conc: 620.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.986 min
Delta R.T.: 0.000 min
Response: 616664
Conc: 558.12 ng/ml




#30 AR-1254-5

R.T.: 8.402 min
Delta R.T.: 0.002 min
Response: 1044659
Conc: 1125.34 ng/ml



#30 AR-1254-5


R.T.: 7.417 min
Delta R.T.: 0.000 min
Response: 1062071
Conc: 687.99 ng/ml



#31 AR-1260-1


R.T.: 7.851 min
Delta R.T.: 0.003 min
Response: 704089
Conc: 781.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.883 min
Delta R.T.: 0.002 min
Response: 903354
Conc: 791.02 ng/ml




#32 AR-1260-2

R.T.: 8.114 min
Delta R.T.: 0.003 min
Response: 653592
Conc: 646.06 ng/ml



#32 AR-1260-2


R.T.: 7.080 min
Delta R.T.: 0.002 min
Response: 827827
Conc: 606.29 ng/ml



#33 AR-1260-3


R.T.: 8.480 min
Delta R.T.: 0.002 min
Response: 301584
Conc: 392.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.236 min
Delta R.T.: 0.000 min
Response: 804841
Conc: 628.56 ng/ml




#34 AR-1260-4

R.T.: 8.706 min
Delta R.T.: 0.000 min
Response: 601432
Conc: 664.13 ng/ml



#34 AR-1260-4


R.T.: 7.721 min
Delta R.T.: 0.002 min
Response: 356365
Conc: 346.04 ng/ml



#35 AR-1260-5


R.T.: 9.023 min
Delta R.T.: 0.002 min
Response: 577884
Conc: 365.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.969 min
Delta R.T.: 0.002 min
Response: 798166
Conc: 362.69 ng/ml




#36 AR-1262-1

R.T.: 8.480 min
Delta R.T.: 0.000 min
Response: 301584
Conc: 278.73 ng/ml



#36 AR-1262-1


R.T.: 7.417 min
Delta R.T.: 0.000 min
Response: 1062071
Conc: 1365.19 ng/ml



#37 AR-1262-2


R.T.: 9.023 min
Delta R.T.: 0.000 min
Response: 577884
Conc: 339.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.721 min
Delta R.T.: 0.000 min
Response: 356365
Conc: 265.62 ng/ml




#38 AR-1262-3

R.T.: 9.342 min
Delta R.T.: 0.011 min
Response: 143014
Conc: 118.26 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T. : 8.254 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 9.418 min
Delta R.T.: 0.003 min
Response: 106408
Conc: 202.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.318 min
Delta R.T.: -0.002 min
Response: 561590
Conc: 303.57 ng/ml




#40 AR-1262-5

R.T.: 10.057 min
Delta R.T.: 0.002 min
Response: 173684
Conc: 313.41 ng/ml



#40 AR-1262-5


R.T.: 8.819 min
Delta R.T.: 0.000 min
Response: 168676
Conc: 196.03 ng/ml



#41 AR-1268-1


R.T.: 9.342 min
Delta R.T.: 0.015 min
Response: 143014
Conc: 65.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 8.255 min
Response: 0
Conc: N.D.




#42 AR-1268-2

R.T.: 9.418 min
Delta R.T.: -0.002 min
Response: 106408
Conc: 51.13 ng/ml



#42 AR-1268-2


R.T.: 8.318 min
Delta R.T.: -0.003 min
Response: 561590
Conc: 206.88 ng/ml



#43 AR-1268-3


R.T.: 9.641 min
Delta R.T.: 0.000 min
Response: 12212
Conc: 6.78 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.529 min
Delta R.T.: 0.001 min
Response: 36822
Conc: 15.72 ng/ml




#44 AR-1268-4

R.T.: 10.057 min
Delta R.T.: 0.001 min
Response: 173684
Conc: 260.59 ng/ml



#44 AR-1268-4


R.T.: 8.819 min
Delta R.T.: 0.000 min
Response: 168676
Conc: 170.46 ng/ml



#45 AR-1268-5

R.T.: 10.459 min
Delta R.T.: 0.002 min
Response: 93170
Conc: 16.82 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.102 min
Delta R.T.: 0.000 min
Response: 125461
Conc: 18.05 ng/ml