

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020923\
 Data File : PP055490.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Feb 2023 13:15
 Operator : YP\AJ
 Sample : PB150751BS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 09 21:31:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.329	3.576	31895960	25424101	18.344	16.936
2)	SA Decachlor...	10.080	8.581	33604191	26649688	19.406	19.732
Target Compounds							
3)	L1 AR-1016-1	5.498	4.655	24215931	18704103	360.302	377.907
4)	L1 AR-1016-2	5.520	4.674	35324569	27056725	368.620	390.885
5)	L1 AR-1016-3	5.582	4.850	21858156	14334200	365.421	385.363
6)	L1 AR-1016-4	5.681	4.892	17534987	12587779	365.937	400.601
7)	L1 AR-1016-5	5.977	5.105	17561061	15580040	346.549	381.074
8)	L2 AR-1221-1	4.532	3.786	2306128	1946879	98.070	98.194
9)	L2 AR-1221-2	4.623	3.872	3018002	2520678	169.866	170.059
10)	L2 AR-1221-3	4.699	3.947	13341800	11113191	247.944	246.333
11)	L3 AR-1232-1	4.699	3.947	13341800	11113191	300.607	301.502
12)	L3 AR-1232-2	5.229	4.674	17877171	27056725	790.508	847.044
13)	L3 AR-1232-3	5.520	4.850	35324569	14334200	804.343	859.830
14)	L3 AR-1232-4	5.681	4.934	17534987	15311629	792.212	949.492
15)	L3 AR-1232-5	5.773	5.105	16196410	15580040	925.137	882.149
16)	L4 AR-1242-1	5.498	4.655	24215931	18704103	432.514	456.618
17)	L4 AR-1242-2	5.520	4.674	35324569	27056725	442.499	473.839
18)	L4 AR-1242-3	5.582	4.850	21858156	14334200	440.143	464.377
19)	L4 AR-1242-4	5.681	4.934	17534987	15311629	437.535	471.314
20)	L4 AR-1242-5	6.422	5.458	2715714	12283500	61.077	328.685 #
21)	L5 AR-1248-1	5.498	4.655	24215931	18704103	580.666	603.297
22)	L5 AR-1248-2	5.773	4.892	16196410	12587779	254.764	275.278
23)	L5 AR-1248-3	5.977	4.934	17561061	15311629	253.926	321.179 #
24)	L5 AR-1248-4	6.382	5.105	2698898	15580040	35.590	282.108 #
25)	L5 AR-1248-5	6.422	5.498	2715714	1532347	36.802	31.164
26)	L6 AR-1254-1	6.356	5.458	13994062	12283500	177.382	152.102
27)	L6 AR-1254-2	6.575	5.605	14848164	11729633	122.689	163.907 #
28)	L6 AR-1254-3	6.944	6.024	7983960	20895383	65.181	187.558 #
29)	L6 AR-1254-4	7.230	6.239	4750462	1635551	53.914	26.591 #
30)	L6 AR-1254-5	7.650	6.657	45923227	38494994	453.963	406.048
31)	L7 AR-1260-1	7.108	6.140	33716993	30468989	353.700	392.923
32)	L7 AR-1260-2	7.367	6.329	38431199	35049120	341.470	389.876
33)	L7 AR-1260-3	7.728	6.482	25197297	32303184	312.937	380.437
34)	L7 AR-1260-4	7.954	6.955	32702002	24591379	336.867	343.442
35)	L7 AR-1260-5	8.271	7.199	56914352	52535101	320.012	341.722
36)	L8 AR-1262-1	7.728	6.749	25197297	14619593	213.568	340.296 #
37)	L8 AR-1262-2	8.271	6.955	56914352	24591379	282.575	265.499
38)	L8 AR-1262-3	8.593	7.484	36565644	11037241	262.928	151.486 #
39)	L8 AR-1262-4	8.671	7.546	9104965	39370416	118.429	301.045 #
40)	L8 AR-1262-5	9.328	8.043	13091608	6056223	183.049	102.081 #
41)	L9 AR-1268-1	8.593	7.484	36565644	11037241	149.610	51.045 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020923\
Data File : PP055490.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Feb 2023 13:15
Operator : YP\AJ
Sample : PB150751BS
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 21:31:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.671	7.546	9104965	39370416	40.088	201.761 #
43) L9	AR-1268-3	8.903	7.755	1341677	647405	6.716	3.769 #
44) L9	AR-1268-4	9.328	8.043	13091608	6056223	166.674	94.951 #
45) L9	AR-1268-5	9.742	8.331	4726151	3467821	7.430	6.585



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP020923\
Data File : PP055490.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Feb 2023 13:15
Operator : YP\AJ
Sample : PB150751BS
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 21:31:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.329 min
Delta R.T.: -0.003 min
Response: 31895960
Conc: 18.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.576 min
Delta R.T.: -0.003 min
Response: 25424101
Conc: 16.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.080 min
Delta R.T.: -0.010 min
Response: 33604191
Conc: 19.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.581 min
Delta R.T.: -0.017 min
Response: 26649688
Conc: 19.73 ng/ml



#3 AR-1016-1


R.T.: 5.498 min
Delta R.T.: -0.005 min
Response: 24215931
Conc: 360.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.655 min
Delta R.T.: -0.008 min
Response: 18704103
Conc: 377.91 ng/ml




#4 AR-1016-2

R.T.: 5.520 min
Delta R.T.: -0.006 min
Response: 35324569
Conc: 368.62 ng/ml



#4 AR-1016-2


R.T.: 4.674 min
Delta R.T.: -0.009 min
Response: 27056725
Conc: 390.89 ng/ml



#5 AR-1016-3


R.T.: 5.582 min
Delta R.T.: -0.006 min
Response: 21858156
Conc: 365.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.850 min
Delta R.T.: -0.010 min
Response: 14334200
Conc: 385.36 ng/ml




#6 AR-1016-4

R.T.: 5.681 min
Delta R.T.: -0.006 min
Response: 17534987
Conc: 365.94 ng/ml



#6 AR-1016-4


R.T.: 4.892 min
Delta R.T.: -0.009 min
Response: 12587779
Conc: 400.60 ng/ml



#7 AR-1016-5


R.T.: 5.977 min
Delta R.T.: -0.007 min
Response: 17561061
Conc: 346.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.105 min
Delta R.T.: -0.010 min
Response: 15580040
Conc: 381.07 ng/ml




#8 AR-1221-1

R.T.: 4.532 min
Delta R.T.: -0.008 min
Response: 2306128
Conc: 98.07 ng/ml



#8 AR-1221-1


R.T.: 3.786 min
Delta R.T.: -0.006 min
Response: 1946879
Conc: 98.19 ng/ml



#9 AR-1221-2


R.T.: 4.623 min
Delta R.T.: -0.003 min
Response: 3018002
Conc: 169.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.872 min
Delta R.T.: -0.005 min
Response: 2520678
Conc: 170.06 ng/ml




#10 AR-1221-3

R.T.: 4.699 min
Delta R.T.: -0.003 min
Response: 13341800
Conc: 247.94 ng/ml



#10 AR-1221-3


R.T.: 3.947 min
Delta R.T.: -0.005 min
Response: 11113191
Conc: 246.33 ng/ml



#11 AR-1232-1


R.T.: 4.699 min
Delta R.T.: -0.005 min
Response: 13341800
Conc: 300.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.947 min
Delta R.T.: -0.006 min
Response: 11113191
Conc: 301.50 ng/ml




#12 AR-1232-2

R.T.: 5.229 min
Delta R.T.: -0.006 min
Response: 17877171
Conc: 790.51 ng/ml



#12 AR-1232-2


R.T.: 4.674 min
Delta R.T.: -0.009 min
Response: 27056725
Conc: 847.04 ng/ml



#13 AR-1232-3


R.T.: 5.520 min
Delta R.T.: -0.006 min
Response: 35324569
Conc: 804.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.850 min
Delta R.T.: -0.009 min
Response: 14334200
Conc: 859.83 ng/ml




#14 AR-1232-4

R.T.: 5.681 min
Delta R.T.: -0.007 min
Response: 17534987
Conc: 792.21 ng/ml



#14 AR-1232-4


R.T.: 4.934 min
Delta R.T.: -0.010 min
Response: 15311629
Conc: 949.49 ng/ml



#15 AR-1232-5


R.T.: 5.773 min
Delta R.T.: -0.007 min
Response: 16196410
Conc: 925.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.105 min
Delta R.T.: -0.010 min
Response: 15580040
Conc: 882.15 ng/ml




#16 AR-1242-1

R.T.: 5.498 min
Delta R.T.: -0.005 min
Response: 24215931
Conc: 432.51 ng/ml



#16 AR-1242-1


R.T.: 4.655 min
Delta R.T.: -0.009 min
Response: 18704103
Conc: 456.62 ng/ml



#17 AR-1242-2


R.T.: 5.520 min
Delta R.T.: -0.005 min
Response: 35324569
Conc: 442.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.674 min
Delta R.T.: -0.009 min
Response: 27056725
Conc: 473.84 ng/ml




#18 AR-1242-3

R.T.: 5.582 min
Delta R.T.: -0.006 min
Response: 21858156
Conc: 440.14 ng/ml



#18 AR-1242-3


R.T.: 4.850 min
Delta R.T.: -0.010 min
Response: 14334200
Conc: 464.38 ng/ml



#19 AR-1242-4


R.T.: 5.681 min
Delta R.T.: -0.006 min
Response: 17534987
Conc: 437.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.934 min
Delta R.T.: -0.010 min
Response: 15311629
Conc: 471.31 ng/ml




#20 AR-1242-5

R.T.: 6.422 min
Delta R.T.: -0.005 min
Response: 2715714
Conc: 61.08 ng/ml



#20 AR-1242-5


R.T.: 5.458 min
Delta R.T.: -0.010 min
Response: 12283500
Conc: 328.68 ng/ml



#21 AR-1248-1


R.T.: 5.498 min
Delta R.T.: -0.004 min
Response: 24215931
Conc: 580.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.655 min
Delta R.T.: -0.008 min
Response: 18704103
Conc: 603.30 ng/ml




#22 AR-1248-2

R.T.: 5.773 min
Delta R.T.: -0.005 min
Response: 16196410
Conc: 254.76 ng/ml



#22 AR-1248-2


R.T.: 4.892 min
Delta R.T.: -0.009 min
Response: 12587779
Conc: 275.28 ng/ml



#23 AR-1248-3


R.T.: 5.977 min
Delta R.T.: -0.005 min
Response: 17561061
Conc: 253.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.934 min
Delta R.T.: -0.009 min
Response: 15311629
Conc: 321.18 ng/ml




#24 AR-1248-4

R.T.: 6.382 min
Delta R.T.: -0.006 min
Response: 2698898
Conc: 35.59 ng/ml



#24 AR-1248-4


R.T.: 5.105 min
Delta R.T.: -0.009 min
Response: 15580040
Conc: 282.11 ng/ml



#25 AR-1248-5


R.T.: 6.422 min
Delta R.T.: -0.005 min
Response: 2715714
Conc: 36.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.498 min
Delta R.T.: -0.010 min
Response: 1532347
Conc: 31.16 ng/ml




#26 AR-1254-1

R.T.: 6.356 min
Delta R.T.: -0.006 min
Response: 13994062
Conc: 177.38 ng/ml



#26 AR-1254-1


R.T.: 5.458 min
Delta R.T.: -0.009 min
Response: 12283500
Conc: 152.10 ng/ml



#27 AR-1254-2


R.T.: 6.575 min
Delta R.T.: -0.006 min
Response: 14848164
Conc: 122.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.605 min
Delta R.T.: -0.010 min
Response: 11729633
Conc: 163.91 ng/ml




#28 AR-1254-3

R.T.: 6.944 min
Delta R.T.: -0.005 min
Response: 7983960
Conc: 65.18 ng/ml



#28 AR-1254-3


R.T.: 6.024 min
Delta R.T.: 0.005 min
Response: 20895383
Conc: 187.56 ng/ml



#29 AR-1254-4


R.T.: 7.230 min
Delta R.T.: -0.006 min
Response: 4750462
Conc: 53.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.239 min
Delta R.T.: -0.010 min
Response: 1635551
Conc: 26.59 ng/ml




#30 AR-1254-5

R.T.: 7.650 min
Delta R.T.: -0.007 min
Response: 45923227
Conc: 453.96 ng/ml



#30 AR-1254-5


R.T.: 6.657 min
Delta R.T.: -0.011 min
Response: 38494994
Conc: 406.05 ng/ml



#31 AR-1260-1


R.T.: 7.108 min
Delta R.T.: -0.008 min
Response: 33716993
Conc: 353.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.140 min
Delta R.T.: -0.012 min
Response: 30468989
Conc: 392.92 ng/ml




#32 AR-1260-2

R.T.: 7.367 min
Delta R.T.: -0.007 min
Response: 38431199
Conc: 341.47 ng/ml



#32 AR-1260-2


R.T.: 6.329 min
Delta R.T.: -0.012 min
Response: 35049120
Conc: 389.88 ng/ml



#33 AR-1260-3


R.T.: 7.728 min
Delta R.T.: -0.007 min
Response: 25197297
Conc: 312.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.482 min
Delta R.T.: -0.013 min
Response: 32303184
Conc: 380.44 ng/ml




#34 AR-1260-4

R.T.: 7.954 min
Delta R.T.: -0.008 min
Response: 32702002
Conc: 336.87 ng/ml



#34 AR-1260-4


R.T.: 6.955 min
Delta R.T.: -0.014 min
Response: 24591379
Conc: 343.44 ng/ml



#35 AR-1260-5


R.T.: 8.271 min
Delta R.T.: -0.006 min
Response: 56914352
Conc: 320.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.199 min
Delta R.T.: -0.013 min
Response: 52535101
Conc: 341.72 ng/ml




#36 AR-1262-1

R.T.: 7.728 min
Delta R.T.: -0.006 min
Response: 25197297
Conc: 213.57 ng/ml



#36 AR-1262-1


R.T.: 6.749 min
Delta R.T.: -0.011 min
Response: 14619593
Conc: 340.30 ng/ml



#37 AR-1262-2


R.T.: 8.271 min
Delta R.T.: -0.006 min
Response: 56914352
Conc: 282.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.955 min
Delta R.T.: -0.012 min
Response: 24591379
Conc: 265.50 ng/ml




#38 AR-1262-3

R.T.: 8.593 min
Delta R.T.: 0.003 min
Response: 36565644
Conc: 262.93 ng/ml



#38 AR-1262-3


R.T.: 7.484 min
Delta R.T.: -0.011 min
Response: 11037241
Conc: 151.49 ng/ml



#39 AR-1262-4


R.T.: 8.671 min
Delta R.T.: -0.007 min
Response: 9104965
Conc: 118.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.546 min
Delta R.T.: -0.013 min
Response: 39370416
Conc: 301.04 ng/ml




#40 AR-1262-5

R.T.: 9.328 min
Delta R.T.: -0.006 min
Response: 13091608
Conc: 183.05 ng/ml



#40 AR-1262-5


R.T.: 8.043 min
Delta R.T.: -0.011 min
Response: 6056223
Conc: 102.08 ng/ml



#41 AR-1268-1


R.T.: 8.593 min
Delta R.T.: 0.008 min
Response: 36565644
Conc: 149.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.484 min
Delta R.T.: -0.010 min
Response: 11037241
Conc: 51.05 ng/ml




#42 AR-1268-2

R.T.: 8.671 min
Delta R.T.: -0.009 min
Response: 9104965
Conc: 40.09 ng/ml



#42 AR-1268-2


R.T.: 7.546 min
Delta R.T.: -0.013 min
Response: 39370416
Conc: 201.76 ng/ml



#43 AR-1268-3


R.T.: 8.903 min
Delta R.T.: -0.004 min
Response: 1341677
Conc: 6.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.755 min
Delta R.T.: -0.012 min
Response: 647405
Conc: 3.77 ng/ml




#44 AR-1268-4

R.T.: 9.328 min
Delta R.T.: -0.005 min
Response: 13091608
Conc: 166.67 ng/ml



#44 AR-1268-4


R.T.: 8.043 min
Delta R.T.: -0.011 min
Response: 6056223
Conc: 94.95 ng/ml



#45 AR-1268-5

R.T.: 9.742 min
Delta R.T.: -0.007 min
Response: 4726151
Conc: 7.43 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.331 min
Delta R.T.: -0.012 min
Response: 3467821
Conc: 6.58 ng/ml