

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021022\
 Data File : PP043569.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Feb 2022 18:36
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 11 02:39:24 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 02:39:15 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.919	3.174	163.2E6	143.0E6	75.111	75.398
2) SA Decachlor...	10.071	8.555	90910914	116.4E6	74.370	74.477
Target Compounds						
21) L5 AR-1248-1	5.174	4.323	29459169	25640426	753.046	754.534
22) L5 AR-1248-2	5.472	4.576	40350283	33192358	748.825	751.834
23) L5 AR-1248-3	5.691	4.619	47638478	34933793	748.142	749.801
24) L5 AR-1248-4	6.130	4.801	49885262	40477607	751.357	753.011
25) L5 AR-1248-5	6.172	5.223	49285202	39437291	750.555	751.393



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021022\
Data File : PP043569.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Feb 2022 18:36
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 11 02:39:24 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 02:39:15 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.919 min
Delta R.T.: 0.000 min
Response: 163227755
Conc: 75.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.174 min
Delta R.T.: 0.000 min
Response: 143012341
Conc: 75.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.071 min
Delta R.T.: 0.000 min
Response: 90910914
Conc: 74.37 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.555 min
Delta R.T.: 0.000 min
Response: 116357432
Conc: 74.48 ng/ml



#21 AR-1248-1


R.T.: 5.174 min
Delta R.T.: 0.000 min
Response: 29459169
Conc: 753.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 4.323 min
Delta R.T.: 0.000 min
Response: 25640426
Conc: 754.53 ng/ml




#22 AR-1248-2

R.T.: 5.472 min
Delta R.T.: 0.000 min
Response: 40350283
Conc: 748.83 ng/ml



#22 AR-1248-2


R.T.: 4.576 min
Delta R.T.: 0.000 min
Response: 33192358
Conc: 751.83 ng/ml



#23 AR-1248-3


R.T.: 5.691 min
Delta R.T.: 0.000 min
Response: 47638478
Conc: 748.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 4.619 min
Delta R.T.: 0.000 min
Response: 34933793
Conc: 749.80 ng/ml




#24 AR-1248-4

R.T.: 6.130 min
Delta R.T.: 0.000 min
Response: 49885262
Conc: 751.36 ng/ml



#24 AR-1248-4


R.T.: 4.801 min
Delta R.T.: 0.000 min
Response: 40477607
Conc: 753.01 ng/ml



#25 AR-1248-5

R.T.: 6.172 min
Delta R.T.: 0.000 min
Response: 49285202
Conc: 750.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 5.223 min
Delta R.T.: 0.000 min
Response: 39437291
Conc: 751.39 ng/ml