

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021022\
 Data File : PP043573.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Feb 2022 19:38
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 11 03:44:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 03:44:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.920	3.175	211.8E6	184.4E6	97.718	97.499
2) SA Decachlor...	10.071	8.555	116.8E6	148.8E6	96.501	95.210
Target Compounds						
26) L6 AR-1254-1	6.102	5.179	67024735	75785692	960.085	953.959
27) L6 AR-1254-2	6.342	5.339	99765092	65160003	958.412	948.214
28) L6 AR-1254-3	6.742	5.772	103.6E6	108.2E6	960.516	959.638
29) L6 AR-1254-4	7.056	6.021	71140949	69404598	962.129	954.835
30) L6 AR-1254-5	7.517	6.474	77823171	103.3E6	966.025	959.158



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021022\
Data File : PP043573.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Feb 2022 19:38
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 11 03:44:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 03:44:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.920 min
Delta R.T.: 0.000 min
Response: 211847973
Conc: 97.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.175 min
Delta R.T.: 0.000 min
Response: 184385621
Conc: 97.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.071 min
Delta R.T.: 0.000 min
Response: 116767262
Conc: 96.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.555 min
Delta R.T.: 0.000 min
Response: 148830951
Conc: 95.21 ng/ml



#26 AR-1254-1


R.T.: 6.102 min
Delta R.T.: 0.000 min
Response: 67024735
Conc: 960.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#26 AR-1254-1

R.T.: 5.179 min
Delta R.T.: 0.000 min
Response: 75785692
Conc: 953.96 ng/ml




#27 AR-1254-2

R.T.: 6.342 min
Delta R.T.: 0.000 min
Response: 99765092
Conc: 958.41 ng/ml



#27 AR-1254-2


R.T.: 5.339 min
Delta R.T.: 0.000 min
Response: 65160003
Conc: 948.21 ng/ml



#28 AR-1254-3


R.T.: 6.742 min
Delta R.T.: 0.000 min
Response: 103568134
Conc: 960.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#28 AR-1254-3

R.T.: 5.772 min
Delta R.T.: 0.000 min
Response: 108160703
Conc: 959.64 ng/ml




#29 AR-1254-4

R.T.: 7.056 min
Delta R.T.: 0.000 min
Response: 71140949
Conc: 962.13 ng/ml



#29 AR-1254-4


R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 69404598
Conc: 954.84 ng/ml



#30 AR-1254-5

R.T.: 7.517 min
Delta R.T.: 0.000 min
Response: 77823171
Conc: 966.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000



#30 AR-1254-5

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 103313533
Conc: 959.16 ng/ml