

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021022\
 Data File : PP043582.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Feb 2022 21:55
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/11/2022
 Supervised By :Ankita Jodhani 02/14/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 11 04:12:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:12:41 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.920	3.174	55372833	48970739	25.781	26.092
2) SA Decachlor...	10.071	8.555	55126083	72951497	26.360	27.013
Target Compounds						
41) L9 AR-1268-1	8.522	7.372	49537525	63514086	262.040m	267.018
42) L9 AR-1268-2	8.621	7.442	44312028	57179029	261.388	264.985
43) L9 AR-1268-3	8.859	7.665	39499539	49315032	262.871	266.446
44) L9 AR-1268-4	9.303	7.983	15950944	23297000	259.474	271.346
45) L9 AR-1268-5	9.728	8.290	115.2E6	150.3E6	259.192	259.837

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021022\
Data File : PP043582.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Feb 2022 21:55
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/11/2022
Supervised By :Ankita Jodhani 02/14/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 11 04:12:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:12:41 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.920 min
Delta R.T.: 0.000 min
Response: 55372833
Conc: 25.78 ng/ml

Instrument :

ECD_P


ClientSampleId :

AR1268ICC250

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 02/11/2022
Supervised By :Ankita Jodhani 02/14/2022




#1 Tetrachloro-m-xylene

R.T.: 3.174 min
Delta R.T.: 0.000 min
Response: 48970739
Conc: 26.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.071 min
Delta R.T.: 0.000 min
Response: 55126083
Conc: 26.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.555 min
Delta R.T.: 0.000 min
Response: 72951497
Conc: 27.01 ng/ml




#41 AR-1268-1

R.T.: 8.522 min
Delta R.T.: 0.000 min
Response: 49537525
Conc: 262.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/11/2022
Supervised By :Ankita Jodhani 02/14/2022




#41 AR-1268-1

R.T.: 7.372 min
Delta R.T.: 0.000 min
Response: 63514086
Conc: 267.02 ng/ml




#42 AR-1268-2

R.T.: 8.621 min
Delta R.T.: 0.000 min
Response: 44312028
Conc: 261.39 ng/ml



#42 AR-1268-2

R.T.: 7.442 min
Delta R.T.: 0.000 min
Response: 57179029
Conc: 264.99 ng/ml




#43 AR-1268-3

R.T.: 8.859 min
Delta R.T.: 0.000 min
Response: 39499539
Conc: 262.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/11/2022
Supervised By :Ankita Jodhani 02/14/2022




#43 AR-1268-3

R.T.: 7.665 min
Delta R.T.: 0.000 min
Response: 49315032
Conc: 266.45 ng/ml




#44 AR-1268-4

R.T.: 9.303 min
Delta R.T.: 0.000 min
Response: 15950944
Conc: 259.47 ng/ml



#44 AR-1268-4

R.T.: 7.983 min
Delta R.T.: 0.000 min
Response: 23297000
Conc: 271.35 ng/ml




#45 AR-1268-5

R.T.: 9.728 min
Delta R.T.: 0.000 min
Response: 115214490
Conc: 259.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/11/2022
Supervised By :Ankita Jodhani 02/14/2022



#45 AR-1268-5

R.T.: 8.290 min
Delta R.T.: 0.000 min
Response: 150279264
Conc: 259.84 ng/ml